

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050818\
 Data File : BF105169.D
 Acq On : 9 May 2018 2:31
 Operator : JU/SJ
 Sample : J2133-09
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: May 09 11:36:58 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 08 16:33:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	202785	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	810090	20.00	ng	0.00
38) Acenaphthene-d10	9.83	164	400637	20.00	ng	0.00
63) Phenanthrene-d10	11.31	188	779463	20.00	ng	0.00
75) Chrysene-d12	13.99	240	668747	20.00	ng	-0.03
86) Perylene-d12	15.51	264	612789	20.00	ng	-0.05

System Monitoring Compounds

5) 2-Fluorophenol	5.40	112	1517931	123.32	ng	0.00
7) Phenol-d6	6.42	99	1822392	124.34	ng	0.00
23) Nitrobenzene-d5	7.36	82	1134953	94.95	ng	0.00
41) 2,4,6-Tribromophenol	10.62	330	580331	131.29	ng	0.00
44) 2-Fluorobiphenyl	9.15	172	2014020	78.02	ng	0.00
78) Terphenyl-d14	12.90	244	2187975	76.00	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.52	88	51378	7.895	ng	91
3) Pyridine	3.27	79	135278	7.751	ng	86
4) n-Nitrosodimethylamine	3.18	42	61296	6.929	ng	82
6) Aniline	6.46	93	177043	8.271	ng	94
8) 2-Chlorophenol	6.57	128	110024	8.426	ng	96
9) Benzaldehyde	6.35	77	56383	5.019	ng	97
10) Phenol	6.43	94	127650	7.826	ng	99
11) bis(2-Chloroethyl)ether	6.53	93	114737	8.535	ng	92
12) 1,3-Dichlorobenzene	6.73	146	142876	9.078	ng	98
13) 1,4-Dichlorobenzene	6.81	146	139922	8.891	ng	97
14) 1,2-Dichlorobenzene	6.96	146	128771	8.870	ng	97
15) Benzyl Alcohol	6.93	79	110084	9.814	ng	99
16) 2,2'-oxybis(1-Chloropropan	7.07	45	205886	8.242	ng	97
17) 2-Methylphenol	7.03	107	95060	8.432	ng	98
18) Hexachloroethane	7.30	117	45419	8.968	ng	96
19) n-Nitroso-di-n-propylamine	7.20	70	86477	8.911	ng	88
20) 3+4-Methylphenols	7.19	107	127934	9.656	ng	96
22) Acetophenone	7.20	105	179836	9.524	ng	# 96
24) Nitrobenzene	7.37	77	130570	9.634	ng	99
25) Isophorone	7.60	82	213364	7.982	ng	94
26) 2-Nitrophenol	7.69	139	46617	16.746	ng	95
27) 2,4-Dimethylphenol	7.72	122	86103	7.248	ng	99
28) bis(2-Chloroethoxy)methane	7.82	93	136107	8.250	ng	98
29) 2,4-Dichlorophenol	7.92	162	93625	8.424	ng	94
30) 1,2,4-Trichlorobenzene	8.01	180	117503	9.168	ng	97
31) Naphthalene	8.09	128	352800	9.045	ng	99
32) Benzoic acid	7.79	122	54759	16.529	ng	96
33) 4-Chloroaniline	8.14	127	137729	8.491	ng	97
34) Hexachlorobutadiene	8.21	225	69550	9.367	ng	97
35) Caprolactam	8.47	113	26059	7.721	ng	# 75
36) 4-Chloro-3-methylphenol	8.61	107	96380	8.330	ng	95
37) 2-Methylnaphthalene	8.78	142	243780	8.967	ng	97
39) 1,2,4,5-Tetrachlorobenzene	8.94	216	119698	9.389	ng	96
40) Hexachlorocyclopentadiene	8.93	237	56198	8.152	ng	96


Data Path : Z:\HPCHEM1\BNA F\DATA\BF050818\
 Data File : BF105169.D
 Acq On : 9 May 2018 2:31
 Operator : JU/SJ
 Sample : J2133-09
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: May 09 11:36:58 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 08 16:33:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.06	196	65356	8.402	ng	90
43) 2,4,5-Trichlorophenol	9.09	196	73550	9.050	ng	# 91
45) 1,1'-Biphenyl	9.24	154	317380	9.723	ng	99
46) 2-Chloronaphthalene	9.27	162	229842	9.421	ng	95
47) 2-Nitroaniline	9.36	65	61234	13.460	ng	# 83
48) Acenaphthylene	9.69	152	358578	9.516	ng	100
49) Dimethylphthalate	9.54	163	261960	9.658	ng	100
50) 2,6-Dinitrotoluene	9.60	165	54080	9.189	ng	97
51) Acenaphthene	9.86	154	231414	9.530	ng	98
52) 3-Nitroaniline	9.77	138	56610	9.132	ng	95
53) 2,4-Dinitrophenol	9.87	184	12330	16.794	ng	# 23
54) Dibenzofuran	10.03	168	319381	8.991	ng	99
55) 4-Nitrophenol	9.92	139	38705	10.865	ng	91
56) 2,4-Dinitrotoluene	10.00	165	66954	11.522	ng	95
57) Fluorene	10.37	166	249911	9.593	ng	99
58) 2,3,4,6-Tetrachlorophenol	10.14	232	59013	8.134	ng	# 88
59) Diethylphthalate	10.25	149	249507	9.361	ng	96
60) 4-Chlorophenyl-phenylether	10.37	204	122584	10.108	ng	89
61) 4-Nitroaniline	10.38	138	52257	8.472	ng	97
62) Azobenzene	10.53	77	244060	8.777	ng	96
64) 4,6-Dinitro-2-methylphenol	10.41	198	25225	17.325	ng	90
65) n-Nitrosodiphenylamine	10.49	169	217461	8.950	ng	98
66) 4-Bromophenyl-phenylether	10.86	248	75616	9.096	ng	# 89
67) Hexachlorobenzene	10.92	284	87967	9.184	ng	# 89
68) Atrazine	11.01	200	43912	5.766	ng	92
69) Pentachlorophenol	11.11	266	45739	10.988	ng	98
70) Phenanthrene	11.33	178	379190	9.227	ng	98
71) Anthracene	11.39	178	371673	8.901	ng	99
72) Carbazole	11.54	167	343095	9.119	ng	98
73) Di-n-butylphthalate	11.87	149	373088	9.271	ng	99
74) Fluoranthene	12.52	202	390812	9.127	ng	95
76) Benzidine	12.64	184	73982	2.963	ng	# 92
77) Pyrene	12.75	202	411962	8.469	ng	99
79) Butylbenzylphthalate	13.39	149	132757	13.366	ng	99
80) Benzo(a)anthracene	13.97	228	347655	8.572	ng	99
81) 3,3'-Dichlorobenzidine	13.94	252	103524	6.790	ng	# 95
82) Chrysene	14.01	228	358716	8.588	ng	98
83) Bis(2-ethylhexyl)phthalate	13.97	149	197280	9.810	ng	98
84) Di-n-octyl phthalate	14.65	149	262459	12.692	ng	# 95
85) Indeno(1,2,3-cd)pyrene	16.94	276	256135	7.738	ng	94
87) Benzo(b)fluoranthene	15.08	252	320312	8.544	ng	98
88) Benzo(k)fluoranthene	15.08	252	320312	8.811	ng	98
89) Benzo(a)pyrene	15.44	252	272353	7.988	ng	97
90) Dibenzo(a,h)anthracene	16.96	278	211792	7.997	ng	95
91) Benzo(g,h,i)perylene	17.37	276	204959	7.911	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.79 min Scan# 800
Delta R.T. -0.00 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Instrument :
BNA_F
ClientSampleId :




Tot Ion:152 Resp: 202785
Ion Ratio Lower Upper
152 100
150 155.0 124.6 186.8
115 60.6 50.1 75.1




#2

1,4-Dioxane
Concen: 7.895 ng
RT: 2.52 min Scan# 73
Delta R.T. -0.00 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion: 88 Resp: 51378
Ion Ratio Lower Upper
88 100
58 87.2 62.6 93.8
43 33.4 24.6 37.0





#3

Pvridine

Concen: 7.751 ng

RT: 3.27 min Scan# 202

Delta R.T. 0.03 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 79 Resp: 135278

Ion Ratio Lower Upper

79 100

52 87.6 59.8 89.6

51 44.2 29.8 44.8



#4

n-Nitrosodimethylamine

Concen: 6.929 ng

RT: 3.18 min Scan# 186

Delta R.T. -0.02 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31




Tgt Ion: 42 Resp: 61296


Ion Ratio Lower Upper

42 100

74 109.1 104.9 157.3

44 9.3 6.9 10.3





#5

2-Fluorophenol

Concen: 123.317 ng

RT: 5.40 min Scan# 563

Delta R.T. -0.00 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :


Tot Ion:112 Resp: 1517931

Ion Ratio Lower Upper

112 100

64 66.6 47.9 71.9




63 35.4 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1



#6

Aniline

Concen: 8.271 ng

RT: 6.46 min Scan# 743

Delta R.T. -0.00 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31


Tgt Ion: 93 Resp: 177043

Ion Ratio Lower Upper

93 100

66 43.8 32.2 48.2



65 23.0 16.4 24.6




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 66.00 (65.70 to 66.70): BF1

Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 124.340 ng

RT: 6.42 min Scan# 737

Delta R.T. -0.00 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31





Instrument :

BNA_F

ClientSampleId :

Tot Ion: 99 Resp: 1822392

Ion	Ratio	Lower	Upper
99	100		
42	22.3	14.5	21.7#
71	35.9	25.4	38.0



#8

2-Chlorophenol

Concen: 8.426 ng

RT: 6.57 min Scan# 763




Delta R.T. -0.01 min


Lab File: BF105169.D

Acq: 9 May 2018 2:31

Tgt Ion: 128 Resp: 110024

Ion	Ratio	Lower	Upper
128	100		
130	31.7	13.0	53.0
64	53.1	29.4	69.4





#9

Benzaldehyde

Concen: 5.019 ng

RT: 6.35 min Scan# 724

Delta R.T. -0.00 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampled :


Tot Ion: 77 Resp: 56383

Ion Ratio Lower Upper

77 100

105 103.2 81.1 121.1




106 97.7 74.5 114.5



Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 106.00 (105.70 to 106.70): BF1



#10

Phenol

Concen: 7.826 ng

RT: 6.43 min Scan# 739

Delta R.T. -0.00 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31


Tgt Ion: 94 Resp: 127650

Ion Ratio Lower Upper

94 100

65 26.9 7.9 47.9



66 36.7 16.2 56.2




Abundance Ion 94.00 (93.70 to 94.70): BF1

Ion 65.00 (64.70 to 65.70): BF1

Ion 66.00 (65.70 to 66.70): BF1





#11

bis(2-Chloroethyl)ether

Concen: 8.535 ng

RT: 6.53 min Scan# 755

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 93 Resp: 114737

Ion Ratio Lower Upper

93 100

63 88.1 58.6 98.6




95 31.7 11.8 51.8



Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 63.00 (62.70 to 63.70): BF1

Ion 95.00 (94.70 to 95.70): BF1



#12

1,3-Dichlorobenzene

Concen: 9.078 ng

RT: 6.73 min Scan# 790

Delta R.T. -0.00 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31


Tgt Ion: 146 Resp: 142876

Ion Ratio Lower Upper

146 100

148 63.4 51.8 77.8



75 28.1 24.2 36.4




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 75.00 (74.70 to 75.70): BF1









#13
 1,4-Dichlorobenzene
 Concen: 8.891 ng
 RT: 6.81 min Scan# 803
 Delta R.T. -0.00 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
146	100		
148	65.6	50.8	76.2
111	41.0	34.2	51.2



#14
 1,2-Dichlorobenzene
 Concen: 8.870 ng
 RT: 6.96 min Scan# 829
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Tgt Ion	Ratio	Lower	Upper
146	100		
148	66.0	50.6	75.8
111	43.4	36.0	54.0





#15

Benzyl Alcohol

Concen: 9.814 ng

RT: 6.93 min Scan# 823

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 79 Resp: 110084

Ion Ratio Lower Upper

79 100

108 80.6 65.1 97.7

77 61.7 50.6 75.8



#16

2,2'-oxybis(1-Chloropropane)

Concen: 8.242 ng

RT: 7.07 min Scan# 847

Delta R.T. -0.00 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31




Tgt Ion: 45 Resp: 205886


Ion Ratio Lower Upper

45 100

77 11.5 0.0 32.9

79 8.6 0.0 29.6









#17
2-Methylphenol
Concen: 8.432 ng
RT: 7.03 min Scan# 841
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Instrument :
BNA_F
ClientSampleId :

Tot Ion:107 Resp: 95060
Ion Ratio Lower Upper
107 100
108 113.4 91.7 137.5
77 44.9 33.8 50.8
79 42.7 33.8 50.8




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BF1
Ion 79.00 (78.70 to 79.70): BF1








#18
Hexachloroethane
Concen: 8.968 ng
RT: 7.30 min Scan# 887
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion:117 Resp: 45419
Ion Ratio Lower Upper
117 100
119 96.2 75.8 113.8
201 101.2 75.7 113.5



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E









#19
n-Nitroso-di-n-propylamine
Concen: 8.911 ng
RT: 7.20 min Scan# 869
Delta R.T. -0.02 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31


Instrument :
BNA_F
ClientSampled :


Tot Ion	Ratio	Lower	Upper
70	100		
42	71.0	47.5	71.3
101	8.7	7.4	11.2
130	18.4	16.6	25.0



#20
3+4-Methylphenols
Concen: 9.656 ng
RT: 7.19 min Scan# 867
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion	Ratio	Lower	Upper
107	100		
108	90.2	68.2	108.2
77	52.5	26.7	66.7
79	27.0	6.3	46.3









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.07 min Scan# 1018
Delta R.T. -0.00 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Instrument :
BNA_F
ClientSampled :

Tot Ion:136	Resp:	810090
Ion Ratio	Lower	Upper
136	100	
137	10.8	8.9 13.3
54	9.1	6.6 9.8
68	6.1	5.0 7.4




Abundance Ion 136.00 (135.70 to 136.70): E
1500000
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1






#22
Acetophenone
Concen: 9.524 ng
RT: 7.20 min Scan# 869
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion:105	Resp:	179836
Ion Ratio	Lower	Upper
105	100	
71	7.1	4.4 6.6#
51	31.3	22.3 33.5
120	21.2	16.9 25.3



Abundance Ion 105.00 (104.70 to 105.70): E
300000
Ion 71.00 (70.70 to 71.70): BF1
Ion 51.00 (50.70 to 51.70): BF1
Ion 120.00 (119.70 to 120.70): E









#23
Nitrobenzene-d5
Concen: 94.952 ng
RT: 7.36 min Scan# 896
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Instrument :
BNA_F
ClientSampleId :

Tot Ion: 82 Resp: 1134953
Ion Ratio Lower Upper
82 100
128 43.5 36.1 54.1
54 59.5 41.4 62.2




Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 128.00 (127.70 to 128.70): BF1
Ion 54.00 (53.70 to 54.70): BF1






#24
Nitrobenzene
Concen: 9.634 ng
RT: 7.37 min Scan# 899
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion: 77 Resp: 130570
Ion Ratio Lower Upper
77 100
123 46.8 37.9 56.9
65 14.0 11.0 16.4



Abundance Ion 77.00 (76.70 to 77.70): BF1
Ion 123.00 (122.70 to 123.70): BF1
Ion 65.00 (64.70 to 65.70): BF1





#25

Isophorone

Concen: 7.982 ng

RT: 7.60 min Scan# 938

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 82 Resp: 213364

Ion Ratio Lower Upper




82 100

95 6.4 5.2 7.8

138 15.1 14.9 22.3



Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 95.00 (94.70 to 95.70): BF1
Ion 138.00 (137.70 to 138.70): BF1



#26

2-Nitrophenol

Concen: 16.746 ng

RT: 7.69 min Scan# 952

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31


Tgt Ion: 139 Resp: 46617

Ion Ratio Lower Upper



139 100


109 31.2 22.7 34.1

65 54.0 40.5 60.7



Abundance Ion 139.00 (138.70 to 139.70): BF1
Ion 109.00 (108.70 to 109.70): BF1
Ion 65.00 (64.70 to 65.70): BF1





#27

2,4-Dimethylphenol

Concen: 7.248 ng

RT: 7.72 min Scan# 957

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :


Tot Ion:122 Resp: 86103

Ion Ratio Lower Upper

122 100

107 113.9 91.2 136.8




121 57.6 47.2 70.8



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 8.250 ng

RT: 7.82 min Scan# 975

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31


Tgt Ion: 93 Resp: 136107

Ion Ratio Lower Upper

93 100

95 31.7 26.3 39.5



123 10.9 9.8 14.6




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 123.00 (122.70 to 123.70): E









#29
 2,4-Dichlorophenol
 Concen: 8.424 ng
 RT: 7.92 min Scan# 992
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Instrument :
 BNA_F
 ClientSampled :

Tot Ion	Ratio	Lower	Upper
162	100		
164	67.6	42.7	82.7
98	34.7	18.1	58.1




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF1






#30
 1,2,4-Trichlorobenzene
 Concen: 9.168 ng
 RT: 8.01 min Scan# 1007
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Tgt Ion	Ratio	Lower	Upper
180	100		
182	93.0	76.8	115.2
145	31.0	26.5	39.7



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 9.045 ng

RT: 8.09 min Scan# 1021

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :


Tot Ion:128 Resp: 352800

Ion Ratio Lower Upper

128 100

129 11.0 8.8 13.2




127 13.0 10.9 16.3



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 16.529 ng

RT: 7.79 min Scan# 969

Delta R.T. -0.05 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31


Tgt Ion:122 Resp: 54759

Ion Ratio Lower Upper

122 100

105 117.3 101.1 141.1



77 94.4 70.7 110.7




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF1









#33
4-Chloroaniline
Concen: 8.491 ng
RT: 8.14 min Scan# 1029
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31




Instrument :
BNA_F
ClientSampled :


Tot Ion:	127	Resp:	137729
Ion	Ratio	Lower	Upper
127	100		
129	32.5	25.9	38.9
65	34.9	25.0	37.4
92	19.8	15.2	22.8



#34
Hexachlorobutadiene
Concen: 9.367 ng
RT: 8.21 min Scan# 1041
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion:	225	Resp:	69550
Ion	Ratio	Lower	Upper
225	100		
223	60.6	50.6	76.0
227	63.5	51.9	77.9









#35
 Caprolactam
 Concen: 7.721 ng
 RT: 8.47 min Scan# 1086
 Delta R.T. -0.05 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:113 Resp: 26059
 Ion Ratio Lower Upper
 113 100
 55 219.5 163.3 203.3#
 56 164.4 115.7 155.7#



#36
 4-Chloro-3-methylphenol
 Concen: 8.330 ng
 RT: 8.61 min Scan# 1109
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Tgt Ion:107 Resp: 96380
 Ion Ratio Lower Upper
 107 100
 144 26.5 20.5 30.7
 142 82.2 62.0 93.0





#37

2-Methylnaphthalene

Concen: 8.967 ng

RT: 8.78 min Scan# 1138

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :


Tot Ion:142 Resp: 243780

Ion Ratio Lower Upper




142 100

141 85.4 70.8 106.2

115 32.9 26.1 39.1



Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 115.00 (114.70 to 115.70): E



#38

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.83 min Scan# 1316

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31


Tgt Ion:164 Resp: 400637

Ion Ratio Lower Upper



164 100


162 102.0 86.1 129.1

160 46.2 38.3 57.5



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E





#39

1,2,4,5-Tetrachlorobenzene

Concen: 9.389 ng

RT: 8.94 min Scan# 1166

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :

Tot Ion:216 Resp: 119698


Ion Ratio Lower Upper

216 100

214 76.8 63.0 94.4

179 20.6 18.1 27.1

108 17.7 17.2 25.8






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 8.152 ng

RT: 8.93 min Scan# 1164

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31


Tgt Ion:237 Resp: 56198

Ion Ratio Lower Upper

237 100

235 62.0 44.8 84.8



272 11.3 0.0 33.3




Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 131.293 ng
 RT: 10.62 min Scan# 1450
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
330	100		
332	95.7	77.0	115.6
141	26.6	29.9	44.9#




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#42
 2,4,6-Trichlorophenol
 Concen: 8.402 ng
 RT: 9.06 min Scan# 1185
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Tgt Ion	Ratio	Lower	Upper
196	100		
198	104.4	75.7	113.5
200	36.3	24.5	36.7



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E









#43
 2,4,5-Trichlorophenol
 Concen: 9.050 ng
 RT: 9.09 min Scan# 1191
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:196 Resp: 73550
 Ion Ratio Lower Upper
 196 100
 198 97.0 74.6 112.0
 97 42.3 42.9 64.3#
 132 26.5 26.3 39.5




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF1
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 78.025 ng
 RT: 9.15 min Scan# 1201
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Tgt Ion:172 Resp: 2014020
 Ion Ratio Lower Upper
 172 100
 171 36.8 29.7 44.5
 170 24.5 19.6 29.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 9.723 ng
 RT: 9.24 min Scan# 1217
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
154	100		
153	40.6	21.3	61.3
76	13.7	0.0	34.0




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF1






#46
 2-Chloronaphthalene
 Concen: 9.421 ng
 RT: 9.27 min Scan# 1222
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Tgt Ion	Ratio	Lower	Upper
162	100		
127	37.5	33.9	50.9
164	34.1	26.5	39.7



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 13.460 ng
RT: 9.36 min Scan# 1237
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31




Instrument :
BNA_F
ClientSampled :


Tot Ion: 65 Resp: 61234
Ion Ratio Lower Upper
65 100
92 63.5 55.8 83.6
138 89.4 91.2 136.8#



#48
Acenaphthylene
Concen: 9.516 ng
RT: 9.69 min Scan# 1292
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion: 152 Resp: 358578
Ion Ratio Lower Upper
152 100
151 20.1 16.3 24.5
153 13.4 10.7 16.1





#49

Dimethylphthalate

Concen: 9.658 ng

RT: 9.54 min Scan# 1268

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampled :


Tot Ion:163 Resp: 261960

Ion Ratio Lower Upper

163 100

194 3.7 2.7 4.1




164 10.3 8.2 12.2



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 9.189 ng

RT: 9.60 min Scan# 1278

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31


Tgt Ion:165 Resp: 54080

Ion Ratio Lower Upper

165 100

63 60.0 44.7 67.1



89 54.7 44.0 66.0




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1





#51

Acenaphthene

Concen: 9.530 ng

RT: 9.86 min Scan# 1321

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :


Tot Ion:154 Resp: 231414

Ion Ratio Lower Upper

154 100

153 114.4 91.2 136.8




152 51.3 43.8 65.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 9.132 ng

RT: 9.77 min Scan# 1307

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31


Tgt Ion:138 Resp: 56610

Ion Ratio Lower Upper

138 100

108 11.2 9.4 14.0



92 105.7 89.4 134.2




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF1





#53

2,4-Dinitrophenol

Concen: 16.794 ng

RT: 9.87 min Scan# 1324

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :


Tot Ion:184 Resp: 12330

Ion Ratio Lower Upper

184 100

63 66.4 54.2 81.2




154 66.2 184.0 276.0#



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 154.00 (153.70 to 154.70): E



#54

Dibenzofuran

Concen: 8.991 ng

RT: 10.03 min Scan# 1350

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31


Tgt Ion:168 Resp: 319381

Ion Ratio Lower Upper

168 100

139 37.0 30.1 45.1



169 13.1 10.9 16.3




Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E









#55
4-Nitrophenol
Concen: 10.865 ng
RT: 9.92 min Scan# 1331
Delta R.T. -0.02 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31




Instrument :
BNA_F
ClientSampleId :


Tot Ion:139 Resp: 38705
Ion Ratio Lower Upper
139 100
109 72.7 48.4 88.4
65 103.9 72.6 112.6



#56
2,4-Dinitrotoluene
Concen: 11.522 ng
RT: 10.00 min Scan# 1346
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion:165 Resp: 66954
Ion Ratio Lower Upper
165 100
63 48.5 36.4 54.6
89 68.0 57.8 86.6
182 2.1 1.6 2.4





#57

Fluorene

Concen: 9.593 ng

RT: 10.37 min Scan# 1408

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :


Tot Ion:166 Resp: 249911

Ion Ratio Lower Upper




166 100

165 99.2 79.8 119.8

167 13.8 10.6 16.0



Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 8.134 ng

RT: 10.14 min Scan# 1370

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Tgt Ion:232 Resp: 59013


Ion Ratio Lower Upper

232 100



131 43.1 43.8 65.8#


130 2.2 2.1 3.1

166 31.2 27.8 41.6



Abundance Ion 232.00 (231.70 to 232.70): E
Ion 131.00 (130.70 to 131.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 166.00 (165.70 to 166.70): E





#59

Diethylphthalate

Concen: 9.361 ng

RT: 10.25 min Scan# 1388

Delta R.T. -0.01 min

Lab File: BF105169.D


Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :




Tot Ion:	149	Resp:	249507
Ion	Ratio	Lower	Upper
149	100		
177	22.5	16.1	24.1
150	12.0	10.1	15.1



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 10.108 ng


RT: 10.37 min Scan# 1408

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31



Tgt Ion:	204	Resp:	122584
Ion	Ratio	Lower	Upper
204	100		
206	32.5	26.5	39.7
141	55.5	54.6	81.8




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E









#61
4-Nitroaniline
Concen: 8.472 ng
RT: 10.38 min Scan# 1410
Delta R.T. -0.02 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31




Instrument :
BNA_F
ClientSampled :


Tot Ion:138 Resp: 52257
Ion Ratio Lower Upper
138 100
92 52.4 30.2 70.2
108 74.2 52.5 92.5



#62
Azobenzene
Concen: 8.777 ng
RT: 10.53 min Scan# 1435
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion: 77 Resp: 244060
Ion Ratio Lower Upper
77 100
182 23.0 1.7 41.7
105 22.7 3.0 43.0
51 34.6 9.9 49.9









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.31 min Scan# 1568
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Instrument :
BNA_F
ClientSampled :

Tot Ion:188 Resp: 779463
Ion Ratio Lower Upper
188 100
94 10.0 8.5 12.7
80 11.2 9.2 13.8




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF1
Ion 80.00 (79.70 to 80.70): BF1






#64
4,6-Dinitro-2-methylphenol
Concen: 17.325 ng
RT: 10.41 min Scan# 1415
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion:198 Resp: 25225
Ion Ratio Lower Upper
198 100
51 55.2 37.7 77.7
105 43.9 36.3 76.3



Abundance Ion 198.00 (197.70 to 198.70): E
Ion 51.00 (50.70 to 51.70): BF1
Ion 105.00 (104.70 to 105.70): E









#65
 n-Nitrosodiphenylamine
 Concen: 8.950 ng
 RT: 10.49 min Scan# 1428
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:169 Resp: 217461
 Ion Ratio Lower Upper
 169 100
 168 70.4 54.4 81.6
 167 38.0 29.8 44.6




Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E






#66
 4-Bromophenyl-phenylether
 Concen: 9.096 ng
 RT: 10.86 min Scan# 1491
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Tgt Ion:248 Resp: 75616
 Ion Ratio Lower Upper
 248 100
 250 95.7 76.9 115.3
 141 71.6 74.4 111.6#



Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 9.184 ng
RT: 10.92 min Scan# 1501
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Instrument :
BNA_F
ClientSampled :

Tot Ion:284 Resp: 87967
Ion Ratio Lower Upper
284 100
142 33.3 34.6 52.0#
249 28.7 24.8 37.2




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 5.766 ng
RT: 11.01 min Scan# 1517
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion:200 Resp: 43912
Ion Ratio Lower Upper
200 100
173 24.4 8.6 48.6
215 50.7 25.1 65.1



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 10.988 ng
 RT: 11.11 min Scan# 1534
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:266 Resp: 45739
 Ion Ratio Lower Upper
 266 100
 268 62.0 51.1 76.7
 264 62.5 49.5 74.3




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 9.227 ng
 RT: 11.33 min Scan# 1572
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Tgt Ion:178 Resp: 379190
 Ion Ratio Lower Upper
 178 100
 176 19.4 15.8 23.8
 179 15.0 13.0 19.6



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 8.901 ng
 RT: 11.39 min Scan# 1581
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:178 Resp: 371673
 Ion Ratio Lower Upper
 178 100
 176 19.3 15.4 23.2
 179 15.3 12.6 19.0



#72
 Carbazole
 Concen: 9.119 ng
 RT: 11.54 min Scan# 1607
 Delta R.T. -0.01 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Tgt Ion:167 Resp: 343095
 Ion Ratio Lower Upper
 167 100
 166 22.7 17.6 26.4
 139 12.3 10.9 16.3





#73

Di-n-butylphthalate

Concen: 9.271 ng

RT: 11.87 min Scan# 1664

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :


Tot Ion:149 Resp: 373088

Ion Ratio Lower Upper

149 100

150 9.5 7.8 11.6




104 5.0 3.8 5.8



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 9.127 ng

RT: 12.52 min Scan# 1774

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31


Tgt Ion:202 Resp: 390812

Ion Ratio Lower Upper

202 100

101 11.1 0.0 34.1



203 16.9 0.0 38.1




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.000 ng

RT: 13.99 min Scan# 2023

Delta R.T. -0.03 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :


Tot Ion:240 Resp: 668747

Ion Ratio Lower Upper

240 100

120 12.1 9.5 14.3




236 26.8 20.7 31.1



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 2.963 ng

RT: 12.64 min Scan# 1795

Delta R.T. -0.01 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31


Tgt Ion:184 Resp: 73982

Ion Ratio Lower Upper

184 100

185 21.4 12.6 18.8#



183 11.8 9.3 13.9




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#77
Pvrene
Concen: 8.469 ng
RT: 12.75 min Scan# 1813
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31




Instrument :
BNA_F
ClientSampleId :


Tot Ion:202 Resp: 411962
Ion Ratio Lower Upper
202 100
200 21.2 17.4 26.2
203 17.4 14.3 21.5



#78
Terphenyl-d14
Concen: 75.998 ng
RT: 12.90 min Scan# 1839
Delta R.T. -0.01 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion:244 Resp: 2187975
Ion Ratio Lower Upper
244 100
212 7.4 5.4 8.0
122 10.8 11.0 16.4#









#79
Butylbenzylphthalate
Concen: 13.366 ng
RT: 13.39 min Scan# 1921
Delta R.T. -0.02 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31




Instrument :
BNA_F
ClientSampled :


Tot Ion	Ratio	Lower	Upper
149	100		
91	71.4	56.8	85.2
206	19.8	14.1	21.1



#80
Benzo(a)anthracene
Concen: 8.572 ng
RT: 13.97 min Scan# 2021
Delta R.T. -0.02 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.9	22.6	33.8
229	19.9	15.8	23.6









#81
 3,3'-Dichlorobenzidine
 Concen: 6.790 ng
 RT: 13.94 min Scan# 2015
 Delta R.T. -0.03 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
252	100		
254	65.9	50.4	75.6
126	10.1	11.3	16.9#



#82
 Chrysene
 Concen: 8.588 ng
 RT: 14.01 min Scan# 2027
 Delta R.T. -0.03 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.9	25.0	37.6
229	19.3	16.2	24.4









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 9.810 ng
 RT: 13.97 min Scan# 2021
 Delta R.T. -0.02 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:149 Resp: 197280
 Ion Ratio Lower Upper
 149 100
 167 25.6 21.3 31.9
 279 3.2 3.0 4.4



#84
 Di-n-octyl phthalate
 Concen: 12.692 ng
 RT: 14.65 min Scan# 2136
 Delta R.T. -0.04 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Tgt Ion:149 Resp: 262459
 Ion Ratio Lower Upper
 149 100
 167 1.6 1.2 1.8
 43 12.7 8.4 12.6#









#85
Indeno(1,2,3-cd)pyrene
Concen: 7.738 ng
RT: 16.94 min Scan# 2525
Delta R.T. -0.07 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
276	100		
138	25.4	25.0	37.6
277	25.6	20.1	30.1



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.51 min Scan# 2282
Delta R.T. -0.05 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.3	19.2	28.8
265	22.2	17.5	26.3









#87
Benzo(b)fluoranthene
Concen: 8.544 ng
RT: 15.08 min Scan# 2209
Delta R.T. -0.05 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31




Instrument :
BNA_F
ClientSampleId :


Tot Ion:252 Resp: 320312
Ion Ratio Lower Upper
252 100
253 22.1 17.0 25.6
125 10.5 9.0 13.6



#88
Benzo(k)fluoranthene
Concen: 8.811 ng
RT: 15.08 min Scan# 2209
Delta R.T. -0.09 min
Lab File: BF105169.D
Acq: 9 May 2018 2:31

Tgt Ion:252 Resp: 320312
Ion Ratio Lower Upper
252 100
253 22.1 17.9 26.9
125 10.5 10.2 15.2





#89

Benzo(a)pyrene

Concen: 7.988 ng

RT: 15.44 min Scan# 2271

Delta R.T. -0.05 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31

Instrument :

BNA_F

ClientSampleId :





Tot Ion:252 Resp: 272353

Ion Ratio Lower Upper

252 100

253 21.9 17.1 25.7

125 10.1 9.8 14.6



#90

Dibenzo(a,h)anthracene

Concen: 7.997 ng

RT: 16.96 min Scan# 2529

Delta R.T. -0.07 min

Lab File: BF105169.D

Acq: 9 May 2018 2:31




Tgt Ion:278 Resp: 211792


Ion Ratio Lower Upper

278 100

139 15.9 15.9 23.9

279 25.0 19.0 28.4





#91
 Benzo(a,h,i)perylene
 Concen: 7.911 ng
 RT: 17.37 min Scan# 2598
 Delta R.T. -0.08 min
 Lab File: BF105169.D
 Acq: 9 May 2018 2:31

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
276	100		
277	24.9	17.9	26.9
138	23.5	21.8	32.8

