

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051119\
 Data File : BF114172.D
 Acq On : 12 May 2019 00:00
 Operator : JU/SJ
 Sample : K2764-16
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P026-SS002-0002-01

Manual Integrations
 APPROVED

Sohil
 5/13/2019 8:36:01 PM

Quant Time: May 13 08:30:35 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 10 15:56:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	305239	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	1126058	20.00	ng	0.00
39) Acenaphthene-d10	9.88	164	517396	20.00	ng	0.00
64) Phenanthrene-d10	11.37	188	911304	20.00	ng	0.00
76) Chrysene-d12	14.02	240	513085	20.00	ng	0.00
87) Perylene-d12	15.49	264	716410	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	1659438	87.49	ng	0.00
7) Phenol-d6	6.49	99	2238101	99.77	ng	0.00
23) Nitrobenzene-d5	7.41	82	1392602	69.40	ng	0.00
42) 2,4,6-Tribromophenol	10.67	330	462487	86.46	ng	0.00
45) 2-Fluorobiphenyl	9.20	172	1963701	57.41	ng	0.00
79) Terphenyl-d14	12.95	244	1257311	50.52	ng	0.00
Target Compounds						
31) Naphthalene	8.15	128	651312	12.382	ng	99
32) Benzoic acid	7.85	122	16439	8.285	ng	90
37) 2-Methylnaphthalene	8.84	142	256228	7.550	ng	98
38) 1-Methylnaphthalene	8.94	142	168902	5.285	ng	98
46) 1,1'-Biphenyl	9.30	154	99161	2.372	ng	95
49) Acenaphthylene	9.75	152	1057412	20.668	ng	98
50) Dimethylphthalate	9.59	163	239056	6.294	ng	100
58) Fluorene	10.43	166	157036m	4.416	ng	
71) Phenanthrene	11.40	178	3856655	86.987	ng	97
72) Anthracene	11.45	178	701366	15.750	ng	99
75) Fluoranthene	12.60	202	4362850	91.380	ng	97
78) Pvrene	12.83	202	6759725	163.772	ng	93
81) Benzo(a)anthracene	14.01	228	3037263	91.522	ng	# 85
83) Chrysene	14.04	228	3353171	103.074	ng	97
86) Indeno(1,2,3-cd)pyrene	16.96	276	2008669	69.864	ng	97
88) Benzo(b)fluoranthene	15.07	252	3932920m	85.690	ng	
89) Benzo(k)fluoranthene	15.09	252	1156611m	27.433	ng	
90) Benzo(a)pyrene	15.43	252	2659590	65.842	ng	99
91) Dibenzo(a,h)anthracene	16.96	278	565104	16.836	ng	97
92) Benzo(a,h,i)perylene	17.42	276	2182668	64.888	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051119\
Data File : BF114172.D
Acq On : 12 May 2019 00:00
Operator : JU/SJ
Sample : K2764-16
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P026-SS002-0002-01

Manual Integrations
APPROVED

Sohil
5/13/2019 8:36:01 PM

Quant Time: May 13 08:30:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 10 15:56:46 2019
Response via : Initial Calibration

