

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051119\
 Data File : BF114149.D
 Acq On : 11 May 2019 11:30
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 13 06:47:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 10 15:56:46 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.683	0.604	11.6	99	0.00
3	Pyridine	1.882	1.726	8.3	106	0.00
4	n-Nitrosodimethylamine	0.908	0.802	11.7	100	0.00
5 S	2-Fluorophenol	1.243	1.202	3.3	108	0.00
6	Aniline	2.123	2.149	-1.2	113	0.00
7 S	Phenol-d6	1.470	1.516	-3.1	117	0.00
8	2-Chlorophenol	1.327	1.407	-6.0	120	0.00
9	Benzaldehyde	1.029	1.025	0.4	107	0.00
10 C	Phenol	1.729	1.742	-0.8	113	0.00
11	bis(2-Chloroethyl)ether	1.313	1.406	-7.1	121	0.00
12	1,3-Dichlorobenzene	1.489	1.476	0.9	113	0.00
13 C	1,4-Dichlorobenzene	1.501	1.485	1.1	113	0.00
14	1,2-Dichlorobenzene	1.371	1.367	0.3	111	0.00
15	Benzyl Alcohol	1.146	1.151	-0.4	112	0.00
16	2,2'-oxybis(1-Chloropropane	2.545	2.600	-2.2	115	0.00
17	2-Methylphenol	1.090	1.085	0.5	113	0.00
18	Hexachloroethane	0.563	0.568	-0.9	114	0.00
19 P	n-Nitroso-di-n-propylamine	0.932	0.928	0.4	116	0.00
20	3+4-Methylphenols	1.259	1.242	1.4	112	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
22	Acetophenone	0.443	0.417	5.9	112	0.00
23 S	Nitrobenzene-d5	0.356	0.339	4.8	112	0.00
24	Nitrobenzene	0.363	0.352	3.0	114	0.00
25	Isophorone	0.661	0.642	2.9	120	0.00
26 C	2-Nitrophenol	0.176	0.177	-0.6	121	0.00
27	2,4-Dimethylphenol	0.277	0.268	3.2	116	0.00
28	bis(2-Chloroethoxy)methane	0.429	0.443	-3.3	125	0.00
29 C	2,4-Dichlorophenol	0.275	0.278	-1.1	119	0.00
30	1,2,4-Trichlorobenzene	0.313	0.311	0.6	118	0.00
31	Naphthalene	0.934	0.950	-1.7	118	0.00
32	Benzoic acid	0.181	0.179	1.1	119	0.01
33	4-Chloroaniline	0.426	0.436	-2.3	119	0.00
34 C	Hexachlorobutadiene	0.178	0.186	-4.5	122	0.00
35	Caprolactam	0.090	0.100	-11.1	127	0.00
36 C	4-Chloro-3-methylphenol	0.277	0.299	-7.9	125	0.00
37	2-Methylnaphthalene	0.603	0.642	-6.5	123	0.00
38	1-Methylnaphthalene	0.568	0.601	-5.8	121	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	127	0.00
40	1,2,4,5-Tetrachlorobenzene	0.606	0.613	-1.2	123	0.00
41 P	Hexachlorocyclopentadiene	0.255	0.244	4.3	114	0.00
42 S	2,4,6-Tribromophenol	0.207	0.199	3.9	122	0.00
43 C	2,4,6-Trichlorophenol	0.417	0.427	-2.4	124	0.00
44	2,4,5-Trichlorophenol	0.427	0.440	-3.0	125	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051119\
 Data File : BF114149.D
 Acq On : 11 May 2019 11:30
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 13 06:47:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 10 15:56:46 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.322	1.205	8.9	116	0.00
46	1,1'-Biphenyl	1.616	1.583	2.0	121	0.00
47	2-Chloronaphthalene	1.300	1.282	1.4	121	0.00
48	2-Nitroaniline	0.461	0.465	-0.9	124	0.00
49	Acenaphthylene	1.978	1.953	1.3	121	0.00
50	Dimethylphthalate	1.468	1.470	-0.1	124	0.00
51	2,6-Dinitrotoluene	0.326	0.328	-0.6	124	0.00
52 C	Acenaphthene	1.214	1.240	-2.1	127	0.00
53	3-Nitroaniline	0.404	0.409	-1.2	125	0.00
54 P	2,4-Dinitrophenol	0.131	0.149	-13.7	135	0.00
55	Dibenzofuran	1.780	1.711	3.9	120	0.00
56 P	4-Nitrophenol	0.286	0.277	3.1	124	0.00
57	2,4-Dinitrotoluene	0.442	0.438	0.9	125	0.00
58	Fluorene	1.375	1.288	6.3	118	0.00
59	2,3,4,6-Tetrachlorophenol	0.369	0.361	2.2	123	0.00
60	Diethylphthalate	1.492	1.437	3.7	123	0.00
61	4-Chlorophenyl-phenylether	0.675	0.649	3.9	121	0.00
62	4-Nitroaniline	0.425	0.405	4.7	123	0.00
63	Azobenzene	1.444	1.413	2.1	124	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	121	0.00
65	4,6-Dinitro-2-methylphenol	0.096	0.107	-11.5	131	0.00
66 c	n-Nitrosodiphenylamine	0.607	0.628	-3.5	121	0.00
67	4-Bromophenyl-phenylether	0.220	0.226	-2.7	122	0.00
68	Hexachlorobenzene	0.244	0.249	-2.0	122	0.00
69	Atrazine	0.189	0.201	-6.3	128	0.00
70 C	Pentachlorophenol	0.115	0.123	-7.0	123	0.00
71	Phenanthrene	0.973	0.972	0.1	118	0.00
72	Anthracene	0.977	0.988	-1.1	118	0.00
73	Carbazole	0.932	0.934	-0.2	118	0.00
74	Di-n-butylphthalate	1.074	1.133	-5.5	123	0.00
75 C	Fluoranthene	1.048	1.027	2.0	116	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	117	-0.01
77	Benzidine	0.898	0.906	-0.9	121	0.00
78	Pyrene	1.609	1.542	4.2	117	0.00
79 S	Terphenyl-d14	0.970	0.912	6.0	115	0.00
80	Butylbenzylphthalate	0.677	0.738	-9.0	128	0.00
81	Benzo(a)anthracene	1.294	1.298	-0.3	115	-0.01
82	3,3'-Dichlorobenzidine	0.502	0.537	-7.0	118	-0.01
83	Chrysene	1.268	1.247	1.7	113	0.00
84	Bis(2-ethylhexyl)phthalate	0.886	0.961	-8.5	126	0.00
85 c	Di-n-octyl phthalate	1.436	1.581	-10.1	125	-0.04
86	Indeno(1,2,3-cd)pyrene	1.121	1.031	8.0	111	-0.04
87 I	Perylene-d12	1.000	1.000	0.0	123	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051119\
Data File : BF114149.D
Acq On : 11 May 2019 11:30
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 13 06:47:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 10 15:56:46 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.281	1.254	2.1	121	-0.03
89	Benzo(k)fluoranthene	1.177	1.174	0.3	116	-0.03
90 C	Benzo(a)pyrene	1.128	1.120	0.7	120	-0.03
91	Dibenzo(a,h)anthracene	0.937	0.848	9.5	113	-0.03
92	Benzo(q,h,i)perylene	0.939	0.828	11.8	113	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0