


Data Path : Z:\SVOASRV\HPCHEM1\BNA\_F\DATA\BF051119\  
 Data File : BF114161.D  
 Acq On : 11 May 2019 19:04  
 Operator : JU/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF051019.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Fri May 10 15:56:46 2019



AutoFind: Scans 1621, 1622, 1623; Background Corrected with Scan 1614

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 45.9      | 210687  | PASS             |
| 68          | 69           | 0.00         | 2            | 1.5       | 2613    | PASS             |
| 69          | 198          | 0.00         | 100          | 36.8      | 168940  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.5       | 761     | PASS             |
| 127         | 198          | 10           | 80           | 48.3      | 221861  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.8       | 3465    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 458944  | PASS             |
| 199         | 198          | 5            | 9            | 6.7       | 30701   | PASS             |
| 275         | 198          | 10           | 60           | 28.7      | 131829  | PASS             |
| 365         | 198          | 1            | 100          | 3.2       | 14511   | PASS             |
| 441         | 198          | 0.01         | 100          | 15.4      | 70781   | PASS             |
| 442         | 198          | 50           | 100          | 96.5      | 443093  | PASS             |
| 443         | 442          | 15           | 24           | 19.9      | 88072   | PASS             |