

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051219\
 Data File : BF114188.D
 Acq On : 12 May 2019 13:28
 Operator : JU/SJ
 Sample : K2766-05
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P026-SS007-0002-01

Manual Integrations
 APPROVED

Sohil
 5/13/2019 8:37:14 PM

Quant Time: May 13 08:12:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 10 15:56:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	281227	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	1079597	20.00	ng	0.00
39) Acenaphthene-d10	9.89	164	526678	20.00	ng	0.00
64) Phenanthrene-d10	11.37	188	1052271	20.00	ng	0.00
76) Chrysene-d12	14.03	240	562652	20.00	ng	0.00
87) Perylene-d12	15.50	264	753003	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.48	112	1596007	91.33	ng	0.00
7) Phenol-d6	6.49	99	2155994	104.31	ng	0.00
23) Nitrobenzene-d5	7.41	82	1354664	70.42	ng	0.00
42) 2,4,6-Tribromophenol	10.67	330	525598	96.53	ng	0.00
45) 2-Fluorobiphenyl	9.20	172	1983787	56.98	ng	0.00
79) Terphenyl-d14	12.96	244	1721968	63.09	ng	0.00

Target Compounds

						Qvalue
31) Naphthalene	8.15	128	946423	18.767	ng	98
32) Benzoic acid	7.85	122	14502	8.184	ng	93
37) 2-Methylnaphthalene	8.85	142	430481	13.231	ng	97
38) 1-Methylnaphthalene	8.94	142	282962	9.235	ng	99
46) 1,1'-Biphenyl	9.30	154	181237	4.260	ng	98
49) Acenaphthylene	9.75	152	1525901	29.299	ng	99
50) Dimethylphthalate	9.60	163	374592	9.689	ng	99
58) Fluorene	10.43	166	246647	6.814	ng	94
71) Phenanthrene	11.41	178	4863522	95.001	ng	97
72) Anthracene	11.45	178	997292	19.395	ng	98
75) Fluoranthene	12.60	202	5683864	103.100	ng	92
78) Pyrene	12.84	202	8667077	191.484	ng	92
81) Benzo(a)anthracene	14.02	228	3537372	97.202	ng	# 81
83) Chrysene	14.06	228	4092465m	114.717	ng	
86) Indeno(1,2,3-cd)pyrene	17.00	276	2460090	78.028	ng	97
88) Benzo(b)fluoranthene	15.09	252	4588051m	95.106	ng	
89) Benzo(k)fluoranthene	15.10	252	1262807m	28.496	ng	
90) Benzo(a)pyrene	15.45	252	3291018	77.515	ng	97
91) Dibenzo(a,h)anthracene	16.99	278	668985	18.962	ng	97
92) Benzo(a,h,i)perylene	17.45	276	2805758	79.359	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051219\
Data File : BF114188.D
Acq On : 12 May 2019 13:28
Operator : JU/SJ
Sample : K2766-05
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P026-SS007-0002-01

Manual Integrations
APPROVED

Sohil
5/13/2019 8:37:14 PM

Quant Time: May 13 08:12:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 10 15:56:46 2019
Response via : Initial Calibration

