

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051219\
 Data File : BF114200.D
 Acq On : 12 May 2019 18:40
 Operator : JU/SJ
 Sample : K2767-04
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P026-SS010-0612-01

Manual Integrations
 APPROVED

Sohil
 5/13/2019 8:37:33 PM

Quant Time: May 13 09:21:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 10 15:56:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	248026	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	950690	20.00	ng	0.00
39) Acenaphthene-d10	9.89	164	434343	20.00	ng	0.00
64) Phenanthrene-d10	11.37	188	800659	20.00	ng	0.00
76) Chrysene-d12	14.02	240	525264	20.00	ng	0.00
87) Perylene-d12	15.49	264	583121	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	1011413	65.62	ng	0.00
7) Phenol-d6	6.49	99	1435019	78.72	ng	0.00
23) Nitrobenzene-d5	7.41	82	887527	52.39	ng	0.00
42) 2,4,6-Tribromophenol	10.67	330	323468	72.04	ng	0.00
45) 2-Fluorobiphenyl	9.20	172	1343539	46.79	ng	0.00
79) Terphenyl-d14	12.95	244	1087672	42.69	ng	0.00
Target Compounds						
31) Naphthalene	8.16	128	141816	3.193	ng	97
37) 2-Methylnaphthalene	8.85	142	66016	2.304	ng	95
49) Acenaphthylene	9.75	152	241048	5.612	ng	99
50) Dimethylphthalate	9.60	163	298406	9.359	ng	99
71) Phenanthrene	11.40	178	850985	21.846	ng	97
72) Anthracene	11.45	178	139496	3.565	ng	97
75) Fluoranthene	12.59	202	1148668	27.383	ng	98
78) Pyrene	12.82	202	1885565	44.624	ng	100
81) Benzo(a)anthracene	14.00	228	682807m	20.098	ng	
83) Chrysene	14.04	228	870288	26.132	ng	98
86) Indeno(1,2,3-cd)pyrene	16.95	276	398526	13.540	ng	97
88) Benzo(b)fluoranthene	15.06	252	952334m	25.492	ng	
89) Benzo(k)fluoranthene	15.08	252	280357m	8.170	ng	
90) Benzo(a)pyrene	15.42	252	596363	18.139	ng	99
91) Dibenzo(a,h)anthracene	16.96	278	108684m	3.978	ng	
92) Benzo(g,h,i)perylene	17.40	276	441395	16.122	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051219\
Data File : BF114200.D
Acq On : 12 May 2019 18:40
Operator : JU/SJ
Sample : K2767-04
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P026-SS010-0612-01

Manual Integrations
APPROVED

Sohil
5/13/2019 8:37:33 PM

Quant Time: May 13 09:21:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 10 15:56:46 2019
Response via : Initial Calibration

