

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051219\
 Data File : BF114216.D
 Acq On : 13 May 2019 2:32
 Operator : JU/SJ
 Sample : K2765-15DL 2X
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P026-SS005-0612-01DL

Manual Integrations
 APPROVED

Sohil
 5/13/2019 8:37:50 PM

Quant Time: May 13 09:34:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 10 15:56:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	264834	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	1001228	20.00	ng	0.00
39) Acenaphthene-d10	9.89	164	452546	20.00	ng	0.00
64) Phenanthrene-d10	11.37	188	803750	20.00	ng	0.00
76) Chrysene-d12	14.02	240	569660	20.00	ng	0.00
87) Perylene-d12	15.49	264	597410	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	957890	58.21	ng	0.00
7) Phenol-d6	6.49	99	1262815	64.88	ng	0.00
23) Nitrobenzene-d5	7.41	82	755677	42.36	ng	0.00
42) 2,4,6-Tribromophenol	10.67	330	268256	57.34	ng	0.00
45) 2-Fluorobiphenyl	9.20	172	1170672	39.13	ng	0.00
79) Terphenyl-d14	12.95	244	924154	33.44	ng	0.00
Target Compounds						
49) Acenaphthylene	9.75	152	446641	9.981	ng	99
50) Dimethylphthalate	9.60	163	195288	5.878	ng	99
71) Phenanthrene	11.40	178	859209	21.973	ng	97
72) Anthracene	11.44	178	216806	5.520	ng	98
75) Fluoranthene	12.59	202	1603246	38.073	ng	100
78) Pyrene	12.82	202	2660312	58.052	ng	99
81) Benzo(a)anthracene	14.00	228	1268451m	34.426	ng	
83) Chrysene	14.04	228	1510496	41.820	ng	99
86) Indeno(1,2,3-cd)pyrene	16.96	276	667693	20.917	ng	97
88) Benzo(b)fluoranthene	15.07	252	1742227m	45.520	ng	
89) Benzo(k)fluoranthene	15.09	252	331776m	9.437	ng	
90) Benzo(a)pyrene	15.43	252	1061163	31.504	ng	100
91) Dibenzo(a,h)anthracene	16.96	278	169179m	6.044	ng	
92) Benzo(g,h,i)perylene	17.42	276	742819	26.482	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051219\
Data File : BF114216.D
Acq On : 13 May 2019 2:32
Operator : JU/SJ
Sample : K2765-15DL 2X
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P026-SS005-0612-01DL

Manual Integrations
APPROVED

Sohil
5/13/2019 8:37:50 PM

Quant Time: May 13 09:34:05 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 10 15:56:46 2019
Response via : Initial Calibration

