


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF051219\
 Data File : BF114186.D
 Acq On : 12 May 2019 11:45
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri May 10 15:56:46 2019



AutoFind: Scans 1621, 1622, 1623; Background Corrected with Scan 1615

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.9	212134	PASS
68	69	0.00	2	1.7	2928	PASS
69	198	0.00	100	37.9	171470	PASS
70	69	0.00	2	0.3	519	PASS
127	198	10	80	49.4	223317	PASS
197	198	0.00	2	0.8	3483	PASS
198	198	100	100	100.0	451882	PASS
199	198	5	9	6.7	30330	PASS
275	198	10	60	27.3	123501	PASS
365	198	1	100	3.1	14160	PASS
441	198	0.01	100	15.0	67826	PASS
442	198	50	100	90.1	407232	PASS
443	442	15	24	19.8	80525	PASS