

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051519\
 Data File : BF114289.D
 Acq On : 15 May 2019 17:21
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 15 23:27:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 10 15:56:46 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.683	0.623	8.8	75	-0.03
3	Pyridine	1.882	1.752	6.9	80	-0.02
4	n-Nitrosodimethylamine	0.908	0.848	6.6	78	-0.02
5 S	2-Fluorophenol	1.243	1.203	3.2	80	0.00
6	Aniline	2.123	2.046	3.6	80	0.00
7 S	Phenol-d6	1.470	1.508	-2.6	86	0.00
8	2-Chlorophenol	1.327	1.374	-3.5	86	0.00
9	Benzaldehyde	1.029	0.947	8.0	73	0.00
10 C	Phenol	1.729	1.706	1.3	82	0.01
11	bis(2-Chloroethyl)ether	1.313	1.396	-6.3	89	0.00
12	1,3-Dichlorobenzene	1.489	1.503	-0.9	85	0.00
13 C	1,4-Dichlorobenzene	1.501	1.511	-0.7	85	0.00
14	1,2-Dichlorobenzene	1.371	1.398	-2.0	84	0.00
15	Benzyl Alcohol	1.146	1.152	-0.5	83	0.00
16	2,2'-oxybis(1-Chloropropane	2.545	2.532	0.5	83	0.00
17	2-Methylphenol	1.090	1.105	-1.4	85	0.00
18	Hexachloroethane	0.563	0.565	-0.4	83	0.00
19 P	n-Nitroso-di-n-propylamine	0.932	0.924	0.9	85	0.00
20	3+4-Methylphenols	1.259	1.329	-5.6	88	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
22	Acetophenone	0.443	0.451	-1.8	88	0.00
23 S	Nitrobenzene-d5	0.356	0.362	-1.7	87	0.00
24	Nitrobenzene	0.363	0.375	-3.3	88	0.00
25	Isophorone	0.661	0.668	-1.1	91	0.00
26 C	2-Nitrophenol	0.176	0.187	-6.3	93	0.00
27	2,4-Dimethylphenol	0.277	0.255	7.9	80	0.00
28	bis(2-Chloroethoxy)methane	0.429	0.434	-1.2	89	0.00
29 C	2,4-Dichlorophenol	0.275	0.277	-0.7	86	0.00
30	1,2,4-Trichlorobenzene	0.313	0.311	0.6	86	0.00
31	Naphthalene	0.934	0.935	-0.1	85	0.00
32	Benzoic acid	0.181	0.171	5.5	83	0.02
33	4-Chloroaniline	0.426	0.444	-4.2	88	0.00
34 C	Hexachlorobutadiene	0.178	0.173	2.8	83	-0.01
35	Caprolactam	0.090	0.103	-14.4	95	0.00
36 C	4-Chloro-3-methylphenol	0.277	0.293	-5.8	89	0.01
37	2-Methylnaphthalene	0.603	0.622	-3.2	87	0.00
38	1-Methylnaphthalene	0.568	0.582	-2.5	86	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
40	1,2,4,5-Tetrachlorobenzene	0.606	0.611	-0.8	85	0.00
41 P	Hexachlorocyclopentadiene	0.255	0.233	8.6	75	0.00
42 S	2,4,6-Tribromophenol	0.207	0.195	5.8	83	0.00
43 C	2,4,6-Trichlorophenol	0.417	0.427	-2.4	86	0.00
44	2,4,5-Trichlorophenol	0.427	0.437	-2.3	86	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051519\
 Data File : BF114289.D
 Acq On : 15 May 2019 17:21
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 15 23:27:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 10 15:56:46 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.322	1.235	6.6	83	0.00
46	1,1'-Biphenyl	1.616	1.621	-0.3	86	0.00
47	2-Chloronaphthalene	1.300	1.310	-0.8	86	0.00
48	2-Nitroaniline	0.461	0.499	-8.2	93	0.00
49	Acenaphthylene	1.978	1.988	-0.5	85	0.00
50	Dimethylphthalate	1.468	1.520	-3.5	89	0.00
51	2,6-Dinitrotoluene	0.326	0.332	-1.8	88	0.00
52 C	Acenaphthene	1.214	1.176	3.1	84	0.00
53	3-Nitroaniline	0.404	0.429	-6.2	91	0.00
54 P	2,4-Dinitrophenol	0.131	0.149	-13.7	94	0.01
55	Dibenzofuran	1.780	1.704	4.3	83	0.00
56 P	4-Nitrophenol	0.286	0.278	2.8	86	0.02
57	2,4-Dinitrotoluene	0.442	0.432	2.3	86	0.01
58	Fluorene	1.375	1.358	1.2	87	0.00
59	2,3,4,6-Tetrachlorophenol	0.369	0.349	5.4	82	0.01
60	Diethylphthalate	1.492	1.453	2.6	87	0.00
61	4-Chlorophenyl-phenylether	0.675	0.672	0.4	87	0.00
62	4-Nitroaniline	0.425	0.447	-5.2	95	0.01
63	Azobenzene	1.444	1.431	0.9	88	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
65	4,6-Dinitro-2-methylphenol	0.096	0.105	-9.4	94	0.01
66 c	n-Nitrosodiphenylamine	0.607	0.614	-1.2	85	0.00
67	4-Bromophenyl-phenylether	0.220	0.219	0.5	86	0.00
68	Hexachlorobenzene	0.244	0.239	2.0	84	0.00
69	Atrazine	0.189	0.185	2.1	85	0.00
70 C	Pentachlorophenol	0.115	0.122	-6.1	88	0.00
71	Phenanthrene	0.973	0.979	-0.6	86	0.00
72	Anthracene	0.977	0.990	-1.3	85	0.00
73	Carbazole	0.932	0.958	-2.8	87	0.01
74	Di-n-butylphthalate	1.074	1.122	-4.5	88	0.00
75 C	Fluoranthene	1.048	1.045	0.3	85	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
77	Benzidine	0.898	0.618	31.2#	59	0.00
78	Pyrene	1.609	1.588	1.3	85	0.00
79 S	Terphenyl-d14	0.970	0.937	3.4	84	0.00
80	Butylbenzylphthalate	0.677	0.734	-8.4	91	0.00
81	Benzo(a)anthracene	1.294	1.365	-5.5	86	0.00
82	3,3'-Dichlorobenzidine	0.502	0.521	-3.8	82	0.00
83	Chrysene	1.268	1.288	-1.6	83	0.00
84	Bis(2-ethylhexyl)phthalate	0.886	0.976	-10.2	91	0.00
85 c	Di-n-octyl phthalate	1.436	1.583	-10.2	89	-0.03
86	Indeno(1,2,3-cd)pyrene	1.121	1.075	4.1	83	0.00
87 I	Perylene-d12	1.000	1.000	0.0	77	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051519\
Data File : BF114289.D
Acq On : 15 May 2019 17:21
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 15 23:27:19 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 10 15:56:46 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.281	1.371	-7.0	83	-0.01
89	Benzo(k)fluoranthene	1.177	1.314	-11.6	81	-0.01
90 C	Benzo(a)pyrene	1.128	1.222	-8.3	82	-0.01
91	Dibenzo(a,h)anthracene	0.937	0.981	-4.7	82	0.00
92	Benzo(q,h,i)perylene	0.939	1.009	-7.5	86	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0