

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051721\
 Data File : BF123969.D
 Acq On : 17 May 2021 18:17
 Operator : JU/CG
 Sample : M2420-05
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CONCESSION-CP-5

Quant Time: May 17 18:41:31 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 19:05:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.904	152	38401	20.00	ng	0.00
21) Naphthalene-d8	8.186	136	152846	20.00	ng	0.00
39) Acenaphthene-d10	9.939	164	88593	20.00	ng	0.00
64) Phenanthrene-d10	11.427	188	167189	20.00	ng	0.00
76) Chrysene-d12	14.068	240	126864	20.00	ng	0.00
86) Perylene-d12	15.562	264	123747	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	205118	75.00	ng	0.00
7) Phenol-d6	6.522	99	257794	72.66	ng	0.00
23) Nitrobenzene-d5	7.463	82	179847	50.70	ng	0.00
42) 2,4,6-Tribromophenol	10.727	330	84459	79.37	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	349199	48.50	ng	0.00
79) Terphenyl-d14	13.016	244	433159	51.71	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.651	163	41266	5.83	ng	Qvalue 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051721\
Data File : BF123969.D
Acq On : 17 May 2021 18:17
Operator : JU/CG
Sample : M2420-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CONCESSION-CP-5

Quant Time: May 17 18:41:31 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 19:05:57 2021
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.904 min Scan# 819
Delta R.T. 0.000 min
Lab File: BF123969.D
Acq: 17 May 2021 18:17




Instrument :
BNA_F
ClientSampleId :
CONCESSION-CP-5


Tgt Ion:152 Resp: 38401
Ion Ratio Lower Upper
152 100
150 158.2 127.0 190.6
115 61.1 54.3 81.5



#5
2-Fluorophenol
Concen: 75.00 ng
RT: 5.516 min Scan# 583
Delta R.T. 0.006 min
Lab File: BF123969.D
Acq: 17 May 2021 18:17

Tgt Ion:112 Resp: 205118
Ion Ratio Lower Upper
112 100
64 68.7 61.4 92.0
63 39.4 33.0 49.4









#7
Phenol-d6
Concen: 72.66 ng
RT: 6.522 min Scan# 754
Delta R.T. 0.000 min
Lab File: BF123969.D
Acq: 17 May 2021 18:17

Instrument :
BNA_F
ClientSampleId :
CONCESSION-CP-5

Tgt Ion: 99 Resp: 257794




Ion	Ratio	Lower	Upper
99	100		
42	23.3	19.8	29.8
71	41.1	31.1	46.7




#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.186 min Scan# 1037
Delta R.T. 0.000 min
Lab File: BF123969.D
Acq: 17 May 2021 18:17

Tgt Ion: 136 Resp: 152846

Ion	Ratio	Lower	Upper
136	100		
137	9.3	8.7	13.1
54	11.8	9.8	14.6
68	6.0	4.1	6.1









#23
Nitrobenzene-d5
Concen: 50.70 ng
RT: 7.463 min Scan# 914
Delta R.T. 0.000 min
Lab File: BF123969.D
Acq: 17 May 2021 18:17




Instrument :
BNA_F
ClientSampleId :
CONCESSION-CP-5


Tgt Ion: 82 Resp: 179847
Ion Ratio Lower Upper
82 100
128 39.2 29.4 44.2
54 68.3 54.1 81.1



#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.939 min Scan# 1335
Delta R.T. -0.006 min
Lab File: BF123969.D
Acq: 17 May 2021 18:17

Tgt Ion:164 Resp: 88593
Ion Ratio Lower Upper
164 100
162 102.6 80.7 121.1
160 46.1 37.3 55.9









#42
2,4,6-Tribromophenol
Concen: 79.37 ng
RT: 10.727 min Scan# 1469
Delta R.T. 0.000 min
Lab File: BF123969.D
Acq: 17 May 2021 18:17




Instrument :
BNA_F
ClientSampleId :
CONCESSION-CP-5


Tgt Ion: 330 Resp: 84459
Ion Ratio Lower Upper
330 100
332 95.4 77.1 115.7
141 38.6 27.9 41.9



#45
2-Fluorobiphenyl
Concen: 48.50 ng
RT: 9.263 min Scan# 1220
Delta R.T. 0.000 min
Lab File: BF123969.D
Acq: 17 May 2021 18:17

Tgt Ion: 172 Resp: 349199
Ion Ratio Lower Upper
172 100
171 38.1 28.1 42.1
170 24.1 19.2 28.8









#50
Dimethylphthalate
Concen: 5.83 ng
RT: 9.651 min Scan# 1286
Delta R.T. -0.012 min
Lab File: BF123969.D
Acq: 17 May 2021 18:17




Instrument :
BNA_F
ClientSampleId :
CONCESSION-CP-5


Tgt Ion:163 Resp: 41266
Ion Ratio Lower Upper
163 100
194 4.3 3.1 4.7
164 11.6 8.2 12.2



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.427 min Scan# 1588
Delta R.T. 0.000 min
Lab File: BF123969.D
Acq: 17 May 2021 18:17

Tgt Ion:188 Resp: 167189
Ion Ratio Lower Upper
188 100
94 8.5 6.9 10.3
80 9.3 8.2 12.2









#76
Chrysene-d12
Concen: 20.00 ng
RT: 14.068 min Scan# 2037
Delta R.T. 0.000 min
Lab File: BF123969.D
Acq: 17 May 2021 18:17




Instrument :
BNA_F
ClientSampleId :
CONCESSION-CP-5


Tgt Ion:240 Resp: 126864
Ion Ratio Lower Upper
240 100
120 10.7 7.2 10.8
236 24.7 19.0 28.4



#79
Terphenyl-d14
Concen: 51.71 ng
RT: 13.016 min Scan# 1858
Delta R.T. 0.000 min
Lab File: BF123969.D
Acq: 17 May 2021 18:17

Tgt Ion:244 Resp: 433159
Ion Ratio Lower Upper
244 100
212 7.7 5.7 8.5
122 11.2 9.7 14.5





#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 15.562 min Scan# 2291
 Delta R.T. 0.012 min
 Lab File: BF123969.D
 Acq: 17 May 2021 18:17

Instrument :
 BNA_F
 ClientSampleId :
 CONCESSION-CP-5

Tgt Ion:264	Resp:	123747
Ion Ratio	Lower	Upper
264	100	
260	23.2	18.7 28.1
265	21.6	17.8 26.8

