

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051724\
 Data File : BF137924.D
 Acq On : 17 May 2024 11:22
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 17 12:09:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF051424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 15 04:12:16 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2	1,4-Dioxane	40.000	40.143	-0.4	103	0.00
3	Pyridine	40.000	43.218	-8.0	109	0.00
4	n-Nitrosodimethylamine	40.000	40.068	-0.2	104	0.00
5 S	2-Fluorophenol	80.000	78.463	1.9	102	0.00
6	Aniline	40.000	43.053	-7.6	111	0.00
7 S	Phenol-d6	80.000	78.759	1.6	103	0.00
8	2-Chlorophenol	40.000	39.447	1.4	102	0.00
9	Benzaldehyde	40.000	39.602	1.0	101	0.00
10 C	Phenol	40.000	39.410	1.5	103	0.00
11	bis(2-Chloroethyl)ether	40.000	39.313	1.7	102	0.00
12	1,3-Dichlorobenzene	40.000	39.003	2.5	102	0.00
13 C	1,4-Dichlorobenzene	40.000	39.689	0.8	104	0.00
14	1,2-Dichlorobenzene	40.000	38.862	2.8	102	0.00
15	Benzyl Alcohol	40.000	41.834	-4.6	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.158	2.1	101	0.00
17	2-Methylphenol	40.000	40.112	-0.3	104	0.00
18	Hexachloroethane	40.000	39.497	1.3	101	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.221	1.9	102	0.00
20	3+4-Methylphenols	40.000	39.804	0.5	103	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	38.394	4.0	103	0.00
23 S	Nitrobenzene-d5	80.000	78.696	1.6	104	0.00
24	Nitrobenzene	40.000	39.458	1.4	104	0.00
25	Isophorone	40.000	39.074	2.3	103	0.00
26 C	2-Nitrophenol	40.000	39.845	0.4	101	0.00
27	2,4-Dimethylphenol	40.000	39.631	0.9	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.838	2.9	103	0.00
29 C	2,4-Dichlorophenol	40.000	39.443	1.4	102	0.00
30	1,2,4-Trichlorobenzene	40.000	38.920	2.7	103	0.00
31	Naphthalene	40.000	38.861	2.8	103	0.00
32	Benzoic acid	40.000	40.742	-1.9	104	0.00
33	4-Chloroaniline	40.000	42.208	-5.5	109	0.00
34 C	Hexachlorobutadiene	40.000	38.700	3.2	102	0.00
35	Caprolactam	40.000	41.309	-3.3	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.046	-2.6	107	0.00
37	2-Methylnaphthalene	40.000	38.809	3.0	103	0.00
38	1-Methylnaphthalene	40.000	38.874	2.8	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.352	4.1	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.691	-1.7	101	0.00
42 S	2,4,6-Tribromophenol	80.000	82.672	-3.3	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.972	2.6	102	0.00
44	2,4,5-Trichlorophenol	40.000	40.116	-0.3	103	0.00
45 S	2-Fluorobiphenyl	80.000	74.949	6.3	102	0.00
46	1,1'-Biphenyl	40.000	38.341	4.1	103	0.00
47	2-Chloronaphthalene	40.000	38.456	3.9	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051724\
 Data File : BF137924.D
 Acq On : 17 May 2024 11:22
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 17 12:09:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF051424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 15 04:12:16 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.593	-1.5	103	0.00
49	Acenaphthylene	40.000	38.929	2.7	102	0.00
50	Dimethylphthalate	40.000	39.330	1.7	103	0.00
51	2,6-Dinitrotoluene	40.000	40.775	-1.9	105	0.00
52 C	Acenaphthene	40.000	39.655	0.9	105	0.00
53	3-Nitroaniline	40.000	42.076	-5.2	107	0.00
54 P	2,4-Dinitrophenol	40.000	38.840	2.9	101	0.00
55	Dibenzofuran	40.000	38.783	3.0	103	0.00
56 P	4-Nitrophenol	40.000	44.556	-11.4	106	0.00
57	2,4-Dinitrotoluene	40.000	42.181	-5.5	104	0.00
58	Fluorene	40.000	39.056	2.4	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.529	-1.3	102	0.00
60	Diethylphthalate	40.000	40.140	-0.4	103	0.00
61	4-Chlorophenyl-phenylether	40.000	39.302	1.7	103	0.00
62	4-Nitroaniline	40.000	43.645	-9.1	107	0.00
63	Azobenzene	40.000	39.808	0.5	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.839	-2.1	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.764	5.6	103	0.00
67	4-Bromophenyl-phenylether	40.000	38.151	4.6	104	0.00
68	Hexachlorobenzene	40.000	38.116	4.7	105	0.00
69	Atrazine	40.000	39.648	0.9	103	0.00
70 C	Pentachlorophenol	40.000	43.743	-9.4	107	0.00
71	Phenanthrene	40.000	38.655	3.4	106	0.00
72	Anthracene	40.000	38.678	3.3	105	0.00
73	Carbazole	40.000	39.912	0.2	107	0.00
74	Di-n-butylphthalate	40.000	41.031	-2.6	106	0.00
75 C	Fluoranthene	40.000	41.247	-3.1	108	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	114	0.00
77	Benzydine	40.000	58.376	-45.9#	117	0.00
78	Pyrene	40.000	36.684	8.3	110	0.00
79 S	Terphenyl-d14	80.000	72.657	9.2	110	0.00
80	Butylbenzylphthalate	40.000	41.721	-4.3	114	0.00
81	Benzo(a)anthracene	40.000	39.250	1.9	115	0.00
82	3,3'-Dichlorobenzidine	40.000	41.208	-3.0	118	0.00
83	Chrysene	40.000	38.517	3.7	114	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.163	-7.9	118	0.00
85 c	Di-n-octyl phthalate	40.000	44.465	-11.2	122	0.00
86 I	Perylene-d12	20.000	20.000	0.0	115	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.068	9.8	107	0.00
88	Benzo(b)fluoranthene	40.000	42.550	-6.4	118	0.00
89	Benzo(k)fluoranthene	40.000	39.431	1.4	115	0.00
90 C	Benzo(a)pyrene	40.000	39.893	0.3	114	0.00
91	Dibenzo(a,h)anthracene	40.000	35.753	10.6	105	0.01
92	Benzo(g,h,i)perylene	40.000	35.528	11.2	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051724\
Data File : BF137924.D
Acq On : 17 May 2024 11:22
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 17 12:09:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF051424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 15 04:12:16 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0