

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052019\
 Data File : BF114428.D
 Acq On : 20 May 2019 10:26
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 20 10:49:49 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 04:47:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69	-0.01
2	1,4-Dioxane	0.683	0.589	13.8	58	0.01
3	Pyridine	1.882	1.687	10.4	63	0.00
4	n-Nitrosodimethylamine	0.908	0.806	11.2	61	0.00
5 S	2-Fluorophenol	1.243	1.197	3.7	65	0.00
6	Aniline	2.123	2.093	1.4	67	0.00
7 S	Phenol-d6	1.470	1.503	-2.2	70	0.00
8	2-Chlorophenol	1.327	1.375	-3.6	71	-0.01
9	Benzaldehyde	1.029	0.861	16.3	54	-0.01
10 C	Phenol	1.729	1.721	0.5	67	-0.01
11	bis(2-Chloroethyl)ether	1.313	1.363	-3.8	71	0.00
12	1,3-Dichlorobenzene	1.489	1.510	-1.4	70	0.00
13 C	1,4-Dichlorobenzene	1.501	1.524	-1.5	70	-0.01
14	1,2-Dichlorobenzene	1.371	1.401	-2.2	69	0.00
15	Benzyl Alcohol	1.146	1.109	3.2	65	0.00
16	2,2'-oxybis(1-Chloropropane	2.545	2.479	2.6	66	-0.01
17	2-Methylphenol	1.090	1.110	-1.8	70	-0.01
18	Hexachloroethane	0.563	0.567	-0.7	69	0.00
19 P	n-Nitroso-di-n-propylamine	0.932	0.913	2.0	69	-0.01
20	3+4-Methylphenols	1.259	1.342	-6.6	73	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	72	0.00
22	Acetophenone	0.443	0.452	-2.0	72	-0.01
23 S	Nitrobenzene-d5	0.356	0.359	-0.8	71	0.00
24	Nitrobenzene	0.363	0.366	-0.8	70	-0.01
25	Isophorone	0.661	0.650	1.7	72	0.00
26 C	2-Nitrophenol	0.176	0.182	-3.4	74	0.00
27	2,4-Dimethylphenol	0.277	0.272	1.8	70	0.00
28	bis(2-Chloroethoxy)methane	0.429	0.429	0.0	71	0.00
29 C	2,4-Dichlorophenol	0.275	0.279	-1.5	71	0.00
30	1,2,4-Trichlorobenzene	0.313	0.310	1.0	70	0.00
31	Naphthalene	0.934	0.963	-3.1	71	-0.01
32	Benzoic acid	0.181	0.193	-6.6	77	-0.01
33	4-Chloroaniline	0.426	0.442	-3.8	72	0.00
34 C	Hexachlorobutadiene	0.178	0.179	-0.6	70	0.00
35	Caprolactam	0.090	0.099	-10.0	74	-0.02
36 C	4-Chloro-3-methylphenol	0.277	0.295	-6.5	73	0.00
37	2-Methylnaphthalene	0.603	0.623	-3.3	71	0.00
38	1-Methylnaphthalene	0.568	0.589	-3.7	71	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	73	0.00
40	1,2,4,5-Tetrachlorobenzene	0.606	0.616	-1.7	71	0.00
41 P	Hexachlorocyclopentadiene	0.255	0.235	7.8	63	0.00
42 S	2,4,6-Tribromophenol	0.207	0.191	7.7	68	0.00
43 C	2,4,6-Trichlorophenol	0.417	0.412	1.2	69	0.00
44	2,4,5-Trichlorophenol	0.427	0.423	0.9	69	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052019\
 Data File : BF114428.D
 Acq On : 20 May 2019 10:26
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 20 10:49:49 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 04:47:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.322	1.273	3.7	71	0.00
46	1,1'-Biphenyl	1.616	1.618	-0.1	71	0.00
47	2-Chloronaphthalene	1.300	1.297	0.2	70	0.00
48	2-Nitroaniline	0.461	0.474	-2.8	73	0.00
49	Acenaphthylene	1.978	1.975	0.2	70	0.00
50	Dimethylphthalate	1.468	1.474	-0.4	72	0.00
51	2,6-Dinitrotoluene	0.326	0.329	-0.9	72	0.00
52 C	Acenaphthene	1.214	1.175	3.2	69	0.00
53	3-Nitroaniline	0.404	0.421	-4.2	74	0.00
54 P	2,4-Dinitrophenol	0.131	0.129	1.5	67	0.00
55	Dibenzofuran	1.780	1.687	5.2	68	0.00
56 P	4-Nitrophenol	0.286	0.269	5.9	69	0.00
57	2,4-Dinitrotoluene	0.442	0.421	4.8	69	0.00
58	Fluorene	1.375	1.374	0.1	72	0.00
59	2,3,4,6-Tetrachlorophenol	0.369	0.352	4.6	69	0.00
60	Diethylphthalate	1.492	1.451	2.7	72	0.00
61	4-Chlorophenyl-phenylether	0.675	0.675	0.0	72	0.00
62	4-Nitroaniline	0.425	0.439	-3.3	77	0.00
63	Azobenzene	1.444	1.414	2.1	72	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	73	0.00
65	4,6-Dinitro-2-methylphenol	0.096	0.100	-4.2	74	0.00
66 c	n-Nitrosodiphenylamine	0.607	0.612	-0.8	71	0.00
67	4-Bromophenyl-phenylether	0.220	0.214	2.7	70	0.00
68	Hexachlorobenzene	0.244	0.237	2.9	69	0.00
69	Atrazine	0.189	0.185	2.1	71	0.00
70 C	Pentachlorophenol	0.115	0.099	13.9	60	0.00
71	Phenanthrene	0.973	0.987	-1.4	72	0.00
72	Anthracene	0.977	0.999	-2.3	72	0.00
73	Carbazole	0.932	0.969	-4.0	73	0.00
74	Di-n-butylphthalate	1.074	1.107	-3.1	72	0.00
75 C	Fluoranthene	1.048	1.077	-2.8	73	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	73	0.00
77	Benzidine	0.898	0.750	16.5	63	0.00
78	Pyrene	1.609	1.528	5.0	73	0.00
79 S	Terphenyl-d14	0.970	0.905	6.7	72	0.00
80	Butylbenzylphthalate	0.677	0.689	-1.8	75	0.00
81	Benzo(a)anthracene	1.294	1.359	-5.0	76	0.00
82	3,3'-Dichlorobenzidine	0.502	0.528	-5.2	73	0.00
83	Chrysene	1.268	1.297	-2.3	74	0.00
84	Bis(2-ethylhexyl)phthalate	0.886	0.927	-4.6	76	0.00
85 c	Di-n-octyl phthalate	1.436	1.532	-6.7	76	0.02
86	Indeno(1,2,3-cd)pyrene	1.121	1.140	-1.7	77	0.02
87 I	Perylene-d12	1.000	1.000	0.0	76	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052019\
Data File : BF114428.D
Acq On : 20 May 2019 10:26
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 20 10:49:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 04:47:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.281	1.330	-3.8	79	0.02
89	Benzo(k)fluoranthene	1.177	1.168	0.8	72	0.02
90 C	Benzo(a)pyrene	1.128	1.154	-2.3	77	0.02
91	Dibenzo(a,h)anthracene	0.937	0.936	0.1	78	0.02
92	Benzo(g,h,i)perylene	0.939	0.939	0.0	79	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0