

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052019\  
 Data File : BF114441.D  
 Acq On : 20 May 2019 16:23  
 Operator : JU/SJ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 LabSampleId :  
 SSTDCCC040

Quant Time: May 20 17:21:53 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon May 20 04:47:40 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | AvqRF | CCRF  | %Dev | Area% | Dev(min) |
|------|-----------------------------|-------|-------|------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0  | 66    | -0.01    |
| 2    | 1,4-Dioxane                 | 0.683 | 0.583 | 14.6 | 55    | 0.00     |
| 3    | Pyridine                    | 1.882 | 1.642 | 12.8 | 58    | 0.00     |
| 4    | n-Nitrosodimethylamine      | 0.908 | 0.783 | 13.8 | 56    | 0.00     |
| 5 S  | 2-Fluorophenol              | 1.243 | 1.211 | 2.6  | 62    | -0.01    |
| 6    | Aniline                     | 2.123 | 2.119 | 0.2  | 64    | 0.00     |
| 7 S  | Phenol-d6                   | 1.470 | 1.502 | -2.2 | 67    | -0.01    |
| 8    | 2-Chlorophenol              | 1.327 | 1.375 | -3.6 | 67    | -0.01    |
| 9    | Benzaldehyde                | 1.029 | 0.928 | 9.8  | 56    | -0.01    |
| 10 C | Phenol                      | 1.729 | 1.748 | -1.1 | 65    | -0.01    |
| 11   | bis(2-Chloroethyl)ether     | 1.313 | 1.371 | -4.4 | 68    | -0.01    |
| 12   | 1,3-Dichlorobenzene         | 1.489 | 1.520 | -2.1 | 67    | -0.01    |
| 13 C | 1,4-Dichlorobenzene         | 1.501 | 1.555 | -3.6 | 68    | -0.01    |
| 14   | 1,2-Dichlorobenzene         | 1.371 | 1.418 | -3.4 | 66    | 0.00     |
| 15   | Benzyl Alcohol              | 1.146 | 1.125 | 1.8  | 63    | -0.01    |
| 16   | 2,2'-oxybis(1-Chloropropane | 2.545 | 2.577 | -1.3 | 66    | -0.01    |
| 17   | 2-Methylphenol              | 1.090 | 1.092 | -0.2 | 66    | -0.01    |
| 18   | Hexachloroethane            | 0.563 | 0.576 | -2.3 | 66    | 0.00     |
| 19 P | n-Nitroso-di-n-propylamine  | 0.932 | 0.949 | -1.8 | 68    | -0.01    |
| 20   | 3+4-Methylphenols           | 1.259 | 1.353 | -7.5 | 70    | -0.01    |
| 21 I | Naphthalene-d8              | 1.000 | 1.000 | 0.0  | 70    | -0.01    |
| 22   | Acetophenone                | 0.443 | 0.450 | -1.6 | 70    | -0.01    |
| 23 S | Nitrobenzene-d5             | 0.356 | 0.355 | 0.3  | 68    | 0.00     |
| 24   | Nitrobenzene                | 0.363 | 0.365 | -0.6 | 68    | -0.01    |
| 25   | Isophorone                  | 0.661 | 0.662 | -0.2 | 71    | -0.01    |
| 26 C | 2-Nitrophenol               | 0.176 | 0.185 | -5.1 | 73    | 0.00     |
| 27   | 2,4-Dimethylphenol          | 0.277 | 0.271 | 2.2  | 68    | 0.00     |
| 28   | bis(2-Chloroethoxy)methane  | 0.429 | 0.434 | -1.2 | 71    | -0.01    |
| 29 C | 2,4-Dichlorophenol          | 0.275 | 0.280 | -1.8 | 69    | 0.00     |
| 30   | 1,2,4-Trichlorobenzene      | 0.313 | 0.313 | 0.0  | 68    | 0.00     |
| 31   | Naphthalene                 | 0.934 | 0.960 | -2.8 | 69    | -0.01    |
| 32   | Benzoic acid                | 0.181 | 0.187 | -3.3 | 72    | -0.01    |
| 33   | 4-Chloroaniline             | 0.426 | 0.438 | -2.8 | 69    | 0.00     |
| 34 C | Hexachlorobutadiene         | 0.178 | 0.180 | -1.1 | 68    | 0.00     |
| 35   | Caprolactam                 | 0.090 | 0.097 | -7.8 | 71    | -0.02    |
| 36 C | 4-Chloro-3-methylphenol     | 0.277 | 0.295 | -6.5 | 71    | -0.01    |
| 37   | 2-Methylnaphthalene         | 0.603 | 0.637 | -5.6 | 71    | 0.00     |
| 38   | 1-Methylnaphthalene         | 0.568 | 0.595 | -4.8 | 70    | 0.00     |
| 39 I | Acenaphthene-d10            | 1.000 | 1.000 | 0.0  | 72    | 0.00     |
| 40   | 1,2,4,5-Tetrachlorobenzene  | 0.606 | 0.626 | -3.3 | 71    | 0.00     |
| 41 P | Hexachlorocyclopentadiene   | 0.255 | 0.238 | 6.7  | 63    | 0.00     |
| 42 S | 2,4,6-Tribromophenol        | 0.207 | 0.198 | 4.3  | 70    | 0.00     |
| 43 C | 2,4,6-Trichlorophenol       | 0.417 | 0.414 | 0.7  | 69    | 0.00     |
| 44   | 2,4,5-Trichlorophenol       | 0.427 | 0.420 | 1.6  | 68    | 0.00     |

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052019\  
 Data File : BF114441.D  
 Acq On : 20 May 2019 16:23  
 Operator : JU/SJ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 LabSampleId :  
 SSTDCCC040

Quant Time: May 20 17:21:53 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon May 20 04:47:40 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                   | AvqRF | CCRF  | %Dev  | Area% | Dev(min) |
|------|----------------------------|-------|-------|-------|-------|----------|
| 45 S | 2-Fluorobiphenyl           | 1.322 | 1.289 | 2.5   | 71    | 0.00     |
| 46   | 1,1'-Biphenyl              | 1.616 | 1.634 | -1.1  | 71    | -0.01    |
| 47   | 2-Chloronaphthalene        | 1.300 | 1.304 | -0.3  | 70    | 0.00     |
| 48   | 2-Nitroaniline             | 0.461 | 0.465 | -0.9  | 71    | 0.00     |
| 49   | Acenaphthylene             | 1.978 | 1.992 | -0.7  | 70    | -0.01    |
| 50   | Dimethylphthalate          | 1.468 | 1.496 | -1.9  | 72    | 0.00     |
| 51   | 2,6-Dinitrotoluene         | 0.326 | 0.330 | -1.2  | 72    | 0.00     |
| 52 C | Acenaphthene               | 1.214 | 1.201 | 1.1   | 70    | 0.00     |
| 53   | 3-Nitroaniline             | 0.404 | 0.411 | -1.7  | 71    | 0.00     |
| 54 P | 2,4-Dinitrophenol          | 0.131 | 0.128 | 2.3   | 66    | 0.00     |
| 55   | Dibenzofuran               | 1.780 | 1.719 | 3.4   | 69    | 0.00     |
| 56 P | 4-Nitrophenol              | 0.286 | 0.258 | 9.8   | 66    | 0.00     |
| 57   | 2,4-Dinitrotoluene         | 0.442 | 0.417 | 5.7   | 68    | 0.00     |
| 58   | Fluorene                   | 1.375 | 1.386 | -0.8  | 73    | 0.00     |
| 59   | 2,3,4,6-Tetrachlorophenol  | 0.369 | 0.350 | 5.1   | 68    | 0.00     |
| 60   | Diethylphthalate           | 1.492 | 1.495 | -0.2  | 73    | 0.00     |
| 61   | 4-Chlorophenyl-phenylether | 0.675 | 0.690 | -2.2  | 73    | 0.00     |
| 62   | 4-Nitroaniline             | 0.425 | 0.423 | 0.5   | 73    | 0.00     |
| 63   | Azobenzene                 | 1.444 | 1.434 | 0.7   | 72    | 0.00     |
| 64 I | Phenanthrene-d10           | 1.000 | 1.000 | 0.0   | 71    | 0.00     |
| 65   | 4,6-Dinitro-2-methylphenol | 0.096 | 0.101 | -5.2  | 73    | 0.00     |
| 66 c | n-Nitrosodiphenylamine     | 0.607 | 0.627 | -3.3  | 71    | 0.00     |
| 67   | 4-Bromophenyl-phenylether  | 0.220 | 0.226 | -2.7  | 72    | 0.00     |
| 68   | Hexachlorobenzene          | 0.244 | 0.246 | -0.8  | 71    | 0.00     |
| 69   | Atrazine                   | 0.189 | 0.189 | 0.0   | 71    | 0.00     |
| 70 C | Pentachlorophenol          | 0.115 | 0.105 | 8.7   | 62    | 0.00     |
| 71   | Phenanthrene               | 0.973 | 1.004 | -3.2  | 72    | 0.00     |
| 72   | Anthracene                 | 0.977 | 1.025 | -4.9  | 72    | 0.00     |
| 73   | Carbazole                  | 0.932 | 0.970 | -4.1  | 72    | 0.00     |
| 74   | Di-n-butylphthalate        | 1.074 | 1.181 | -10.0 | 76    | 0.00     |
| 75 C | Fluoranthene               | 1.048 | 1.069 | -2.0  | 71    | 0.00     |
| 76 I | Chrysene-d12               | 1.000 | 1.000 | 0.0   | 67    | 0.00     |
| 77   | Benzidine                  | 0.898 | 0.793 | 11.7  | 60    | 0.00     |
| 78   | Pyrene                     | 1.609 | 1.637 | -1.7  | 71    | 0.00     |
| 79 S | Terphenyl-d14              | 0.970 | 1.008 | -3.9  | 72    | 0.00     |
| 80   | Butylbenzylphthalate       | 0.677 | 0.749 | -10.6 | 74    | 0.00     |
| 81   | Benzo(a)anthracene         | 1.294 | 1.383 | -6.9  | 70    | 0.00     |
| 82   | 3,3'-Dichlorobenzidine     | 0.502 | 0.515 | -2.6  | 65    | 0.00     |
| 83   | Chrysene                   | 1.268 | 1.301 | -2.6  | 67    | 0.00     |
| 84   | Bis(2-ethylhexyl)phthalate | 0.886 | 0.998 | -12.6 | 75    | 0.00     |
| 85 c | Di-n-octyl phthalate       | 1.436 | 1.525 | -6.2  | 69    | 0.01     |
| 86   | Indeno(1,2,3-cd)pyrene     | 1.121 | 1.150 | -2.6  | 71    | 0.00     |
| 87 I | Perylene-d12               | 1.000 | 1.000 | 0.0   | 65    | 0.01     |

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052019\  
Data File : BF114441.D  
Acq On : 20 May 2019 16:23  
Operator : JU/SJ  
Sample : SSTDCCC040  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_F  
LabSampleId :  
SSTDCCC040

Quant Time: May 20 17:21:53 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon May 20 04:47:40 2019  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound               | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|------------------------|-------|-------|------|-------|----------|
| 88   | Benzo(b)fluoranthene   | 1.281 | 1.310 | -2.3 | 67    | 0.00     |
| 89   | Benzo(k)fluoranthene   | 1.177 | 1.154 | 2.0  | 61    | 0.00     |
| 90 C | Benzo(a)pyrene         | 1.128 | 1.151 | -2.0 | 66    | 0.01     |
| 91   | Dibenzo(a,h)anthracene | 0.937 | 1.008 | -7.6 | 72    | 0.00     |
| 92   | Benzo(g,h,i)perylene   | 0.939 | 1.012 | -7.8 | 73    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0