

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052024\
Data File : BF137958.D
Acq On : 20 May 2024 13:59
Operator : CG\JU
Sample : P2518-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-MH-LL

Quant Time: May 20 14:34:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF051424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 15 04:12:16 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.040	152	169018	20.000	ng	0.00
21) Naphthalene-d8	8.322	136	659100	20.000	ng	0.00
39) Acenaphthene-d10	10.081	164	386625	20.000	ng	0.00
64) Phenanthrene-d10	11.575	188	694082	20.000	ng	0.00
76) Chrysene-d12	14.221	240	372696	20.000	ng	0.00
86) Perylene-d12	15.786	264	361409	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.675	112	1013275	101.835	ng	0.01
7) Phenol-d6	6.657	99	1167127	95.221	ng	0.00
23) Nitrobenzene-d5	7.604	82	832750	74.355	ng	0.00
42) 2,4,6-Tribromophenol	10.875	330	508592	130.464	ng	0.00
45) 2-Fluorobiphenyl	9.398	172	1862821	74.218	ng	0.00
79) Terphenyl-d14	13.169	244	2279159	90.128	ng	0.00


Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052024\
Data File : BF137958.D
Acq On : 20 May 2024 13:59
Operator : CG\JU
Sample : P2518-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-MH-LL

Quant Time: May 20 14:34:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF051424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 15 04:12:16 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.040 min Scan# 84
Delta R.T. -0.005 min
Lab File: BF137958.D
Acq: 20 May 2024 13:59




Instrument :
BNA_F
ClientSampleId :
WC-MH-LL


Tgt Ion:152 Resp: 169018
Ion Ratio Lower Upper
152 100
150 157.3 125.7 188.5
115 61.4 47.0 70.6



#5
2-Fluorophenol
Concen: 101.835 ng
RT: 5.675 min Scan# 610
Delta R.T. 0.012 min
Lab File: BF137958.D
Acq: 20 May 2024 13:59

Tgt Ion:112 Resp: 1013275
Ion Ratio Lower Upper
112 100
64 55.6 42.4 63.6
63 28.8 21.4 32.2









#7
Phenol-d6
Concen: 95.221 ng
RT: 6.657 min Scan# 71
Delta R.T. 0.000 min
Lab File: BF137958.D
Acq: 20 May 2024 13:59

Instrument :
BNA_F
ClientSampleId :
WC-MH-LL

Tgt Ion: 99 Resp: 1167127




Ion	Ratio	Lower	Upper
99	100		
42	15.8	12.6	18.8
71	33.4	27.1	40.7



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.322 min Scan# 1060
Delta R.T. -0.006 min
Lab File: BF137958.D
Acq: 20 May 2024 13:59

Tgt Ion: 136 Resp: 659100

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.7	13.1
54	7.1	5.4	8.2
68	5.8	4.6	7.0









#23
Nitrobenzene-d5
Concen: 74.355 ng
RT: 7.604 min Scan# 91
Delta R.T. -0.006 min
Lab File: BF137958.D
Acq: 20 May 2024 13:59




Instrument :
BNA_F
ClientSampleId :
WC-MH-LL


Tgt Ion: 82 Resp: 832750
Ion Ratio Lower Upper
82 100
128 49.8 40.3 60.5
54 46.5 36.8 55.2



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.081 min Scan# 1359
Delta R.T. -0.005 min
Lab File: BF137958.D
Acq: 20 May 2024 13:59

Tgt Ion:164 Resp: 386625
Ion Ratio Lower Upper
164 100
162 99.6 79.4 119.2
160 42.5 33.1 49.7









#42
2,4,6-Tribromophenol
Concen: 130.464 ng
RT: 10.875 min Scan# 1494
Delta R.T. -0.006 min
Lab File: BF137958.D
Acq: 20 May 2024 13:59




Instrument :
BNA_F
ClientSampleId :
WC-MH-LL


Tgt Ion:330 Resp: 508592
Ion Ratio Lower Upper
330 100
332 97.0 75.8 113.8
141 27.6 23.8 35.6



#45
2-Fluorobiphenyl
Concen: 74.218 ng
RT: 9.398 min Scan# 1243
Delta R.T. 0.000 min
Lab File: BF137958.D
Acq: 20 May 2024 13:59

Tgt Ion:172 Resp: 1862821
Ion Ratio Lower Upper
172 100
171 36.4 28.9 43.3
170 23.6 18.9 28.3





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.575 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF137958.D

Acq: 20 May 2024 13:59

Instrument :

BNA_F

ClientSampleId :

WC-MH-LL





Tgt Ion:188 Resp: 694082

Ion Ratio Lower Upper

188 100

94 10.1 7.0 10.4

80 11.0 7.7 11.5



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.221 min Scan# 2063

Delta R.T. -0.006 min

Lab File: BF137958.D

Acq: 20 May 2024 13:59




Tgt Ion:240 Resp: 372696


Ion Ratio Lower Upper

240 100

120 12.6 8.9 13.3

236 24.9 21.0 31.4









#79
 Terphenyl-d14
 Concen: 90.128 ng
 RT: 13.169 min Scan# 1883
 Delta R.T. 0.006 min
 Lab File: BF137958.D
 Acq: 20 May 2024 13:59

Instrument :
 BNA_F
 ClientSampleId :
 WC-MH-LL

Tgt Ion:244 Resp: 2279159
 Ion Ratio Lower Upper
 244 100
 212 6.8 5.6 8.4
 122 9.8 8.8 13.2



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.786 min Scan# 2329
 Delta R.T. 0.000 min
 Lab File: BF137958.D
 Acq: 20 May 2024 13:59

Tgt Ion:264 Resp: 361409
 Ion Ratio Lower Upper
 264 100
 260 23.8 18.3 27.5
 265 21.4 17.6 26.4

