

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052224\
 Data File : BF137995.D
 Acq On : 22 May 2024 09:46
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 22 10:10:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF051424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 15 04:12:16 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	92	-0.01
2	1,4-Dioxane	40.000	39.250	1.9	91	-0.02
3	Pyridine	40.000	41.532	-3.8	94	-0.02
4	n-Nitrosodimethylamine	40.000	37.469	6.3	88	-0.02
5 S	2-Fluorophenol	80.000	79.177	1.0	93	-0.01
6	Aniline	40.000	42.453	-6.1	98	0.00
7 S	Phenol-d6	80.000	78.268	2.2	92	0.00
8	2-Chlorophenol	40.000	39.850	0.4	93	0.00
9	Benzaldehyde	40.000	41.209	-3.0	94	0.00
10 C	Phenol	40.000	39.206	2.0	92	0.00
11	bis(2-Chloroethyl)ether	40.000	39.907	0.2	93	0.00
12	1,3-Dichlorobenzene	40.000	39.442	1.4	92	-0.01
13 C	1,4-Dichlorobenzene	40.000	39.729	0.7	93	-0.01
14	1,2-Dichlorobenzene	40.000	39.562	1.1	93	0.00
15	Benzyl Alcohol	40.000	42.058	-5.1	93	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.529	-1.3	94	0.00
17	2-Methylphenol	40.000	39.520	1.2	92	0.00
18	Hexachloroethane	40.000	40.485	-1.2	93	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.872	0.3	93	-0.01
20	3+4-Methylphenols	40.000	39.881	0.3	93	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	88	0.00
22	Acetophenone	40.000	40.512	-1.3	93	-0.01
23 S	Nitrobenzene-d5	80.000	81.689	-2.1	92	-0.01
24	Nitrobenzene	40.000	40.431	-1.1	91	-0.01
25	Isophorone	40.000	41.225	-3.1	93	0.00
26 C	2-Nitrophenol	40.000	42.261	-5.7	92	0.00
27	2,4-Dimethylphenol	40.000	41.066	-2.7	92	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.868	-2.2	93	0.00
29 C	2,4-Dichlorophenol	40.000	41.395	-3.5	92	0.00
30	1,2,4-Trichlorobenzene	40.000	41.026	-2.6	94	0.00
31	Naphthalene	40.000	40.527	-1.3	92	-0.01
32	Benzoic acid	40.000	45.450	-13.6	99	0.00
33	4-Chloroaniline	40.000	43.350	-8.4	96	0.00
34 C	Hexachlorobutadiene	40.000	41.368	-3.4	94	0.00
35	Caprolactam	40.000	42.502	-6.3	92	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.586	-4.0	93	0.00
37	2-Methylnaphthalene	40.000	41.045	-2.6	94	0.00
38	1-Methylnaphthalene	40.000	40.833	-2.1	93	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	89	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.217	-0.5	94	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.426	-6.1	91	0.00
42 S	2,4,6-Tribromophenol	80.000	84.451	-5.6	92	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.690	-4.2	94	0.00
44	2,4,5-Trichlorophenol	40.000	41.692	-4.2	93	0.00
45 S	2-Fluorobiphenyl	80.000	78.850	1.4	93	0.00
46	1,1'-Biphenyl	40.000	40.144	-0.4	94	0.00
47	2-Chloronaphthalene	40.000	40.326	-0.8	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052224\
 Data File : BF137995.D
 Acq On : 22 May 2024 09:46
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 22 10:10:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF051424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 15 04:12:16 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.503	-3.8	91	0.00
49	Acenaphthylene	40.000	40.417	-1.0	92	0.00
50	Dimethylphthalate	40.000	41.163	-2.9	93	0.00
51	2,6-Dinitrotoluene	40.000	41.833	-4.6	93	0.00
52 C	Acenaphthene	40.000	41.088	-2.7	94	0.00
53	3-Nitroaniline	40.000	42.024	-5.1	92	0.00
54 P	2,4-Dinitrophenol	40.000	42.248	-5.6	96	0.00
55	Dibenzofuran	40.000	40.102	-0.3	92	0.00
56 P	4-Nitrophenol	40.000	43.093	-7.7	89	0.00
57	2,4-Dinitrotoluene	40.000	42.266	-5.7	90	0.00
58	Fluorene	40.000	41.010	-2.5	93	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.409	-3.5	90	0.00
60	Diethylphthalate	40.000	41.837	-4.6	93	0.00
61	4-Chlorophenyl-phenylether	40.000	41.158	-2.9	93	0.00
62	4-Nitroaniline	40.000	42.311	-5.8	90	0.00
63	Azobenzene	40.000	40.365	-0.9	91	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	87	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.876	-12.2	94	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.951	-2.4	92	0.00
67	4-Bromophenyl-phenylether	40.000	41.605	-4.0	94	0.00
68	Hexachlorobenzene	40.000	41.061	-2.7	93	0.00
69	Atrazine	40.000	39.436	1.4	85	0.00
70 C	Pentachlorophenol	40.000	44.423	-11.1	89	0.00
71	Phenanthrene	40.000	40.240	-0.6	91	0.00
72	Anthracene	40.000	40.575	-1.4	91	0.00
73	Carbazole	40.000	40.914	-2.3	91	0.00
74	Di-n-butylphthalate	40.000	42.657	-6.6	91	0.00
75 C	Fluoranthene	40.000	40.388	-1.0	87	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	79	0.00
77	Benzydine	40.000	77.097	-92.7#	107	0.00
78	Pyrene	40.000	41.981	-5.0	87	0.00
79 S	Terphenyl-d14	80.000	84.123	-5.2	88	0.00
80	Butylbenzylphthalate	40.000	44.171	-10.4	84	0.00
81	Benzo(a)anthracene	40.000	41.403	-3.5	84	0.00
82	3,3'-Dichlorobenzidine	40.000	45.523	-13.8	90	0.00
83	Chrysene	40.000	41.009	-2.5	84	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.575	-6.4	80	0.00
85 c	Di-n-octyl phthalate	40.000	46.723	-16.8	89	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	45.053	-12.6	111	0.00
88	Benzo(b)fluoranthene	40.000	40.211	-0.5	92	0.00
89	Benzo(k)fluoranthene	40.000	39.757	0.6	96	0.00
90 C	Benzo(a)pyrene	40.000	41.754	-4.4	99	0.00
91	Dibenzo(a,h)anthracene	40.000	44.731	-11.8	109	0.01
92	Benzo(g,h,i)perylene	40.000	44.450	-11.1	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052224\
Data File : BF137995.D
Acq On : 22 May 2024 09:46
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 22 10:10:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF051424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 15 04:12:16 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0