

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052319\
 Data File : BF114534.D
 Acq On : 23 May 2019 20:15
 Operator : JU/SJ
 Sample : K2935-04
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PURGE-WATER-MUD

Quant Time: May 24 07:47:54 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 24 07:23:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	192229	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	746717	20.00	ng	0.00
39) Acenaphthene-d10	9.85	164	361271	20.00	ng	0.00
64) Phenanthrene-d10	11.33	188	761602	20.00	ng	0.00
76) Chrysene-d12	13.98	240	659030	20.00	ng	0.00
87) Perylene-d12	15.44	264	652876	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.44	112	762376	63.82	ng	0.00
7) Phenol-d6	6.46	99	1003545	71.03	ng	0.00
23) Nitrobenzene-d5	7.37	82	821871	61.77	ng	0.00
42) 2,4,6-Tribromophenol	10.64	330	250654	67.11	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	1501228	62.86	ng	0.00
79) Terphenyl-d14	12.92	244	1772348	55.44	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.56	163	98099	3.699	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052319\
Data File : BF114534.D
Acq On : 23 May 2019 20:15
Operator : JU/SJ
Sample : K2935-04
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PURGE-WATER-MUD

Quant Time: May 24 07:47:54 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 24 07:23:33 2019
Response via : Initial Calibration

