

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052319\  
 Data File : BF114546.D  
 Acq On : 24 May 2019 1:33  
 Operator : JU/SJ  
 Sample : K2978-04  
 Misc :  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 CONCRETE-COMP

Quant Time: May 24 08:17:30 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri May 24 07:23:33 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min)  |
|-----------------------------|-------|------|----------|-------|-------|-----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.81  | 152  | 164647   | 20.00 | ng    | 0.00      |
| 21) Naphthalene-d8          | 8.09  | 136  | 648732   | 20.00 | ng    | 0.00      |
| 39) Acenaphthene-d10        | 9.85  | 164  | 324272   | 20.00 | ng    | 0.00      |
| 64) Phenanthrene-d10        | 11.33 | 188  | 693780   | 20.00 | ng    | 0.00      |
| 76) Chrysene-d12            | 13.98 | 240  | 476263   | 20.00 | ng    | 0.00      |
| 87) Perylene-d12            | 15.44 | 264  | 467441   | 20.00 | ng    | 0.00      |
| System Monitoring Compounds |       |      |          |       |       |           |
| 5) 2-Fluorophenol           | 5.45  | 112  | 460459   | 45.01 | ng    | 0.00      |
| 7) Phenol-d6                | 6.46  | 99   | 952527   | 78.72 | ng    | 0.00      |
| 23) Nitrobenzene-d5         | 7.37  | 82   | 655592   | 56.71 | ng    | 0.00      |
| 42) 2,4,6-Tribromophenol    | 10.64 | 330  | 57627    | 17.19 | ng    | 0.00      |
| 45) 2-Fluorobiphenyl        | 9.17  | 172  | 1259698  | 58.76 | ng    | 0.00      |
| 79) Terphenyl-d14           | 12.92 | 244  | 1468128  | 63.55 | ng    | 0.00      |
| Target Compounds            |       |      |          |       |       |           |
| 50) Dimethylphthalate       | 9.56  | 163  | 165734   | 6.962 | ng    | Qvalue 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052319\  
Data File : BF114546.D  
Acq On : 24 May 2019 1:33  
Operator : JU/SJ  
Sample : K2978-04  
Misc :  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
CONCRETE-COMP

Quant Time: May 24 08:17:30 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri May 24 07:23:33 2019  
Response via : Initial Calibration

