

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF052319\
 Data File : BF114554.D
 Acq On : 24 May 2019 5:04
 Operator : JU/SJ
 Sample : K3008-01 2X
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COMP-19

Manual Integrations
 APPROVED

mohammad
 5/24/2019 1:37:04 PM

Quant Time: Jun 10 04:19:46 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 24 07:23:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	173946	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	698693	20.00	ng	0.00
39) Acenaphthene-d10	9.85	164	331944	20.00	ng	0.00
64) Phenanthrene-d10	11.34	188	605964	20.00	ng	0.00
76) Chrysene-d12	13.99	240	378645	20.00	ng	0.00
87) Perylene-d12	15.46	264	311661	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	267598	24.76	ng	0.00
7) Phenol-d6	6.46	99	397787	31.12	ng	0.00
23) Nitrobenzene-d5	7.37	82	226190	18.17	ng	0.00
42) 2,4,6-Tribromophenol	10.64	330	65216	19.00	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	437627	19.94	ng	0.00
79) Terphenyl-d14	12.92	244	386703	21.05	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.56	163	56081	2.301	ng	99
71) Phenanthrene	11.36	178	500637	16.982	ng	98
72) Anthracene	11.42	178	124089	4.191	ng	95
75) Fluoranthene	12.56	202	1251357	39.416	ng	99
78) Pyrene	12.79	202	1121675	36.824	ng	99
81) Benzo(a)anthracene	13.97	228	702237	28.674	ng	99
83) Chrysene	14.01	228	617838	25.735	ng	97
86) Indeno(1,2,3-cd)pyrene	16.94	276	206821	9.748	ng	94
88) Benzo(b)fluoranthene	15.03	252	675698m	33.841	ng	
89) Benzo(k)fluoranthene	15.06	252	249311m	13.593	ng	
90) Benzo(a)pyrene	15.40	252	411821	23.436	ng	98
91) Dibenzo(a,h)anthracene	16.93	278	59970	4.107	ng	# 89
92) Benzo(a,h,i)perylene	17.39	276	192080	13.126	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF052319\
Data File : BF114554.D
Acq On : 24 May 2019 5:04
Operator : JU/SJ
Sample : K3008-01 2X
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
COMP-19

Manual Integrations
APPROVED

mohammad
5/24/2019 1:37:04 PM

Quant Time: Jun 10 04:19:46 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 24 07:23:33 2019
Response via : Initial Calibration

