

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052519\
 Data File : BF114617.D
 Acq On : 25 May 2019 21:08
 Operator : JU/SJ
 Sample : PB119929BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: May 26 23:47:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 24 07:23:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	141046	20.00	ng	-0.02
21) Naphthalene-d8	8.08	136	530843	20.00	ng	-0.02
39) Acenaphthene-d10	9.84	164	248578	20.00	ng	-0.01
64) Phenanthrene-d10	11.33	188	478330	20.00	ng	-0.01
76) Chrysene-d12	13.97	240	350782	20.00	ng	-0.02
87) Perylene-d12	15.43	264	335688	20.00	ng	-0.02

System Monitoring Compounds						
5) 2-Fluorophenol	5.44	112	1136883	129.71	ng	0.00
7) Phenol-d6	6.45	99	1426959	137.65	ng	-0.01
23) Nitrobenzene-d5	7.37	82	861498	91.08	ng	-0.01
42) 2,4,6-Tribromophenol	10.63	330	330579	128.64	ng	-0.01
45) 2-Fluorobiphenyl	9.16	172	1516782	92.30	ng	-0.02
79) Terphenyl-d14	12.90	244	1527996	89.80	ng	-0.01

Target Compounds				Qvalue		
2) 1,4-Dioxane	2.62	88	212898	44.192	ng	99
3) Pyridine	3.36	79	403020	30.363	ng	99
4) n-Nitrosodimethylamine	3.31	42	222250	34.722	ng	100
6) Aniline	6.47	93	411655	27.491	ng	# 58
8) 2-Chlorophenol	6.59	128	414350	44.284	ng	95
9) Benzaldehyde	6.35	77	199925	27.549	ng	96
10) Phenol	6.46	94	503680	41.304	ng	87
11) bis(2-Chloroethyl)ether	6.53	93	393873	42.544	ng	99
12) 1,3-Dichlorobenzene	6.74	146	428549	40.803	ng	98
13) 1,4-Dichlorobenzene	6.82	146	445077	42.053	ng	97
14) 1,2-Dichlorobenzene	6.97	146	403713	41.752	ng	99
15) Benzyl Alcohol	6.95	79	322105	39.840	ng	99
16) 2,2'-oxybis(1-Chloropropan	7.07	45	691311	38.512	ng	57
17) 2-Methylphenol	7.07	107	321467	41.824	ng	# 90
18) Hexachloroethane	7.30	117	159771	40.217	ng	95
19) n-Nitroso-di-n-propylamine	7.22	70	276180	42.033	ng	97
20) 3+4-Methylphenols	7.22	107	425841	47.953	ng	# 67
22) Acetophenone	7.21	105	548822	46.669	ng	# 89
24) Nitrobenzene	7.39	77	426279	44.247	ng	98
25) Isophorone	7.62	82	760881	43.373	ng	99
26) 2-Nitrophenol	7.70	139	218600	46.768	ng	99
27) 2,4-Dimethylphenol	7.75	122	353258	48.121	ng	100
28) bis(2-Chloroethoxy)methane	7.83	93	483031	42.436	ng	98
29) 2,4-Dichlorophenol	7.95	162	333278	45.592	ng	100
30) 1,2,4-Trichlorobenzene	8.03	180	359171	43.209	ng	97
31) Naphthalene	8.10	128	1147555	46.279	ng	98
32) Benzoic acid	7.90	122	184634	37.134	ng	97
33) 4-Chloroaniline	8.16	127	299270	26.485	ng	98
34) Hexachlorobutadiene	8.22	225	197669	41.794	ng	98
35) Caprolactam	8.53	113	58264	24.444	ng	92
36) 4-Chloro-3-methylphenol	8.66	107	348087	47.306	ng	100
37) 2-Methylnaphthalene	8.80	142	763696	47.735	ng	98
38) 1-Methylnaphthalene	8.90	142	727102	48.261	ng	100
40) 1,2,4,5-Tetrachlorobenzene	8.97	216	349049	46.355	ng	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052519\
 Data File : BF114617.D
 Acq On : 25 May 2019 21:08
 Operator : JU/SJ
 Sample : PB119929BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: May 26 23:47:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 24 07:23:33 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	8.95	237	147375	44.064	ng	99
43) 2,4,6-Trichlorophenol	9.09	196	212148	40.961	ng	99
44) 2,4,5-Trichlorophenol	9.13	196	222293	41.850	ng	95
46) 1,1'-Biphenyl	9.26	154	929383	46.280	ng	99
47) 2-Chloronaphthalene	9.29	162	700666	43.354	ng	98
48) 2-Nitroaniline	9.39	65	234559	40.923	ng	91
49) Acenaphthylene	9.70	152	1114882	45.356	ng	97
50) Dimethylphthalate	9.56	163	772128	42.313	ng	99
51) 2,6-Dinitrotoluene	9.63	165	175870	43.453	ng	97
52) Acenaphthene	9.87	154	655500	43.457	ng	100
53) 3-Nitroaniline	9.80	138	134993	26.869	ng	98
54) 2,4-Dinitrophenol	9.93	184	136848	68.317	ng	91
55) Dibenzofuran	10.05	168	942744	42.623	ng	99
56) 4-Nitrophenol	9.99	139	170096	47.873	ng	97
57) 2,4-Dinitrotoluene	10.04	165	216521	39.414	ng	# 73
58) Fluorene	10.39	166	778961	45.595	ng	98
59) 2,3,4,6-Tetrachlorophenol	10.17	232	189221	41.287	ng	98
60) Diethylphthalate	10.26	149	774597	41.777	ng	98
61) 4-Chlorophenyl-phenylether	10.37	204	364150	43.398	ng	97
62) 4-Nitroaniline	10.42	138	189649	35.922	ng	96
63) Azobenzene	10.54	77	747998	41.679	ng	96
65) 4,6-Dinitro-2-methylphenol	10.46	198	97874	42.583	ng	80
66) n-Nitrosodiphenylamine	10.50	169	679779	46.825	ng	99
67) 4-Bromophenyl-phenylether	10.86	248	227653	43.226	ng	97
68) Hexachlorobenzene	10.95	284	243951	41.871	ng	93
69) Atrazine	11.02	200	195719	43.322	ng	97
70) Pentachlorophenol	11.15	266	146728	53.126	ng	98
71) Phenanthrene	11.35	178	1086225	46.676	ng	97
72) Anthracene	11.40	178	1116197	47.754	ng	98
73) Carbazole	11.56	167	1050281	47.120	ng	98
74) Di-n-butylphthalate	11.88	149	1173154	45.685	ng	99
75) Fluoranthene	12.54	202	1167111	46.572	ng	99
77) Benzidine	12.66	184	579847	36.825	ng	97
78) Pyrene	12.77	202	1182244	41.896	ng	99
80) Butylbenzylphthalate	13.37	149	511915	43.100	ng	100
81) Benzo(a)anthracene	13.96	228	985866	43.452	ng	99
82) 3,3'-Dichlorobenzidine	13.92	252	286378	32.538	ng	100
83) Chrysene	14.00	228	984835	44.280	ng	99
84) Bis(2-ethylhexyl)phthalate	13.93	149	689030	44.356	ng	99
85) Di-n-octyl phthalate	14.54	149	1151725	45.736	ng	98
86) Indeno(1,2,3-cd)pyrene	16.89	276	910460	46.319	ng	100
88) Benzo(b)fluoranthene	15.00	252	915970	42.591	ng	99
89) Benzo(k)fluoranthene	15.03	252	964534	48.824	ng	97
90) Benzo(a)pyrene	15.37	252	895423	47.309	ng	98
91) Dibenzo(a,h)anthracene	16.89	278	758489	48.227	ng	99
92) Benzo(g,h,i)perylene	17.33	276	787623	49.972	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052519\
Data File : BF114617.D
Acq On : 25 May 2019 21:08
Operator : JU/SJ
Sample : PB119929BS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: May 26 23:47:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 24 07:23:33 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.80 min Scan# 801
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Instrument :
BNA_F
ClientSampleId :

Tot Ion:152 Resp: 141046
Ion Ratio Lower Upper
152 100
150 158.8 124.3 186.5
115 59.9 49.1 73.7




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 44.192 ng
RT: 2.62 min Scan# 90
Delta R.T. 0.06 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion: 88 Resp: 212898
Ion Ratio Lower Upper
88 100
58 85.2 68.7 103.1
43 35.1 28.2 42.2



Abundance Ion 88.00 (87.70 to 88.70): BF1
Ion 58.00 (57.70 to 58.70): BF1
Ion 43.00 (42.70 to 43.70): BF1





#3

Pvridine

Concen: 30.363 ng

RT: 3.36 min Scan# 217

Delta R.T. 0.06 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 79 Resp: 403020

Ion Ratio Lower Upper

79 100

52 86.9 70.6 105.8




51 43.2 34.2 51.2



Abundance Ion 79.00 (78.70 to 79.70): BF1

Ion 52.00 (51.70 to 52.70): BF1

Ion 51.00 (50.70 to 51.70): BF1



#4

n-Nitrosodimethylamine

Concen: 34.722 ng

RT: 3.31 min Scan# 208

Delta R.T. 0.05 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Tgt Ion: 42 Resp: 222250

Ion Ratio Lower Upper

42 100

74 110.5 88.6 132.8



44 7.3 5.5 8.3




Abundance Ion 42.00 (41.70 to 42.70): BF1

Ion 74.00 (73.70 to 74.70): BF1

Ion 44.00 (43.70 to 44.70): BF1





#5

2-Fluorophenol

Concen: 129.712 ng

RT: 5.44 min Scan# 570

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampled :


Tot Ion:112 Resp: 1136883

Ion Ratio Lower Upper

112 100

64 66.2 49.0 73.6




63 33.3 24.6 37.0



Abundance Ion 112.00 (111.70 to 112.70): BF1

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1



#6

Aniline

Concen: 27.491 ng

RT: 6.47 min Scan# 745

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Tgt Ion: 93 Resp: 411655

Ion Ratio Lower Upper

93 100

66 67.0 33.2 49.8#



65 43.3 17.6 26.4#




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 66.00 (65.70 to 66.70): BF1

Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 137.655 ng

RT: 6.45 min Scan# 742

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Instrument :

BNA_F

ClientSampleId :

Tot Ion: 99 Resp: 1426959

Ion	Ratio	Lower	Upper
99	100		
42	21.9	17.9	26.9
71	34.0	26.2	39.2






Abundance

Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#8

2-Chlorophenol

Concen: 44.284 ng

RT: 6.59 min Scan# 766


Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Tgt Ion: 128 Resp: 414350

Ion	Ratio	Lower	Upper
128	100		
130	32.6	12.6	52.6
64	49.7	35.9	75.9






Abundance

Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF1





#9

Benzaldehyde

Concen: 27.549 ng

RT: 6.35 min Scan# 725

Delta R.T. -0.02 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 77 Resp: 199925

Ion Ratio Lower Upper

77 100

105 96.6 79.3 119.3




106 91.1 77.0 117.0



Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 106.00 (105.70 to 106.70): BF1



#10

Phenol

Concen: 41.304 ng

RT: 6.46 min Scan# 744

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Tgt Ion: 94 Resp: 503680

Ion Ratio Lower Upper

94 100

65 39.8 14.7 54.7



66 60.3 29.6 69.6




Abundance Ion 94.00 (93.70 to 94.70): BF1

Ion 65.00 (64.70 to 65.70): BF1

Ion 66.00 (65.70 to 66.70): BF1









#11
bis(2-Chloroethyl)ether
Concen: 42.544 ng
RT: 6.53 min Scan# 756
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08




Instrument :
BNA_F
ClientSampled :


Tot Ion: 93 Resp: 393873
Ion Ratio Lower Upper
93 100
63 83.5 63.6 103.6
95 31.9 13.1 53.1



#12
1,3-Dichlorobenzene
Concen: 40.803 ng
RT: 6.74 min Scan# 791
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion: 146 Resp: 428549
Ion Ratio Lower Upper
146 100
148 63.3 51.6 77.4
75 32.1 24.2 36.2





#13

1,4-Dichlorobenzene

Concen: 42.053 ng

RT: 6.82 min Scan# 804

Delta R.T. -0.02 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :


Tot Ion:146 Resp: 445077

Ion Ratio Lower Upper




146 100

148 61.4 50.2 75.2

111 41.9 35.8 53.6



Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 41.752 ng

RT: 6.97 min Scan# 830

Delta R.T. -0.02 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Tgt Ion:146 Resp: 403713

Ion Ratio Lower Upper



146 100


148 62.8 51.1 76.7

111 45.0 35.3 52.9



Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 39.840 ng

RT: 6.95 min Scan# 827

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampled :





Tot Ion: 79 Resp: 322105

Ion Ratio Lower Upper

79 100

108 76.0 62.1 93.1

77 63.5 51.0 76.6



#16

2,2'-oxybis(1-chloropropane)

Concen: 38.512 ng

RT: 7.07 min Scan# 848

Delta R.T. -0.02 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08




Tgt Ion: 45 Resp: 691311


Ion Ratio Lower Upper

45 100

77 27.1 0.0 31.1

79 24.3 0.0 28.4









#17
2-Methylphenol
Concen: 41.824 ng
RT: 7.07 min Scan# 847
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
107	100		
108	113.9	90.5	135.7
77	62.5	37.0	55.4
79	58.1	35.8	53.8



#18
Hexachloroethane
Concen: 40.217 ng
RT: 7.30 min Scan# 887
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion	Ratio	Lower	Upper
117	100		
119	94.7	77.4	116.2
201	85.2	74.5	111.7









#19
n-Nitroso-di-n-propylamine
Concen: 42.033 ng
RT: 7.22 min Scan# 872
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08




Instrument :
BNA_F
ClientSampleId :


Tot Ion:	70	Resp:	276180
Ion	Ratio	Lower	Upper
70	100		
42	70.5	58.6	87.8
101	9.3	7.8	11.8
130	22.4	17.1	25.7



#20
3+4-Methylphenols
Concen: 47.953 ng
RT: 7.22 min Scan# 873
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion:	107	Resp:	425841
Ion	Ratio	Lower	Upper
107	100		
108	84.4	73.8	113.8
77	67.1	111.5	151.5#
79	26.0	12.0	52.0









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.08 min Scan# 1019
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08




Instrument :
BNA_F
ClientSampleId :


Tot Ion:136	Resp: 530843
Ion Ratio	Lower Upper
136 100	
137 11.2	8.8 13.2
54 9.9	7.8 11.8
68 5.8	5.1 7.7



#22
Acetophenone
Concen: 46.669 ng
RT: 7.21 min Scan# 871
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt	Ion:105	Resp:	548822
Ion	Ratio	Lower	Upper
105	100		
71	6.6	5.3	7.9
51	33.7	36.2	54.4#
120	22.9	18.6	28.0









#23
 Nitrobenzene-d5
 Concen: 91.077 ng
 RT: 7.37 min Scan# 898
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Instrument :
 BNA_F
 ClientSampled :

Tot Ion	82	Resp	861498
Ion Ratio	100		
Lower			
Upper			
128	45.4	35.5	53.3
54	57.4	48.5	72.7




Abundance Ion 82.00 (81.70 to 82.70): BF1
 Ion 128.00 (127.70 to 128.70): BF1
 Ion 54.00 (53.70 to 54.70): BF1






#24
 Nitrobenzene
 Concen: 44.247 ng
 RT: 7.39 min Scan# 901
 Delta R.T. -0.02 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Tgt Ion	77	Resp	426279
Ion Ratio <td>100</td> <td></td> <td></td>	100		
Lower			
Upper			
123	47.4	36.9	55.3
65	14.1	11.8	17.6



Abundance Ion 77.00 (76.70 to 77.70): BF1
 Ion 123.00 (122.70 to 123.70): BF1
 Ion 65.00 (64.70 to 65.70): BF1





#25

Isophorone

Concen: 43.373 ng

RT: 7.62 min Scan# 941

Delta R.T. -0.02 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :





Tot Ion: 82 Resp: 760881

Ion Ratio Lower Upper

82 100

95 7.0 5.6 8.4

138 17.9 14.8 22.2



#26

2-Nitrophenol

Concen: 46.768 ng

RT: 7.70 min Scan# 955

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08




Tgt Ion: 139 Resp: 218600


Ion Ratio Lower Upper

139 100

109 27.1 20.7 31.1

65 50.3 40.2 60.2









#27
 2,4-Dimethylphenol
 Concen: 48.121 ng
 RT: 7.75 min Scan# 962
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:122 Resp: 353258
 Ion Ratio Lower Upper
 122 100
 107 110.2 88.4 132.6
 121 57.1 45.9 68.9



#28
 bis(2-Chloroethoxy)methane
 Concen: 42.436 ng
 RT: 7.83 min Scan# 976
 Delta R.T. -0.02 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Tgt Ion: 93 Resp: 483031
 Ion Ratio Lower Upper
 93 100
 95 32.1 26.8 40.2
 123 10.9 9.5 14.3









#29
2,4-Dichlorophenol
Concen: 45.592 ng
RT: 7.95 min Scan# 997
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
162	100		
164	63.2	43.6	83.6
98	40.7	20.6	60.6



#30
1,2,4-Trichlorobenzene
Concen: 43.209 ng
RT: 8.03 min Scan# 1010
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion	Ratio	Lower	Upper
180	100		
182	93.6	76.6	114.8
145	29.2	25.1	37.7





#31

Naphthalene

Concen: 46.279 ng

RT: 8.10 min Scan# 1023

Delta R.T. -0.02 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :


Tot Ion:128 Resp: 1147555

Ion Ratio Lower Upper

128 100

129 11.2 9.6 14.4




127 13.6 11.3 16.9



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 37.134 ng

RT: 7.90 min Scan# 988

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Tgt Ion:122 Resp: 184634

Ion Ratio Lower Upper

122 100

105 122.3 99.9 139.9



77 93.3 69.4 109.4




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF1









#33
4-Chloroaniline
Concen: 26.485 ng
RT: 8.16 min Scan# 1033
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Instrument :
BNA_F
ClientSampleId :

Tot Ion:	127	Resp:	299270
Ion	Ratio	Lower	Upper
127	100		
129	33.7	26.7	40.1
65	29.4	25.4	38.0
92	18.8	15.4	23.2




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BF1
Ion 92.00 (91.70 to 92.70): BF1






#34
Hexachlorobutadiene
Concen: 41.794 ng
RT: 8.22 min Scan# 1043
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion:	225	Resp:	197669
Ion	Ratio	Lower	Upper
225	100		
223	64.3	50.2	75.2
227	66.9	52.6	79.0



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 24.444 ng
 RT: 8.53 min Scan# 1095
 Delta R.T. -0.02 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:113 Resp: 58264
 Ion Ratio Lower Upper
 113 100
 55 190.6 183.5 223.5
 56 139.3 128.8 168.8



#36
 4-Chloro-3-methylphenol
 Concen: 47.306 ng
 RT: 8.66 min Scan# 1117
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Tgt Ion:107 Resp: 348087
 Ion Ratio Lower Upper
 107 100
 144 25.4 20.4 30.6
 142 80.7 64.2 96.4









#37
2-Methylnaphthalene
Concen: 47.735 ng
RT: 8.80 min Scan# 1141
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Instrument :
BNA_F
ClientSampleId :

Tot Ion:142 Resp: 763696
Ion Ratio Lower Upper
142 100
141 88.7 72.1 108.1
115 32.9 28.0 42.0




Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 115.00 (114.70 to 115.70): E






#38
1-Methylnaphthalene
Concen: 48.261 ng
RT: 8.90 min Scan# 1158
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion:142 Resp: 727102
Ion Ratio Lower Upper
142 100
141 91.3 73.4 110.0
116 3.6 3.1 4.7



Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 116.00 (115.70 to 116.70): E





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.84 min Scan# 1318

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :


Tot Ion:164 Resp: 248578

Ion Ratio Lower Upper

164 100

162 101.7 80.6 120.8




160 46.4 36.5 54.7



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#40

1,2,4,5-Tetrachlorobenzene

Concen: 46.355 ng

RT: 8.97 min Scan# 1170

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Tgt Ion:216 Resp: 349049


Ion Ratio Lower Upper

216 100

214 78.4 63.9 95.9

179 21.4 17.0 25.6

108 18.3 14.8 22.2






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E









#41
Hexachlorocyclopentadiene
Concen: 44.064 ng
RT: 8.95 min Scan# 1167
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Instrument :
BNA_F
ClientSampled :

Tot Ion	Ratio	Lower	Upper
237	100		
235	60.9	42.0	82.0
272	12.9	0.0	32.4




Abundance Ion 236.80 (236.50 to 237.50): E
Ion 234.90 (234.60 to 235.60): E
Ion 271.75 (271.45 to 272.45): E






#42
2,4,6-Tribromophenol
Concen: 128.636 ng
RT: 10.63 min Scan# 1453
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion	Ratio	Lower	Upper
330	100		
332	102.4	81.4	122.2
141	33.8	26.6	40.0



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E






#43
 2,4,6-Trichlorophenol
 Concen: 40.961 ng
 RT: 9.09 min Scan# 1190
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	196	Resp	212148
Ion Ratio	Lower	Upper	
196	100		
198	96.2	75.9	113.9
200	29.8	23.5	35.3






Abundance

Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E



#44
 2,4,5-Trichlorophenol
 Concen: 41.850 ng
 RT: 9.13 min Scan# 1198
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Tgt Ion	196	Resp	222293
Ion Ratio	Lower	Upper	
196	100		
198	91.4	77.7	116.5
97	50.5	38.2	57.2
132	28.7	21.4	32.2




Abundance


Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 97.00 (96.70 to 97.70): BF1

Ion 132.00 (131.70 to 132.70): E









#45
2-Fluorobiphenyl
Concen: 92.300 ng
RT: 9.16 min Scan# 1202
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Instrument :
BNA_F
ClientSampleId :

Tot Ion:172 Resp: 1516782
Ion Ratio Lower Upper
172 100
171 36.0 30.7 46.1
170 23.8 20.5 30.7




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#46
1,1'-Biphenyl
Concen: 46.280 ng
RT: 9.26 min Scan# 1219
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion:154 Resp: 929383
Ion Ratio Lower Upper
154 100
153 40.7 21.9 61.9
76 12.7 0.0 32.8



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 76.00 (75.70 to 76.70): BF1









#47
 2-Chloronaphthalene
 Concen: 43.354 ng
 RT: 9.29 min Scan# 1224
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
162	100		
127	41.0	34.0	51.0
164	32.6	26.7	40.1



#48
 2-Nitroaniline
 Concen: 40.923 ng
 RT: 9.39 min Scan# 1242
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Tgt Ion	Ratio	Lower	Upper
65	100		
92	69.6	53.9	80.9
138	114.6	81.4	122.0





#49

Acenaphthylene

Concen: 45.356 ng

RT: 9.70 min Scan# 1295

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :


Tot Ion:152 Resp: 1114882

Ion Ratio Lower Upper

152 100

151 19.9 16.8 25.2




153 13.0 11.4 17.2



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 42.313 ng

RT: 9.56 min Scan# 1270

Delta R.T. -0.02 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Tgt Ion:163 Resp: 772128

Ion Ratio Lower Upper

163 100

194 3.9 3.4 5.0



164 10.3 8.2 12.2




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E






#51
 2,6-Dinitrotoluene
 Concen: 43.453 ng
 RT: 9.63 min Scan# 1282
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
165	100		
63	64.9	50.2	75.4
89	55.8	46.6	69.8






Abundance

Ion 165.00 (164.70 to 165.70): E


Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1



#52
 Acenaphthene
 Concen: 43.457 ng
 RT: 9.87 min Scan# 1324
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Tgt Ion	Ratio	Lower	Upper
154	100		
153	112.5	90.5	135.7
152	54.4	43.4	65.2






Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E









#53
3-Nitroaniline
Concen: 26.869 ng
RT: 9.80 min Scan# 1312
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Instrument :
BNA_F
ClientSampleId :

Tot Ion:138 Resp: 134993
Ion Ratio Lower Upper
138 100
108 9.9 7.8 11.6
92 111.8 91.2 136.8




Abundance Ion 138.00 (137.70 to 138.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 92.00 (91.70 to 92.70): BF1






#54
2,4-Dinitrophenol
Concen: 68.317 ng
RT: 9.93 min Scan# 1333
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion:184 Resp: 136848
Ion Ratio Lower Upper
184 100
63 59.4 55.6 83.4
154 61.1 46.0 69.0



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 154.00 (153.70 to 154.70): E









#55
Dibenzofuran
Concen: 42.623 ng
RT: 10.05 min Scan# 1353
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Instrument :
BNA_F
ClientSampled :

Tot Ion:168 Resp: 942744
Ion Ratio Lower Upper
168 100
139 38.6 31.1 46.7
169 13.7 11.4 17.0




Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E






#56
4-Nitrophenol
Concen: 47.873 ng
RT: 9.99 min Scan# 1343
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion:139 Resp: 170096
Ion Ratio Lower Upper
139 100
109 66.0 43.9 83.9
65 109.1 85.7 125.7



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF1





#57

2,4-Dinitrotoluene

Concen: 39.414 ng

RT: 10.04 min Scan# 1352

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :

Tot Ion:165 Resp: 216521



Ion Ratio Lower Upper

165 100

63 69.6 35.5 53.3#

89 85.8 55.2 82.8#

182 2.4 2.2 3.2





Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1

Ion 182.00 (181.70 to 182.70): E



#58

Fluorene

Concen: 45.595 ng

RT: 10.39 min Scan# 1411

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08



Tgt Ion:166 Resp: 778961

Ion Ratio Lower Upper

166 100

165 93.5 76.1 114.1


167 13.3 11.1 16.7




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E






#59
 2,3,4,6-Tetrachlorophenol
 Concen: 41.287 ng
 RT: 10.17 min Scan# 1375
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
232	100		
131	43.5	35.6	53.4
130	2.0	1.6	2.4
166	28.4	24.1	36.1






Abundance

Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E


Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



#60
 Diethylphthalate
 Concen: 41.777 ng
 RT: 10.26 min Scan# 1389
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Tgt Ion	Ratio	Lower	Upper
149	100		
177	20.6	17.8	26.6
150	12.4	10.2	15.2






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E






#61
 4-Chlorophenyl-phenylether
 Concen: 43.398 ng
 RT: 10.37 min Scan# 1409
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Instrument :
 BNA_F
 ClientSampled :

Tot Ion	Ratio	Lower	Upper
204	100		
206	33.1	26.6	39.8
141	60.9	46.4	69.6






Abundance

Ion 204.00 (203.70 to 204.70): E


Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E



#62
 4-Nitroaniline
 Concen: 35.922 ng
 RT: 10.42 min Scan# 1417
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Tgt Ion	Ratio	Lower	Upper
138	100		
92	47.5	33.0	73.0
108	64.3	44.9	84.9






Abundance

Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF1

Ion 108.00 (107.70 to 108.70): E





#63

Azobenzene

Concen: 41.679 ng

RT: 10.54 min Scan# 1437

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :

Tot Ion: 77 Resp: 747998


Ion	Ratio	Lower	Upper
77	100		
182	28.4	4.7	44.7
105	23.5	2.9	42.9
51	33.0	14.6	54.6

77 100

182 28.4 4.7 44.7

105 23.5 2.9 42.9

51 33.0 14.6 54.6






Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 182.00 (181.70 to 182.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 51.00 (50.70 to 51.70): BF1



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.33 min Scan# 1571

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Tgt Ion: 188 Resp: 478330

Ion	Ratio	Lower	Upper
188	100		
94	8.9	8.4	12.6
80	9.4	9.2	13.8

188 100

94 8.9 8.4 12.6



80 9.4 9.2 13.8




Abundance Ion 188.00 (187.70 to 188.70): BF1

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1









#65
 4,6-Dinitro-2-methylphenol
 Concen: 42.583 ng
 RT: 10.46 min Scan# 1423
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:198 Resp: 97874
 Ion Ratio Lower Upper
 198 100
 51 65.9 28.0 68.0
 105 48.6 21.1 61.1




Abundance Ion 198.00 (197.70 to 198.70): E
 Ion 51.00 (50.70 to 51.70): BF1
 Ion 105.00 (104.70 to 105.70): E






#66
 n-Nitrosodiphenylamine
 Concen: 46.825 ng
 RT: 10.50 min Scan# 1430
 Delta R.T. -0.01 min
 Lab File: BF114617.D
 Acq: 25 May 2019 21:08

Tgt Ion:169 Resp: 679779
 Ion Ratio Lower Upper
 169 100
 168 66.7 54.2 81.2
 167 36.3 29.8 44.8



Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E






#67
4-Bromophenyl-phenylether
Concen: 43.226 ng
RT: 10.86 min Scan# 1492
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
248	100		
250	95.6	77.4	116.2
141	73.2	55.6	83.4






Abundance

Ion 248.00 (247.70 to 248.70): E


Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E



#68
Hexachlorobenzene
Concen: 41.871 ng
RT: 10.95 min Scan# 1506
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion	Ratio	Lower	Upper
284	100		
142	29.5	28.4	42.6
249	29.6	25.2	37.8






Abundance

Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E






#69
Atrazine
Concen: 43.322 ng
RT: 11.02 min Scan# 1519
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
200	100		
173	28.4	8.9	48.9
215	47.7	30.7	70.7






Abundance

Ion 200.00 (199.70 to 200.70): E


Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E



#70
Pentachlorophenol
Concen: 53.126 ng
RT: 11.15 min Scan# 1540
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion	Ratio	Lower	Upper
266	100		
268	64.5	51.1	76.7
264	66.7	51.1	76.7






Abundance

Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E





#71

Phenanthrene

Concen: 46.676 ng

RT: 11.35 min Scan# 1575

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :


Tot Ion:178 Resp: 1086225

Ion	Ratio	Lower	Upper
178	100		
176	19.3	16.4	24.6
179	15.2	13.2	19.8

178 100

176 19.3 16.4 24.6




179 15.2 13.2 19.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Anthracene

Concen: 47.754 ng

RT: 11.40 min Scan# 1584

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Tgt Ion:178 Resp: 1116197

Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.8	23.8
179	15.4	13.0	19.4

178 100

176 18.9 15.8 23.8



179 15.4 13.0 19.4




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#73

Carbazole

Concen: 47.120 ng

RT: 11.56 min Scan# 1610

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :


Tot Ion:167 Resp: 1050281

Ion Ratio Lower Upper

167 100

166 21.5 18.2 27.2




139 12.9 10.6 16.0



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#74

Di-n-butylphthalate

Concen: 45.685 ng

RT: 11.88 min Scan# 1665

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Tgt Ion:149 Resp: 1173154

Ion Ratio Lower Upper

149 100

150 9.0 7.7 11.5



104 4.3 3.6 5.4




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E









#75
Fluoranthene
Concen: 46.572 ng
RT: 12.54 min Scan# 1777
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08




Instrument :
BNA_F
ClientSampleId :


Tot Ion:202 Resp: 1167111
Ion Ratio Lower Upper
202 100
101 11.6 0.0 31.0
203 17.8 0.0 38.4



#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.97 min Scan# 2020
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion:240 Resp: 350782
Ion Ratio Lower Upper
240 100
120 11.6 8.3 12.5
236 25.7 20.8 31.2





#77

Benzidine

Concen: 36.825 ng

RT: 12.66 min Scan# 1797

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Instrument :

BNA_F

ClientSampleId :

Tot Ion:184 Resp: 579847

Ion	Ratio	Lower	Upper
184	100		
185	15.4	13.6	20.4
183	10.8	9.2	13.8






Abundance

Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 41.896 ng

RT: 12.77 min Scan# 1816


Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Tgt Ion:202 Resp: 1182244

Ion	Ratio	Lower	Upper
202	100		
200	20.5	16.9	25.3
203	17.7	14.5	21.7






Abundance

Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#79

Terphenyl-d14

Concen: 89.795 ng

RT: 12.90 min Scan# 1839

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :


Tot Ion:244 Resp: 1527996

Ion Ratio Lower Upper

244 100

212 7.0 5.8 8.6




122 12.8 9.4 14.2



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 43.100 ng

RT: 13.37 min Scan# 1919

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Tgt Ion:149 Resp: 511915

Ion Ratio Lower Upper

149 100

91 68.4 55.0 82.4



206 21.1 16.9 25.3




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF1

Ion 206.00 (205.70 to 206.70): E






#81
Benzo(a)anthracene
Concen: 43.452 ng
RT: 13.96 min Scan# 2018
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
228	100		
226	26.9	22.2	33.2
229	19.9	16.4	24.6






Abundance

Ion 228.00 (227.70 to 228.70): E


Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82
3,3'-Dichlorobenzidine
Concen: 32.538 ng
RT: 13.92 min Scan# 2011
Delta R.T. -0.01 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion	Ratio	Lower	Upper
252	100		
254	64.7	51.9	77.9
126	10.1	7.9	11.9






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#83

Chrysene

Concen: 44.280 ng

RT: 14.00 min Scan# 2025

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :


Tot Ion:228 Resp: 984835

Ion Ratio Lower Upper

228 100

226 31.5 25.0 37.6




229 20.1 17.0 25.6



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Bis(2-ethylhexyl)phthalate

Concen: 44.356 ng

RT: 13.93 min Scan# 2014

Delta R.T. -0.01 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Tgt Ion:149 Resp: 689030

Ion Ratio Lower Upper

149 100

167 27.5 22.6 33.8



279 4.3 3.0 4.6




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E





#85

Di-n-octyl phthalate

Concen: 45.736 ng

RT: 14.54 min Scan# 2117

Delta R.T. -0.02 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :


Tot Ion:149 Resp: 1151725

Ion Ratio Lower Upper

149 100

167 1.5 1.3 1.9




43 11.6 10.1 15.1



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF1



#86

Indeno(1,2,3-cd)pyrene

Concen: 46.319 ng

RT: 16.89 min Scan# 2517

Delta R.T. -0.04 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Tgt Ion:276 Resp: 910460

Ion Ratio Lower Upper

276 100

138 24.7 20.2 30.2



277 25.0 20.0 30.0




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#87
Pervlene-d12
Concen: 20.000 ng
RT: 15.43 min Scan# 2268
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Instrument :
BNA_F
ClientSampleId :

Tot Ion:264 Resp: 335688




Ion	Ratio	Lower	Upper
264	100		
260	23.6	19.4	29.0
265	22.4	16.9	25.3




#88
Benzo(b)fluoranthene
Concen: 42.591 ng
RT: 15.00 min Scan# 2196
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion:252 Resp: 915970

Ion	Ratio	Lower	Upper
252	100		
253	22.7	17.8	26.8
125	10.9	8.3	12.5









#89
Benzo(k)fluoranthene
Concen: 48.824 ng
RT: 15.03 min Scan# 2201
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08




Instrument :
BNA_F
ClientSampleId :


Tot Ion:252 Resp: 964534
Ion Ratio Lower Upper
252 100
253 21.3 18.2 27.2
125 9.4 8.6 13.0



#90
Benzo(a)pyrene
Concen: 47.309 ng
RT: 15.37 min Scan# 2258
Delta R.T. -0.02 min
Lab File: BF114617.D
Acq: 25 May 2019 21:08

Tgt Ion:252 Resp: 895423
Ion Ratio Lower Upper
252 100
253 22.2 18.1 27.1
125 10.6 9.5 14.3





#91

Dibenzo(a,h)anthracene

Concen: 48.227 ng

RT: 16.89 min Scan# 2517

Delta R.T. -0.04 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08

Instrument :

BNA_F

ClientSampleId :


Tot Ion:278 Resp: 758489

Ion Ratio Lower Upper

278 100

139 16.6 13.5 20.3




279 24.3 19.0 28.6



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#92

Benzo(g,h,i)perylene

Concen: 49.972 ng

RT: 17.33 min Scan# 2591

Delta R.T. -0.04 min

Lab File: BF114617.D

Acq: 25 May 2019 21:08


Tgt Ion:276 Resp: 787623

Ion Ratio Lower Upper

276 100

277 24.1 19.0 28.4

138 21.7 17.8 26.6



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

