


Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF052519\  
 Data File : BF114623.D  
 Acq On : 26 May 2019 1:03  
 Operator : JU/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF051019.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Fri May 24 07:23:33 2019



AutoFind: Scans 1612, 1613, 1614; Background Corrected with Scan 1607

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 45.1         | 57770      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.6          | 786        | PASS                |
| 69             | 198             | 0.00            | 100             | 37.9         | 48496      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 264        | PASS                |
| 127            | 198             | 10              | 80              | 49.9         | 63890      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.7          | 958        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 128104     | PASS                |
| 199            | 198             | 5               | 9               | 6.9          | 8833       | PASS                |
| 275            | 198             | 10              | 60              | 27.6         | 35402      | PASS                |
| 365            | 198             | 1               | 100             | 3.1          | 3912       | PASS                |
| 441            | 198             | 0.01            | 100             | 13.6         | 17456      | PASS                |
| 442            | 198             | 50              | 100             | 85.6         | 109600     | PASS                |
| 443            | 442             | 15              | 24              | 18.7         | 20545      | PASS                |