

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF053019\
 Data File : BF114652.D
 Acq On : 30 May 2019 11:06
 Operator : HP/JU
 Sample : PB119288BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB119288BL

Quant Time: May 30 14:30:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF052919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 30 14:17:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.69	152	523328	20.00	ng	0.00
21) Naphthalene-d8	7.97	136	2021884	20.00	ng	0.00
39) Acenaphthene-d10	9.72	164	1029473	20.00	ng	0.00
64) Phenanthrene-d10	11.19	188	1963901	20.00	ng	0.00
76) Chrysene-d12	13.82	240	1533758	20.00	ng	0.00
87) Perylene-d12	15.20	264	1620354	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.31	112	2809789	94.27	ng	0.02
7) Phenol-d6	6.33	99	3415804	95.10	ng	0.00
23) Nitrobenzene-d5	7.25	82	2177373	67.19	ng	0.00
42) 2,4,6-Tribromophenol	10.50	330	1001421	102.09	ng	0.00
45) 2-Fluorobiphenyl	9.05	172	3881547	62.64	ng	0.00
79) Terphenyl-d14	12.78	244	4157566	50.97	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF053019\
Data File : BF114652.D
Acq On : 30 May 2019 11:06
Operator : HP/JU
Sample : PB119288BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB119288BL

Quant Time: May 30 14:30:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF052919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 14:17:39 2019
Response via : Initial Calibration

