

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF053119\
 Data File : BF114701.D
 Acq On : 31 May 2019 12:00
 Operator : HP/JU
 Sample : PB120137BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB120137BL

Quant Time: Jun 01 00:37:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF052919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 29 19:08:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	450544	20.00	ng	0.00
21) Naphthalene-d8	7.97	136	1766477	20.00	ng	0.00
39) Acenaphthene-d10	9.72	164	921963	20.00	ng	0.00
64) Phenanthrene-d10	11.20	188	1759401	20.00	ng	0.00
76) Chrysene-d12	13.82	240	1184472	20.00	ng	0.00
87) Perylene-d12	15.22	264	1003154	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	2777419	108.23	ng	0.02
7) Phenol-d6	6.34	99	3394775	109.78	ng	0.01
23) Nitrobenzene-d5	7.26	82	2222238	78.49	ng	0.00
42) 2,4,6-Tribromophenol	10.51	330	1063349	121.04	ng	0.00
45) 2-Fluorobiphenyl	9.06	172	3964858	79.81	ng	0.00
79) Terphenyl-d14	12.79	244	4088347	64.90	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF053119\
Data File : BF114701.D
Acq On : 31 May 2019 12:00
Operator : HP/JU
Sample : PB120137BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB120137BL

Quant Time: Jun 01 00:37:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF052919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 29 19:08:51 2019
Response via : Initial Calibration

