

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF060119\
 Data File : BF114759.D
 Acq On : 1 Jun 2019 16:12
 Operator : HP/JU
 Sample : K3097-22
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 0442-PR-5(PASSAIC)

Manual Integrations
 APPROVED

mohammad
 6/3/2019 8:51:27 AM

Quant Time: Jun 03 03:58:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF052919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 29 19:08:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.69	152	362918	20.00	ng	0.00
21) Naphthalene-d8	7.97	136	1267573	20.00	ng	0.00
39) Acenaphthene-d10	9.72	164	555362	20.00	ng	0.00
64) Phenanthrene-d10	11.19	188	1011843	20.00	ng	0.00
76) Chrysene-d12	13.82	240	938889	20.00	ng	0.00
87) Perylene-d12	15.20	264	937758	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	1774240	85.83	ng	0.00
7) Phenol-d6	6.33	99	2108847	84.66	ng	0.00
23) Nitrobenzene-d5	7.25	82	1260154	62.02	ng	0.00
42) 2,4,6-Tribromophenol	10.50	330	479624	90.64	ng	0.00
45) 2-Fluorobiphenyl	9.05	172	1967281	62.26	ng	0.00
79) Terphenyl-d14	12.77	244	2046245	40.98	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.44	163	372483	9.625	ng	99
75) Fluoranthene	12.40	202	225937	4.227	ng	94
78) Pyrene	12.62	202	248640	3.131	ng	95
84) Bis(2-ethylhexyl)phthalate	13.82	149	545605	11.057	ng	99
88) Benzo(b)fluoranthene	14.81	252	124562m	2.109	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF060119\
Data File : BF114759.D
Acq On : 1 Jun 2019 16:12
Operator : HP/JU
Sample : K3097-22
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
0442-PR-5(PASSAIC)

Manual Integrations
APPROVED

mohammad
6/3/2019 8:51:27 AM

Quant Time: Jun 03 03:58:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF052919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 29 19:08:51 2019
Response via : Initial Calibration

