

Data File : BF114797.D
Acq On : 4 Jun 2019 18:07
Operator : HP/JU
Sample : K2641-14MS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS

Quant Time: Jun 05 04:23:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 04 16:57:15 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	168638	20.00	ng	0.00
21) Naphthalene-d8	7.95	136	637975	20.00	ng	0.00
39) Acenaphthene-d10	9.69	164	305840	20.00	ng	0.00
64) Phenanthrene-d10	11.16	188	614789	20.00	ng	0.00
76) Chrysene-d12	13.79	240	510032	20.00	ng	0.00
87) Perylene-d12	15.16	264	615109	20.00	ng	-0.01

System Monitoring Compounds						
5) 2-Fluorophenol	5.27	112	1142531	118.38	ng	0.00
7) Phenol-d6	6.30	99	1275173	108.59	ng	0.00
23) Nitrobenzene-d5	7.22	82	899885	86.83	ng	-0.01
42) 2,4,6-Tribromophenol	10.47	330	438992	133.19	ng	0.00
45) 2-Fluorobiphenyl	9.02	172	1788190	109.53	ng	0.00
79) Terphenyl-d14	12.74	244	2194457	83.67	ng	0.00

Target Compounds					Qvalue
2) 1,4-Dioxane	2.31	88	167103	36.265 ng	98
3) Pyridine	3.00	79	366736	28.960 ng	97
4) n-Nitrosodimethylamine	2.89	42	176673	38.120 ng	92
6) Aniline	6.32	93	123984	7.441 ng	93
8) 2-Chlorophenol	6.45	128	475655	42.656 ng	99
9) Benzaldehyde	6.20	77	121417	14.905 ng	98
10) Phenol	6.31	94	468539	34.889 ng	# 76
11) bis(2-Chloroethyl)ether	6.40	93	431207	41.143 ng	97
12) 1,3-Dichlorobenzene	6.60	146	494266	38.039 ng	99
13) 1,4-Dichlorobenzene	6.68	146	506551	38.744 ng	98
14) 1,2-Dichlorobenzene	6.83	146	476530	40.357 ng	99
15) Benzyl Alcohol	6.80	79	382255	43.044 ng	98
16) 2,2'-oxybis(1-Chloropropan	6.95	45	581088	38.766 ng	99
17) 2-Methylphenol	6.92	107	364601	38.252 ng	99
18) Hexachloroethane	7.18	117	177193	36.007 ng	95
19) n-Nitroso-di-n-propylamine	7.08	70	305959	39.359 ng	99
20) 3+4-Methylphenols	7.07	107	449748	42.780 ng	# 85
22) Acetophenone	7.07	105	642693	45.757 ng	97
24) Nitrobenzene	7.24	77	472506	44.009 ng	99
25) Isophorone	7.49	82	851568	41.563 ng	99
26) 2-Nitrophenol	7.56	139	253820	43.649 ng	92
27) 2,4-Dimethylphenol	7.60	122	400392	45.302 ng	98
28) bis(2-Chloroethoxy)methane	7.70	93	539507	41.193 ng	98
29) 2,4-Dichlorophenol	7.80	162	397244	42.721 ng	97
30) 1,2,4-Trichlorobenzene	7.89	180	442626	41.486 ng	99
31) Naphthalene	7.97	128	1383817	48.180 ng	96
32) Benzoic acid	7.69	122	175005	27.186 ng	99
33) 4-Chloroaniline	8.02	127	80352	5.864 ng	95
34) Hexachlorobutadiene	8.10	225	239881	39.546 ng	98
35) Caprolactam	8.36	113	97299	32.032 ng	93
36) 4-Chloro-3-methylphenol	8.50	107	390917	42.248 ng	100
37) 2-Methylnaphthalene	8.66	142	926102	46.487 ng	99
38) 1-Methylnaphthalene	8.75	142	879084	46.531 ng	100
40) 1,2,4,5-Tetrachlorobenzene	8.82	216	403682	45.801 ng	99

Data File : BF114797.D
Acq On : 4 Jun 2019 18:07
Operator : HP/JU
Sample : K2641-14MS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS


Quant Time: Jun 05 04:23:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 04 16:57:15 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	8.82	237	404321	75.637	ng	99
43) 2,4,6-Trichlorophenol	8.93	196	285408	41.644	ng	99
44) 2,4,5-Trichlorophenol	8.97	196	304472	44.112	ng	100
46) 1,1'-Biphenyl	9.12	154	1106785	48.827	ng	96
47) 2-Chloronaphthalene	9.14	162	866456	44.985	ng	97
48) 2-Nitroaniline	9.23	65	241062	41.939	ng	91
49) Acenaphthylene	9.55	152	1366893	48.937	ng	96
50) Dimethylphthalate	9.42	163	1007371	45.098	ng	99
51) 2,6-Dinitrotoluene	9.47	165	225922	43.006	ng	97
52) Acenaphthene	9.72	154	885704	48.407	ng	98
53) 3-Nitroaniline	9.63	138	114107	18.594	ng	93
54) 2,4-Dinitrophenol	9.75	184	173474	72.976	ng	95
55) Dibenzofuran	9.90	168	1216043	47.795	ng	93
56) 4-Nitrophenol	9.80	139	377510	83.951	ng	99
57) 2,4-Dinitrotoluene	9.87	165	309062	46.800	ng	95
58) Fluorene	10.23	166	884632	56.629	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.02	232	258818	45.769	ng	99
60) Diethylphthalate	10.12	149	999747	44.603	ng	97
61) 4-Chlorophenyl-phenylether	10.23	204	437718	44.918	ng	97
62) 4-Nitroaniline	10.24	138	236866	39.794	ng	95
63) Azobenzene	10.39	77	888117	46.548	ng	96
65) 4,6-Dinitro-2-methylphenol	10.28	198	116652	36.527	ng	92
66) n-Nitrosodiphenylamine	10.34	169	830667	44.684	ng	99
67) 4-Bromophenyl-phenylether	10.72	248	285937	40.951	ng	100
68) Hexachlorobenzene	10.79	284	313094	40.971	ng	96
69) Atrazine	10.87	200	274100	42.395	ng	98
70) Pentachlorophenol	10.97	266	350960	79.445	ng	99
71) Phenanthrene	11.19	178	1445247	49.243	ng	97
72) Anthracene	11.24	178	1478934	47.882	ng	96
73) Carbazole	11.39	167	1359580	51.348	ng	96
74) Di-n-butylphthalate	11.74	149	1679923	48.238	ng	97
75) Fluoranthene	12.37	202	1559315	51.046	ng	97
77) Benzidine	12.49	184	187101	7.410	ng	# 81
78) Pyrene	12.59	202	1579122	35.690	ng	97
80) Butylbenzylphthalate	13.23	149	788827	35.305	ng	94
81) Benzo(a)anthracene	13.77	228	1524939	38.799	ng	97
82) 3,3'-Dichlorobenzidine	13.74	252	230307	14.451	ng	99
83) Chrysene	13.81	228	1507415	39.426	ng	98
84) Bis(2-ethylhexyl)phthalate	13.79	149	1103994	39.075	ng	98
85) Di-n-octyl phthalate	14.40	149	1896350	39.502	ng	98
86) Indeno(1,2,3-cd)pyrene	16.47	276	1839008	42.155	ng	100
88) Benzo(b)fluoranthene	14.78	252	1761851	44.992	ng	99
89) Benzo(k)fluoranthene	14.81	252	1399735	42.092	ng	97
90) Benzo(a)pyrene	15.11	252	1561797	46.080	ng	99
91) Dibenzo(a,h)anthracene	16.49	278	1537388	47.897	ng	99
92) Benzo(g,h,i)perylene	16.86	276	1544954	47.864	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 05 04:23:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 04 16:57:15 2019
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.66 min Scan# 778
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS




Tgt Ion:152 Resp: 168638
Ion Ratio Lower Upper
152 100
150 153.8 124.0 186.0
115 57.9 46.3 69.5




#2

1,4-Dioxane
Concen: 36.265 ng
RT: 2.31 min Scan# 38
Delta R.T. 0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Tgt Ion: 88 Resp: 167103
Ion Ratio Lower Upper
88 100
58 67.2 53.8 80.8
43 28.8 20.7 31.1





#3

Pyridine

Concen: 28.960 ng

RT: 3.00 min Scan# 156

Delta R.T. 0.03 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion: 79 Resp: 366736

Ion Ratio Lower Upper

79 100

52 66.8 51.0 76.4




51 32.1 25.7 38.5



Abundance Ion 79.00 (78.70 to 79.70): BF1

Ion 52.00 (51.70 to 52.70): BF1

Ion 51.00 (50.70 to 51.70): BF1



#4

n-Nitrosodimethylamine

Concen: 38.120 ng

RT: 2.89 min Scan# 136

Delta R.T. -0.04 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion: 42 Resp: 176673

Ion Ratio Lower Upper

42 100

74 146.9 126.2 189.2



44 7.6 5.9 8.9




Abundance Ion 42.00 (41.70 to 42.70): BF1

Ion 74.00 (73.70 to 74.70): BF1

Ion 44.00 (43.70 to 44.70): BF1





#5

2-Fluorophenol

Concen: 118.384 ng

RT: 5.27 min Scan# 541

Delta R.T. 0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion:112 Resp: 1142531

Ion Ratio Lower Upper

112 100

64 50.9 44.6 66.8




63 25.7 23.0 34.4



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1



#6

Aniline

Concen: 7.441 ng

RT: 6.32 min Scan# 720

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion: 93 Resp: 123984

Ion Ratio Lower Upper

93 100

66 46.1 32.8 49.2



65 24.0 16.9 25.3




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 66.00 (65.70 to 66.70): BF1

Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 108.586 ng

RT: 6.30 min Scan# 716

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion: 99 Resp: 1275173

Ion Ratio Lower Upper

99 100

42 15.7 13.3 19.9




71 31.9 26.3 39.5



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#8

2-Chlorophenol

Concen: 42.656 ng

RT: 6.45 min Scan# 741

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion: 128 Resp: 475655

Ion Ratio Lower Upper

128 100

130 32.1 12.7 52.7



64 42.2 22.5 62.5




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF1





#9

Benzaldehyde

Concen: 14.905 ng

RT: 6.20 min Scan# 700

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion: 77 Resp: 121417

Ion Ratio Lower Upper

77 100

105 109.0 85.8 125.8




106 100.4 81.6 121.6



Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 106.00 (105.70 to 106.70): BF1



#10

Phenol

Concen: 34.889 ng

RT: 6.31 min Scan# 718

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion: 94 Resp: 468539

Ion Ratio Lower Upper

94 100

65 26.5 16.1 56.1



66 34.9 36.3 76.3#




Abundance Ion 94.00 (93.70 to 94.70): BF1

Ion 65.00 (64.70 to 65.70): BF1

Ion 66.00 (65.70 to 66.70): BF1









#11
bis(2-Chloroethyl)ether
Concen: 41.143 ng
RT: 6.40 min Scan# 733
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS

Tgt Ion	Ratio	Lower	Upper
93	100		
63	72.0	49.6	89.6
95	31.9	13.6	53.6




Abundance Ion 93.00 (92.70 to 93.70): BF1
Ion 63.00 (62.70 to 63.70): BF1
Ion 95.00 (94.70 to 95.70): BF1






#12
1,3-Dichlorobenzene
Concen: 38.039 ng
RT: 6.60 min Scan# 768
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Tgt Ion	Ratio	Lower	Upper
146	100		
148	64.5	52.7	79.1
75	27.1	22.0	33.0



Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 75.00 (74.70 to 75.70): BF1





#13

1,4-Dichlorobenzene

Concen: 38.744 ng

RT: 6.68 min Scan# 781

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS





Tgt Ion:146 Resp: 506551

Ion Ratio Lower Upper

146 100

148 64.5 53.0 79.4

111 39.1 32.2 48.4



#14

1,2-Dichlorobenzene

Concen: 40.357 ng

RT: 6.83 min Scan# 807

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07




Tgt Ion:146 Resp: 476530


Ion Ratio Lower Upper

146 100

148 65.7 52.5 78.7

111 42.0 34.8 52.2





#15

Benzyl Alcohol

Concen: 43.044 ng

RT: 6.80 min Scan# 802

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS





Tgt Ion: 79 Resp: 382255

Ion Ratio Lower Upper

79 100

108 81.5 66.8 100.2

77 61.3 50.6 76.0



#16

2,2'-oxybis(1-Chloropropane)

Concen: 38.766 ng

RT: 6.95 min Scan# 826

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07




Tgt Ion: 45 Resp: 581088


Ion Ratio Lower Upper

45 100

77 15.2 0.0 35.8

79 11.7 0.0 32.0





#17

2-Methylphenol

Concen: 38.252 ng

RT: 6.92 min Scan# 822

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS

Tgt Ion:107 Resp: 364601


Ion Ratio Lower Upper

107 100

108 114.1 90.2 135.2

77 42.4 34.5 51.7

79 41.2 33.8 50.6






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF1

Ion 79.00 (78.70 to 79.70): BF1



#18

Hexachloroethane

Concen: 36.007 ng

RT: 7.18 min Scan# 866

Delta R.T. -0.00 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion:117 Resp: 177193

Ion Ratio Lower Upper

117 100

119 96.3 77.0 115.4



201 104.7 76.0 114.0




Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E






#19
n-Nitroso-di-n-propylamine
Concen: 39.359 ng
RT: 7.08 min Scan# 849
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS

Tgt Ion:	70	Resp:	305959
Ion	Ratio	Lower	Upper
70	100		
42	52.8	41.8	62.8
101	10.4	8.2	12.4
130	24.8	18.5	27.7






Abundance

Ion 70.00 (69.70 to 70.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E



#20
3+4-Methylphenols
Concen: 42.780 ng
RT: 7.07 min Scan# 848
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Tgt Ion:	107	Resp:	449748
Ion	Ratio	Lower	Upper
107	100		
108	89.1	73.5	113.5
77	113.8	67.2	107.2#
79	27.0	10.6	50.6




Abundance


Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF1

Ion 79.00 (78.70 to 79.70): BF1





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 7.95 min Scan# 996

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS

Tgt Ion:136 Resp: 637975



Ion Ratio Lower Upper

136 100

137 11.0 9.1 13.7

54 6.4 5.4 8.0

68 5.4 4.6 6.8





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#22

Acetophenone

Concen: 45.757 ng

RT: 7.07 min Scan# 847

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Tgt Ion:105 Resp: 642693



Ion Ratio Lower Upper

105 100

71 2.8 2.0 3.0

51 28.8 25.4 38.0

120 23.0 18.7 28.1





Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF1

Ion 51.00 (50.70 to 51.70): BF1

Ion 120.00 (119.70 to 120.70): E





#23

Nitrobenzene-d5

Concen: 86.830 ng

RT: 7.22 min Scan# 873

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion: 82 Resp: 899885

Ion Ratio Lower Upper

82 100

128 50.3 41.8 62.6




54 45.8 37.9 56.9



Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 128.00 (127.70 to 128.70): BF1

Ion 54.00 (53.70 to 54.70): BF1



#24

Nitrobenzene

Concen: 44.009 ng

RT: 7.24 min Scan# 876

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion: 77 Resp: 472506

Ion Ratio Lower Upper

77 100

123 50.9 41.5 62.3



65 13.1 10.8 16.2




Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 123.00 (122.70 to 123.70): BF1

Ion 65.00 (64.70 to 65.70): BF1





#25

Isophorone

Concen: 41.563 ng

RT: 7.49 min Scan# 918

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion: 82 Resp: 851568

Ion Ratio Lower Upper




82 100

95 7.0 5.8 8.8

138 20.8 16.9 25.3



Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 95.00 (94.70 to 95.70): BF1
Ion 138.00 (137.70 to 138.70): BF1



#26

2-Nitrophenol

Concen: 43.649 ng

RT: 7.56 min Scan# 930

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion: 139 Resp: 253820

Ion Ratio Lower Upper



139 100


109 26.7 20.4 30.6

65 45.2 30.5 45.7



Abundance Ion 139.00 (138.70 to 139.70): BF1
Ion 109.00 (108.70 to 109.70): BF1
Ion 65.00 (64.70 to 65.70): BF1





#27

2,4-Dimethylphenol

Concen: 45.302 ng

RT: 7.60 min Scan# 938

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion:122 Resp: 400392

Ion Ratio Lower Upper

122 100

107 112.8 87.9 131.9




121 59.1 46.4 69.6



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 41.193 ng

RT: 7.70 min Scan# 954

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion: 93 Resp: 539507

Ion Ratio Lower Upper

93 100

95 32.6 26.9 40.3



123 11.9 10.4 15.6




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 123.00 (122.70 to 123.70): E









#29
2,4-Dichlorophenol
Concen: 42.721 ng
RT: 7.80 min Scan# 971
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS

Tgt Ion:162 Resp: 397244
Ion Ratio Lower Upper
162 100
164 66.0 45.1 85.1
98 37.7 14.6 54.6




Abundance Ion 162.00 (161.70 to 162.70): E
Ion 164.00 (163.70 to 164.70): E
Ion 98.00 (97.70 to 98.70): BF1






#30
1,2,4-Trichlorobenzene
Concen: 41.486 ng
RT: 7.89 min Scan# 986
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Tgt Ion:180 Resp: 442626
Ion Ratio Lower Upper
180 100
182 94.5 76.5 114.7
145 31.1 25.8 38.8



Abundance Ion 180.00 (179.70 to 180.70): E
Ion 182.00 (181.70 to 182.70): E
Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 48.180 ng

RT: 7.97 min Scan# 1000

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion:128 Resp: 1383817

Ion Ratio Lower Upper

128 100

129 10.9 10.0 15.0




127 13.4 12.2 18.2



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 27.186 ng

RT: 7.69 min Scan# 952

Delta R.T. -0.06 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion:122 Resp: 175005

Ion Ratio Lower Upper

122 100

105 117.9 98.7 138.7



77 85.7 64.8 104.8




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF1






#33
4-Chloroaniline
Concen: 5.864 ng
RT: 8.02 min Scan# 1008
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS

Tgt Ion:	127	Resp:	80352
Ion	Ratio	Lower	Upper
127	100		
129	31.6	26.9	40.3
65	23.4	21.7	32.5
92	15.9	14.8	22.2






Abundance

Ion 127.00 (126.70 to 127.70): E

Ion 129.00 (128.70 to 129.70): E


Ion 65.00 (64.70 to 65.70): BF1

Ion 92.00 (91.70 to 92.70): BF1



#34
Hexachlorobutadiene
Concen: 39.546 ng
RT: 8.10 min Scan# 1022
Delta R.T. -0.00 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Tgt Ion:	225	Resp:	239881
Ion	Ratio	Lower	Upper
225	100		
223	61.9	51.0	76.6
227	64.3	52.6	78.8






Abundance

Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 32.032 ng
 RT: 8.36 min Scan# 1066
 Delta R.T. -0.05 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07




Instrument :
 BNA_F
 ClientSampleId :
 CL-02-053119-AMS


Tgt Ion:113 Resp: 97299
 Ion Ratio Lower Upper
 113 100
 55 148.8 121.1 161.1
 56 110.3 82.6 122.6



#36
 4-Chloro-3-methylphenol
 Concen: 42.248 ng
 RT: 8.50 min Scan# 1090
 Delta R.T. -0.01 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07

Tgt Ion:107 Resp: 390917
 Ion Ratio Lower Upper
 107 100
 144 28.0 22.1 33.1
 142 85.1 68.4 102.6





#37

2-Methylnaphthalene

Concen: 46.487 ng

RT: 8.66 min Scan# 1117

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion:142 Resp: 926102

Ion Ratio Lower Upper

142 100

141 89.5 71.6 107.4

115 32.3 28.2 42.4






Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

1-Methylnaphthalene

Concen: 46.531 ng

RT: 8.75 min Scan# 1133

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion:142 Resp: 879084

Ion Ratio Lower Upper

142 100

141 91.8 73.8 110.6

116 3.6 3.3 4.9






Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 116.00 (115.70 to 116.70): E









#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.69 min Scan# 1293
 Delta R.T. -0.01 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07




Instrument :
 BNA_F
 ClientSampleId :
 CL-02-053119-AMS


Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.0	82.3	123.5
160	45.5	36.7	55.1



#40
 1,2,4,5-Tetrachlorobenzene
 Concen: 45.801 ng
 RT: 8.82 min Scan# 1145
 Delta R.T. -0.01 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07

Tgt Ion	Ratio	Lower	Upper
216	100		
214	78.5	63.4	95.2
179	22.2	18.3	27.5
108	20.6	15.5	23.3





#41

Hexachlorocyclopentadiene

Concen: 75.637 ng

RT: 8.82 min Scan# 1144

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion: 237 Resp: 404321

Ion Ratio Lower Upper




237 100

235 63.2 44.0 84.0

272 12.6 0.0 33.4



Abundance Ion 236.80 (236.50 to 237.50): E
Ion 234.90 (234.60 to 235.60): E
Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 133.193 ng

RT: 10.47 min Scan# 1426

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion: 330 Resp: 438992

Ion Ratio Lower Upper



330 100


332 95.7 77.3 115.9

141 33.5 26.2 39.4



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#43
 2,4,6-Trichlorophenol
 Concen: 41.644 ng
 RT: 8.93 min Scan# 1164
 Delta R.T. -0.01 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07

Instrument :
 BNA_F
 ClientSampleId :
 CL-02-053119-AMS

Tgt Ion:196 Resp: 285408
 Ion Ratio Lower Upper
 196 100
 198 95.5 77.1 115.7
 200 30.8 24.6 36.8




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E






#44
 2,4,5-Trichlorophenol
 Concen: 44.112 ng
 RT: 8.97 min Scan# 1170
 Delta R.T. -0.01 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07

Tgt Ion:196 Resp: 304472
 Ion Ratio Lower Upper
 196 100
 198 95.2 76.2 114.4
 97 47.0 38.2 57.4
 132 29.3 23.7 35.5



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF1
 Ion 132.00 (131.70 to 132.70): E









#45
2-Fluorobiphenyl
Concen: 109.527 ng
RT: 9.02 min Scan# 1179
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07




Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS


Tgt Ion:172 Resp: 1788190
Ion Ratio Lower Upper
172 100
171 36.5 32.3 48.5
170 25.2 22.4 33.6



#46
1,1'-Biphenyl
Concen: 48.827 ng
RT: 9.12 min Scan# 1196
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Tgt Ion:154 Resp: 1106785
Ion Ratio Lower Upper
154 100
153 41.3 23.9 63.9
76 11.8 0.0 33.2









#47
 2-Chloronaphthalene
 Concen: 44.985 ng
 RT: 9.14 min Scan# 1199
 Delta R.T. -0.01 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07

Instrument :
 BNA_F
 ClientSampleId :
 CL-02-053119-AMS

Tgt Ion	Ratio	Lower	Upper
162	100		
127	41.3	34.6	51.8
164	32.9	27.8	41.8




Abundance Ion 162.00 (161.70 to 162.70): E
 1500000 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E






#48
 2-Nitroaniline
 Concen: 41.939 ng
 RT: 9.23 min Scan# 1215
 Delta R.T. -0.01 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07

Tgt Ion	Ratio	Lower	Upper
65	100		
92	78.7	60.4	90.6
138	140.9	101.2	151.8



Abundance Ion 65.00 (64.70 to 65.70): BF1
 500000 Ion 92.00 (91.70 to 92.70): BF1
 Ion 138.00 (137.70 to 138.70): E





#49

Acenaphthylene

Concen: 48.937 ng

RT: 9.55 min Scan# 1269

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion:152 Resp: 1366893

Ion Ratio Lower Upper




152 100

151 21.0 18.3 27.5

153 13.5 12.3 18.5



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 45.098 ng

RT: 9.42 min Scan# 1247

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion:163 Resp: 1007371

Ion Ratio Lower Upper



163 100


194 3.9 3.4 5.2

164 10.3 8.6 13.0



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E





#51

2,6-Dinitrotoluene

Concen: 43.006 ng

RT: 9.47 min Scan# 1256

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion:165 Resp: 225922

Ion Ratio Lower Upper

165 100

63 46.7 35.1 52.7




89 49.0 38.2 57.2



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1



#52

Acenaphthene

Concen: 48.407 ng

RT: 9.72 min Scan# 1298

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion:154 Resp: 885704

Ion Ratio Lower Upper

154 100

153 112.6 88.5 132.7



152 54.3 44.1 66.1




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E





#53

3-Nitroaniline

Concen: 18.594 ng

RT: 9.63 min Scan# 1283

Delta R.T. -0.02 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion:138 Resp: 114107

Ion Ratio Lower Upper

138 100

108 9.9 7.9 11.9




92 107.4 79.9 119.9



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF1



#54

2,4-Dinitrophenol

Concen: 72.976 ng

RT: 9.75 min Scan# 1302

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion:184 Resp: 173474

Ion Ratio Lower Upper

184 100

63 48.4 44.1 66.1



154 58.9 48.1 72.1




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 154.00 (153.70 to 154.70): E









#55
Dibenzenofuran
Concen: 47.795 ng
RT: 9.90 min Scan# 1328
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS

Tgt Ion:168 Resp: 1216043
Ion Ratio Lower Upper
168 100
139 37.0 33.8 50.6
169 13.3 11.9 17.9




Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E






#56
4-Nitrophenol
Concen: 83.951 ng
RT: 9.80 min Scan# 1311
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Tgt Ion:139 Resp: 377510
Ion Ratio Lower Upper
139 100
109 62.1 43.5 83.5
65 83.0 63.2 103.2



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF1









#57
 2,4-Dinitrotoluene
 Concen: 46.800 ng
 RT: 9.87 min Scan# 1324
 Delta R.T. -0.01 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07

Instrument :
 BNA_F
 ClientSampleId :
 CL-02-053119-AMS

Tgt Ion:	165	Resp:	309062
Ion	Ratio	Lower	Upper
165	100		
63	34.8	27.0	40.4
89	63.4	47.0	70.6
182	2.8	2.2	3.2




Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BF1
 Ion 89.00 (88.70 to 89.70): BF1
 Ion 182.00 (181.70 to 182.70): E






#58
 Fluorene
 Concen: 56.629 ng
 RT: 10.23 min Scan# 1385
 Delta R.T. -0.01 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07

Tgt Ion:	166	Resp:	884632
Ion	Ratio	Lower	Upper
166	100		
165	98.6	77.9	116.9
167	14.1	11.3	16.9



Abundance Ion 166.00 (165.70 to 166.70): E
 Ion 165.00 (164.70 to 165.70): E
 Ion 167.00 (166.70 to 167.70): E









#59
2,3,4,6-Tetrachlorophenol
Concen: 45.769 ng
RT: 10.02 min Scan# 1348
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS

Tgt Ion: 232 Resp: 258818
Ion Ratio Lower Upper
232 100
131 46.3 36.8 55.2
130 2.3 1.8 2.6
166 30.4 24.6 37.0




Abundance Ion 232.00 (231.70 to 232.70): E
Ion 131.00 (130.70 to 131.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 166.00 (165.70 to 166.70): E






#60
Diethylphthalate
Concen: 44.603 ng
RT: 10.12 min Scan# 1366
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Tgt Ion: 149 Resp: 999747
Ion Ratio Lower Upper
149 100
177 22.6 19.3 28.9
150 12.6 10.7 16.1



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E









#61
 4-Chlorophenyl-phenylether
 Concen: 44.918 ng
 RT: 10.23 min Scan# 1385
 Delta R.T. -0.01 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07

Instrument :
 BNA_F
 ClientSampleId :
 CL-02-053119-AMS

Tgt Ion	Ratio	Lower	Upper
204	100		
206	32.4	26.5	39.7
141	60.2	51.0	76.4




Abundance Ion 204.00 (203.70 to 204.70): E
 Ion 206.00 (205.70 to 206.70): E
 Ion 141.00 (140.70 to 141.70): E






#62
 4-Nitroaniline
 Concen: 39.794 ng
 RT: 10.24 min Scan# 1387
 Delta R.T. -0.01 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07

Tgt Ion	Ratio	Lower	Upper
138	100		
92	43.8	26.5	66.5
108	61.4	45.4	85.4



Abundance Ion 138.00 (137.70 to 138.70): E
 Ion 92.00 (91.70 to 92.70): BF1
 Ion 108.00 (107.70 to 108.70): E





#63

Azobenzene

Concen: 46.548 ng

RT: 10.39 min Scan# 1412

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS

Tgt Ion: 77 Resp: 888117


Ion Ratio Lower Upper

77 100

182 29.4 11.2 51.2

105 24.5 5.7 45.7

51 25.4 8.0 48.0






Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 182.00 (181.70 to 182.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 51.00 (50.70 to 51.70): BF1



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.16 min Scan# 1543

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion: 188 Resp: 614789

Ion Ratio Lower Upper

188 100

94 9.1 7.6 11.4



80 10.0 8.3 12.5




Abundance Ion 188.00 (187.70 to 188.70): BF1

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1





#65

4,6-Dinitro-2-methylphenol

Concen: 36.527 ng

RT: 10.28 min Scan# 1393

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion:198 Resp: 116652

Ion Ratio Lower Upper

198 100

51 30.2 14.9 54.9




105 35.7 21.0 61.0



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BF1

Ion 105.00 (104.70 to 105.70): E



#66

n-Nitrosodiphenylamine

Concen: 44.684 ng

RT: 10.34 min Scan# 1404

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion:169 Resp: 830667

Ion Ratio Lower Upper

169 100

168 67.0 54.3 81.5



167 36.5 29.9 44.9




Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E









#67
4-Bromophenyl-phenylether
Concen: 40.951 ng
RT: 10.72 min Scan# 1467
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS

Tgt Ion:248 Resp: 285937
Ion Ratio Lower Upper
248 100
250 96.6 77.5 116.3
141 77.7 61.8 92.8




Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E






#68
Hexachlorobenzene
Concen: 40.971 ng
RT: 10.79 min Scan# 1479
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Tgt Ion:284 Resp: 313094
Ion Ratio Lower Upper
284 100
142 29.1 25.3 37.9
249 29.5 24.9 37.3



Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E









#69
Atrazine
Concen: 42.395 ng
RT: 10.87 min Scan# 1494
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07




Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS


Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.4	8.0	48.0
215	47.6	28.7	68.7



#70
Pentachlorophenol
Concen: 79.445 ng
RT: 10.97 min Scan# 1511
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Tgt Ion	Ratio	Lower	Upper
266	100		
268	64.9	51.0	76.6
264	62.6	49.6	74.4









#71
Phenanthrene
Concen: 49.243 ng
RT: 11.19 min Scan# 1547
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07




Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS


Tgt Ion:178 Resp: 1445247
Ion Ratio Lower Upper
178 100
176 20.4 17.7 26.5
179 16.4 13.7 20.5



#72
Anthracene
Concen: 47.882 ng
RT: 11.24 min Scan# 1556
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Tgt Ion:178 Resp: 1478934
Ion Ratio Lower Upper
178 100
176 19.8 17.4 26.0
179 16.1 14.2 21.4









#73
 Carbazole
 Concen: 51.348 ng
 RT: 11.39 min Scan# 1582
 Delta R.T. -0.01 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07




Instrument :
 BNA_F
 ClientSampleId :
 CL-02-053119-AMS


Tgt Ion:167 Resp: 1359580
 Ion Ratio Lower Upper
 167 100
 166 22.1 19.1 28.7
 139 13.4 12.4 18.6



#74
 Di-n-butylphthalate
 Concen: 48.238 ng
 RT: 11.74 min Scan# 1641
 Delta R.T. -0.01 min
 Lab File: BF114797.D
 Acq: 4 Jun 2019 18:07

Tgt Ion:149 Resp: 1679923
 Ion Ratio Lower Upper
 149 100
 150 9.7 8.9 13.3
 104 4.6 4.3 6.5









#75
Fluoranthene
Concen: 51.046 ng
RT: 12.37 min Scan# 1748
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07




Instrument :
BNA_F
ClientSampleId :
CL-02-053119-AMS


Tgt Ion: 202 Resp: 1559315
Ion Ratio Lower Upper
202 100
101 12.1 0.0 32.8
203 18.5 0.2 40.2



#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.79 min Scan# 1989
Delta R.T. -0.01 min
Lab File: BF114797.D
Acq: 4 Jun 2019 18:07

Tgt Ion: 240 Resp: 510032
Ion Ratio Lower Upper
240 100
120 10.5 9.0 13.4
236 26.2 21.4 32.0





#77

Benzidine

Concen: 7.410 ng

RT: 12.49 min Scan# 1768

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion:184 Resp: 187101

Ion Ratio Lower Upper

184 100

185 30.9 14.1 21.1#




183 11.6 9.9 14.9



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 35.690 ng

RT: 12.59 min Scan# 1786

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion:202 Resp: 1579122

Ion Ratio Lower Upper

202 100

200 22.3 19.2 28.8



203 18.5 15.9 23.9




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#79

Terphenyl-d14

Concen: 83.675 ng

RT: 12.74 min Scan# 1812

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion:244 Resp: 2194457

Ion Ratio Lower Upper

244 100

212 7.2 6.5 9.7




122 12.3 10.5 15.7



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 35.305 ng

RT: 13.23 min Scan# 1894

Delta R.T. -0.00 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion:149 Resp: 788827

Ion Ratio Lower Upper

149 100

91 62.2 54.6 82.0



206 21.2 17.2 25.8




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF1

Ion 206.00 (205.70 to 206.70): E





#81

Benzo(a)anthracene

Concen: 38.799 ng

RT: 13.77 min Scan# 1987

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS





Tgt Ion:228 Resp: 1524939

Ion Ratio Lower Upper

228 100

226 27.7 23.4 35.2

229 20.3 17.4 26.0



#82

3,3'-Dichlorobenzidine

Concen: 14.451 ng

RT: 13.74 min Scan# 1981

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07




Tgt Ion:252 Resp: 230307


Ion Ratio Lower Upper

252 100

254 64.5 51.9 77.9

126 11.2 10.5 15.7





#83

Chrysene

Concen: 39.426 ng

RT: 13.81 min Scan# 1993

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion:228 Resp: 1507415

Ion Ratio Lower Upper

228 100

226 30.9 26.0 39.0




229 20.5 17.1 25.7



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Bis(2-ethylhexyl)phthalate

Concen: 39.075 ng

RT: 13.79 min Scan# 1990

Delta R.T. -0.00 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion:149 Resp: 1103994

Ion Ratio Lower Upper

149 100

167 27.5 22.7 34.1



279 4.1 3.1 4.7




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E





#85

Di-n-octyl phthalate

Concen: 39.502 ng

RT: 14.40 min Scan# 2093

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS





Tgt Ion:149 Resp: 1896350

Ion Ratio Lower Upper

149 100

167 1.6 1.4 2.0

43 8.1 7.0 10.6



#86

Indeno(1,2,3-cd)pyrene

Concen: 42.155 ng

RT: 16.47 min Scan# 2445

Delta R.T. -0.02 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07




Tgt Ion:276 Resp: 1839008


Ion Ratio Lower Upper

276 100

138 28.0 22.2 33.4

277 25.5 20.6 30.8





#87

Perylene-d12

Concen: 20.000 ng

RT: 15.16 min Scan# 2223

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion:264 Resp: 615109

Ion Ratio Lower Upper

264 100

260 24.3 20.1 30.1




265 22.5 17.7 26.5



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#88

Benzo(b)fluoranthene

Concen: 44.992 ng

RT: 14.78 min Scan# 2158

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion:252 Resp: 1761851

Ion Ratio Lower Upper

252 100

253 22.9 19.0 28.6



125 11.5 9.4 14.0




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Benzo(k)fluoranthene

Concen: 42.092 ng

RT: 14.81 min Scan# 2163

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS





Tgt Ion:252 Resp: 1399735

Ion Ratio Lower Upper

252 100

253 23.0 19.3 28.9

125 11.3 10.1 15.1



#90

Benzo(a)pyrene

Concen: 46.080 ng

RT: 15.11 min Scan# 2214

Delta R.T. -0.01 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07




Tgt Ion:252 Resp: 1561797


Ion Ratio Lower Upper

252 100

253 22.5 18.5 27.7

125 12.4 10.1 15.1





#91

Dibenzo(a,h)anthracene

Concen: 47.897 ng

RT: 16.49 min Scan# 2448

Delta R.T. -0.02 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07

Instrument :

BNA_F

ClientSampleId :

CL-02-053119-AMS


Tgt Ion:278 Resp: 1537388

Ion Ratio Lower Upper

278 100

139 18.3 14.6 22.0




279 24.3 20.1 30.1



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#92

Benzo(g,h,i)perylene

Concen: 47.864 ng

RT: 16.86 min Scan# 2512

Delta R.T. -0.02 min

Lab File: BF114797.D

Acq: 4 Jun 2019 18:07


Tgt Ion:276 Resp: 1544954

Ion Ratio Lower Upper

276 100

277 24.2 19.8 29.6

138 23.7 19.4 29.2



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

