

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF060419\
 Data File : BF114801.D
 Acq On : 4 Jun 2019 19:54
 Operator : HP/JU
 Sample : K3135-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ROLL-OFF-COMP

Quant Time: Jun 05 04:45:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF060419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 04 16:57:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	229317	20.00	ng	0.00
21) Naphthalene-d8	7.94	136	877729	20.00	ng	-0.01
39) Acenaphthene-d10	9.69	164	450296	20.00	ng	-0.01
64) Phenanthrene-d10	11.16	188	910636	20.00	ng	0.00
76) Chrysene-d12	13.78	240	763466	20.00	ng	-0.01
87) Perylene-d12	15.16	264	808486	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.27	112	1558505	118.76	ng	0.00
7) Phenol-d6	6.30	99	1724753	108.01	ng	0.00
23) Nitrobenzene-d5	7.22	82	1270529	89.11	ng	-0.01
42) 2,4,6-Tribromophenol	10.47	330	646496	133.23	ng	0.00
45) 2-Fluorobiphenyl	9.02	172	2425957	100.92	ng	0.00
79) Terphenyl-d14	12.74	244	2999537	76.41	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF060419\
Data File : BF114801.D
Acq On : 4 Jun 2019 19:54
Operator : HP/JU
Sample : K3135-02
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ROLL-OFF-COMP

Quant Time: Jun 05 04:45:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF060419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 04 16:57:15 2019
Response via : Initial Calibration

