


Data Path : Z:\SVOASRV\HPCHEM1\BNA\_F\DATA\BF060419\  
 Data File : BF114787.D  
 Acq On : 4 Jun 2019 10:46  
 Operator : HP/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF060419.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Tue Jun 04 16:57:15 2019



AutoFind: Scans 1590, 1591, 1592; Background Corrected with Scan 1584

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 32.2         | 140621     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.8          | 2627       | PASS                |
| 69             | 198             | 0.00            | 100             | 33.5         | 146228     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.4          | 563        | PASS                |
| 127            | 198             | 10              | 80              | 48.2         | 210240     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.5          | 2154       | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 436480     | PASS                |
| 199            | 198             | 5               | 9               | 6.8          | 29512      | PASS                |
| 275            | 198             | 10              | 60              | 27.1         | 118434     | PASS                |
| 365            | 198             | 1               | 100             | 3.2          | 14080      | PASS                |
| 441            | 198             | 0.01            | 100             | 14.0         | 61317      | PASS                |
| 442            | 198             | 50              | 100             | 90.9         | 396632     | PASS                |
| 443            | 442             | 15              | 24              | 19.7         | 78053      | PASS                |