

Data Path : Z:\HPCHEM1\BNA_F\Data\BF060616\
 Data File : BF087760.D
 Acq On : 6 Jun 2016 19:17
 Operator : UM/SJ
 Sample : H3341-02RE
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-2RE

Manual Integrations
 APPROVED

sohil
 6/7/2016 7:24:56 PM

Quant Time: Jun 07 11:54:21 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 04 11:07:28 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	103304	20.00	ng	-0.01
21) Naphthalene-d8	8.15	136	234154	20.00	ng	0.00
38) Acenaphthene-d10	9.90	164	187225	20.00	ng	-0.01
63) Phenanthrene-d10	11.37	188	271523	20.00	ng	-0.01
75) Chrysene-d12	14.00	240	230676	20.00	ng	-0.02
86) Perylene-d12	15.43	264	176683	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	714142	111.74	ng	0.00
7) Phenol-d6	6.50	99	899363	112.20	ng	0.01
23) Nitrobenzene-d5	7.43	82	501998	124.04	ng	0.00
41) 2,4,6-Tribromophenol	10.68	330	165336	87.97	ng	0.00
44) 2-Fluorobiphenyl	9.22	172	754143	55.62	ng	-0.01
78) Terphenyl-d14	12.96	244	526159	55.66	ng	-0.01
Target Compounds						Qvalue
10) Phenol	6.51	94	104604	11.46	ng	90
20) 3+4-Methylphenols	7.28	107	87488m	12.09	ng	
22) Acetophenone	7.27	105	581201	109.59	ng	# 84
27) 2,4-Dimethylphenol	7.82	122	112660m	28.88	ng	
31) Naphthalene	8.17	128	1486994	130.61	ng	99
37) 2-Methylnaphthalene	8.86	142	550297	73.20	ng	# 95
45) 1,1'-Biphenyl	9.31	154	78958	5.20	ng	97
49) Dimethylphthalate	9.61	163	85323	6.27	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\Data\BF060616\
Data File : BF087760.D
Acq On : 6 Jun 2016 19:17
Operator : UM/SJ
Sample : H3341-02RE
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-2RE

Manual Integrations
APPROVED

sohil
6/7/2016 7:24:56 PM

Quant Time: Jun 07 11:54:21 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 04 11:07:28 2016
Response via : Initial Calibration

