


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF060619\
 Data File : BF114864.D
 Acq On : 6 Jun 2019 15:56
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060419.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jun 04 16:57:15 2019



AutoFind: Scans 1591, 1592, 1593; Background Corrected with Scan 1585

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.3	113555	PASS
68	69	0.00	2	1.8	2159	PASS
69	198	0.00	100	31.6	118608	PASS
70	69	0.00	2	0.6	711	PASS
127	198	10	80	46.4	174320	PASS
197	198	0.00	2	0.5	1858	PASS
198	198	100	100	100.0	375381	PASS
199	198	5	9	6.7	25096	PASS
275	198	10	60	27.6	103669	PASS
365	198	1	100	3.2	11836	PASS
441	198	0.01	100	15.1	56640	PASS
442	198	50	100	96.9	363648	PASS
443	442	15	24	19.3	70253	PASS