

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060623\
 Data File : BF133614.D
 Acq On : 07 Jun 2023 04:45
 Operator : CG\JU
 Sample : 03002-05
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC04

Quant Time: Jun 07 05:20:52 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 06 14:05:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.834	152	126197	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	435322	20.000	ng	0.00
39) Acenaphthene-d10	9.875	164	184548	20.000	ng	0.00
64) Phenanthrene-d10	11.363	188	308254	20.000	ng	0.00
76) Chrysene-d12	14.010	240	156498	20.000	ng	0.00
86) Perylene-d12	15.474	264	133758	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.451	112	701206	92.707	ng	0.00
7) Phenol-d6	6.463	99	843533	90.596	ng	0.00
23) Nitrobenzene-d5	7.398	82	538739	82.630	ng	0.00
42) 2,4,6-Tribromophenol	10.669	330	163180	88.704	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	883434	73.885	ng	0.00
79) Terphenyl-d14	12.957	244	697298	79.220	ng	0.00
Target Compounds						Qvalue
9) Benzaldehyde	6.387	77	2029	Below Cal	#	48
19) n-Nitroso-di-n-propyla...	7.398	70	73505	12.201	ng	# 79
25) Isophorone	7.398	82	538733	35.454	ng	# 65
26) 2-Nitrophenol	7.504	139	148	4.784	ng	# 33
32) Benzoic acid	7.898	122	66	7.362	ng	# 35
48) 2-Nitroaniline	9.428	65	53	4.005	ng	# 6
51) 2,6-Dinitrotoluene	9.875	165	23895	12.628	ng	# 18
53) 3-Nitroaniline	9.780	138	62	3.199	ng	# 1
57) 2,4-Dinitrotoluene	9.875	165	23895	11.781	ng	# 16
62) 4-Nitroaniline	10.427	138	58	3.002	ng	# 19
63) Azobenzene	10.592	77	71873	5.332	ng	# 1
65) 4,6-Dinitro-2-methylph...	10.663	198	375	8.955	ng	# 1
66) n-Nitrosodiphenylamine	10.533	169	251101	25.091	ng	99
67) 4-Bromophenyl-phenylether	10.663	248	10327	3.155	ng	# 1
71) Phenanthrene	11.386	178	124393	7.967	ng	99
72) Anthracene	11.439	178	32589	2.028	ng	99
74) Di-n-butylphthalate	11.927	149	128832	7.170	ng	99
75) Fluoranthene	12.580	202	243971	14.147	ng	99
77) Benzidine	12.957	184	9983	2.181	ng	# 92
78) Pyrene	12.810	202	216922	16.289	ng	99
81) Benzo(a)anthracene	13.998	228	94586	9.073	ng	98
83) Chrysene	13.998	228	94586	8.987	ng	98
87) Indeno(1,2,3-cd)pyrene	16.933	276	29934	3.513	ng	99
88) Benzo(b)fluoranthene	15.051	252	119390	13.707	ng	# 94
89) Benzo(k)fluoranthene	15.051	252	119390	14.058	ng	# 95
90) Benzo(a)pyrene	15.415	252	67147	8.472	ng	97
92) Benzo(g,h,i)perylene	17.380	276	33627	4.828	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060623\
Data File : BF133614.D
Acq On : 07 Jun 2023 04:45
Operator : CG\JU
Sample : 03002-05
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC04

Quant Time: Jun 07 05:20:52 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 06 14:05:12 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.834 min Scan# 808
Delta R.T. -0.000 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45




Instrument :
BNA_F
ClientSampleId :
WC04


Tgt Ion:152 Resp: 126197
Ion Ratio Lower Upper
152 100
150 159.1 124.9 187.3
115 64.1 48.1 72.1



#5
2-Fluorophenol
Concen: 92.707 ng
RT: 5.451 min Scan# 572
Delta R.T. 0.006 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45

Tgt Ion:112 Resp: 701206
Ion Ratio Lower Upper
112 100
64 63.0 50.1 75.1
63 35.1 27.5 41.3









#7
Phenol-d6
Concen: 90.596 ng
RT: 6.463 min Scan# 744
Delta R.T. 0.000 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45




Instrument :
BNA_F
ClientSampleId :
WC04

Tgt Ion: 99 Resp: 843533
Ion Ratio Lower Upper
99 100
42 24.2 19.7 29.5
71 39.2 29.8 44.8



#9
Benzaldehyde
Concen: Below Cal
RT: 6.387 min Scan# 731
Delta R.T. 0.000 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45

Tgt Ion: 77 Resp: 2029
Ion Ratio Lower Upper
77 100
105 63.3 81.9 121.9#
106 32.6 77.8 117.8#









#19
n-Nitroso-di-n-propylamine
Concen: 12.201 ng
RT: 7.398 min Scan# 90
Delta R.T. 0.147 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45




Instrument :
BNA_F
ClientSampleId :
WC04

Tgt Ion:	70	Resp:	73505
Ion	Ratio	Lower	Upper
70	100		
42	57.4	57.2	85.8
101	0.0	7.3	10.9#
130	2.2	15.0	22.4#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.122 min Scan# 1026
Delta R.T. 0.000 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45

Tgt Ion:	136	Resp:	435322
Ion	Ratio	Lower	Upper
136	100		
137	11.4	8.2	12.4
54	8.6	6.6	10.0
68	5.6	4.6	7.0





#23

Nitrobenzene-d5

Concen: 82.630 ng

RT: 7.398 min Scan# 903

Delta R.T. -0.006 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45

Instrument :

BNA_F

ClientSampleId :

WC04





Tgt Ion: 82 Resp: 538739

Ion Ratio Lower Upper

82 100

128 41.8 34.2 51.4

54 58.6 48.2 72.4



#25

Isophorone

Concen: 35.454 ng

RT: 7.398 min Scan# 903

Delta R.T. -0.259 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45




Tgt Ion: 82 Resp: 538733


Ion Ratio Lower Upper

82 100

95 0.0 5.7 8.5#

138 0.0 14.2 21.4#









#26
2-Nitrophenol
Concen: 4.784 ng
RT: 7.504 min Scan# 91
Delta R.T. -0.229 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45




Instrument :
BNA_F
ClientSampleId :
WC04


Tgt Ion:139 Resp: 148
Ion Ratio Lower Upper
139 100
109 0.0 24.0 36.0#
65 0.0 41.7 62.5#



#32
Benzoic acid
Concen: 7.362 ng
RT: 7.898 min Scan# 988
Delta R.T. 0.012 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45

Tgt Ion:122 Resp: 66
Ion Ratio Lower Upper
122 100
105 0.0 103.3 143.3#
77 102.7 75.7 115.7









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.875 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45




Instrument :
BNA_F
ClientSampleId :
WC04


Tgt Ion:164 Resp: 184548
Ion Ratio Lower Upper
164 100
162 99.3 79.7 119.5
160 44.9 35.4 53.0



#42
2,4,6-Tribromophenol
Concen: 88.704 ng
RT: 10.669 min Scan# 1459
Delta R.T. 0.000 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45

Tgt Ion:330 Resp: 163180
Ion Ratio Lower Upper
330 100
332 96.5 77.8 116.6
141 32.6 26.6 40.0









#45
2-Fluorobiphenyl
Concen: 73.885 ng
RT: 9.198 min Scan# 1209
Delta R.T. 0.000 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45




Instrument :
BNA_F
ClientSampleId :
WC04


Tgt Ion:172 Resp: 883434
Ion Ratio Lower Upper
172 100
171 35.1 28.4 42.6
170 23.9 19.3 28.9



#48
2-Nitroaniline
Concen: 4.005 ng
RT: 9.428 min Scan# 1248
Delta R.T. 0.012 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45

Tgt Ion: 65 Resp: 53
Ion Ratio Lower Upper
65 100
92 0.0 50.8 76.2#
138 0.0 82.6 124.0#









#51
2,6-Dinitrotoluene
Concen: 12.628 ng
RT: 9.875 min Scan# 11
Delta R.T. 0.212 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45




Instrument :
BNA_F
ClientSampleId :
WC04


Tgt Ion:165 Resp: 23895
Ion Ratio Lower Upper
165 100
63 1.5 57.0 85.4#
89 2.0 50.8 76.2#



#53
3-Nitroaniline
Concen: 3.199 ng
RT: 9.780 min Scan# 1308
Delta R.T. -0.053 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45

Tgt Ion:138 Resp: 62
Ion Ratio Lower Upper
138 100
108 0.0 11.9 17.9#
92 0.0 93.1 139.7#









#57
2,4-Dinitrotoluene
Concen: 11.781 ng
RT: 9.875 min Scan# 11
Delta R.T. -0.188 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45




Instrument :
BNA_F
ClientSampleId :
WC04


Tgt Ion:165 Resp: 23895
Ion Ratio Lower Upper
165 100
63 1.5 48.1 72.1#
89 2.0 68.4 102.6#
182 0.0 1.9 2.9#



#62
4-Nitroaniline
Concen: 3.002 ng
RT: 10.427 min Scan# 1418
Delta R.T. -0.018 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45

Tgt Ion:138 Resp: 58
Ion Ratio Lower Upper
138 100
92 0.0 28.4 68.4#
108 0.0 54.2 94.2#









#63
Azobenzene
Concen: 5.332 ng
RT: 10.592 min Scan# 1446
Delta R.T. 0.012 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45




Instrument :
BNA_F
ClientSampleId :
WC04


Tgt Ion: 77 Resp: 71873
Ion Ratio Lower Upper
77 100
182 91.3 5.6 45.6#
105 162.1 2.6 42.6#
51 45.0 14.7 54.7



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.363 min Scan# 1577
Delta R.T. 0.000 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45

Tgt Ion: 188 Resp: 308254
Ion Ratio Lower Upper
188 100
94 10.0 8.6 12.8
80 10.7 9.2 13.8





#65

4,6-Dinitro-2-methylphenol

Concen: 8.955 ng

RT: 10.663 min Scan# 1458

Delta R.T. 0.188 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45

Instrument :

BNA_F

ClientSampleId :

WC04





Tgt Ion:198 Resp: 375

Ion Ratio Lower Upper

198 100

51 315.3 53.6 93.6#

105 200.2 35.9 75.9#



#66

n-Nitrosodiphenylamine

Concen: 25.091 ng

RT: 10.533 min Scan# 1436

Delta R.T. -0.006 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45




Tgt Ion:169 Resp: 251101


Ion Ratio Lower Upper

169 100

168 66.5 53.7 80.5

167 35.3 29.6 44.4









#67
4-Bromophenyl-phenylether
Concen: 3.155 ng
RT: 10.663 min Scan# 1458
Delta R.T. -0.247 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45




Instrument :
BNA_F
ClientSampleId :
WC04


Tgt Ion:248 Resp: 10327
Ion Ratio Lower Upper
248 100
250 192.6 79.7 119.5#
141 556.5 66.6 100.0#



#71
Phenanthrene
Concen: 7.967 ng
RT: 11.386 min Scan# 1581
Delta R.T. -0.006 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45

Tgt Ion:178 Resp: 124393
Ion Ratio Lower Upper
178 100
176 18.7 15.6 23.4
179 15.2 12.4 18.6





#72

Anthracene

Concen: 2.028 ng

RT: 11.439 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45

Instrument :

BNA_F

ClientSampleId :

WC04





Tgt Ion:178 Resp: 32589

Ion Ratio Lower Upper

178 100

176 19.3 15.3 22.9

179 16.8 12.6 18.8



#74

Di-n-butylphthalate

Concen: 7.170 ng

RT: 11.927 min Scan# 1673

Delta R.T. 0.000 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45




Tgt Ion:149 Resp: 128832


Ion Ratio Lower Upper

149 100

150 8.9 7.4 11.2

104 4.8 3.9 5.9









#75
Fluoranthene
Concen: 14.147 ng
RT: 12.580 min Scan# 1784
Delta R.T. 0.000 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45




Instrument :
BNA_F
ClientSampleId :
WC04


Tgt Ion:202 Resp: 243971
Ion Ratio Lower Upper
202 100
101 11.8 0.0 32.2
203 17.0 0.0 37.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.010 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BF133614.D
Acq: 07 Jun 2023 04:45

Tgt Ion:240 Resp: 156498
Ion Ratio Lower Upper
240 100
120 10.9 10.3 15.5
236 24.7 20.5 30.7





#77

Benzidine

Concen: 2.181 ng

RT: 12.957 min Scan# 1848

Delta R.T. 0.259 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45

Instrument :

BNA_F

ClientSampleId :

WC04





Tgt Ion:184 Resp: 9983

Ion Ratio Lower Upper

184 100

185 17.4 12.2 18.2

183 6.4 8.7 13.1#



#78

Pyrene

Concen: 16.289 ng

RT: 12.810 min Scan# 1823

Delta R.T. 0.000 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45




Tgt Ion:202 Resp: 216922


Ion Ratio Lower Upper

202 100

200 20.3 16.3 24.5

203 18.4 14.3 21.5









#79
 Terphenyl-d14
 Concen: 79.220 ng
 RT: 12.957 min Scan# 1848
 Delta R.T. 0.000 min
 Lab File: BF133614.D
 Acq: 07 Jun 2023 04:45




Instrument :
 BNA_F
 ClientSampleId :
 WC04


Tgt Ion:244 Resp: 697298
 Ion Ratio Lower Upper
 244 100
 212 6.6 5.8 8.6
 122 10.2 8.5 12.7



#81
 Benzo(a)anthracene
 Concen: 9.073 ng
 RT: 13.998 min Scan# 2025
 Delta R.T. 0.000 min
 Lab File: BF133614.D
 Acq: 07 Jun 2023 04:45

Tgt Ion:228 Resp: 94586
 Ion Ratio Lower Upper
 228 100
 226 29.2 22.3 33.5
 229 20.5 16.1 24.1





#83

Chrysene

Concen: 8.987 ng

RT: 13.998 min Scan# 2032

Delta R.T. -0.041 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45

Instrument :

BNA_F

ClientSampleId :

WC04





Tgt Ion:228 Resp: 94586

Ion Ratio Lower Upper

228 100

226 29.2 24.1 36.1

229 20.5 15.8 23.8



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.474 min Scan# 2276

Delta R.T. 0.006 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45




Tgt Ion:264 Resp: 133758


Ion Ratio Lower Upper

264 100

260 23.7 19.0 28.4

265 21.0 16.8 25.2





#87

Indeno(1,2,3-cd)pyrene

Concen: 3.513 ng

RT: 16.933 min Scan# 21

Delta R.T. 0.000 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45

Instrument :

BNA_F

ClientSampleId :

WC04





Tgt Ion:276 Resp: 29934

Ion Ratio Lower Upper

276 100

138 26.1 21.4 32.2

277 24.4 20.0 30.0



#88

Benzo(b)fluoranthene

Concen: 13.707 ng

RT: 15.051 min Scan# 2204

Delta R.T. 0.006 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45




Tgt Ion:252 Resp: 119390


Ion Ratio Lower Upper

252 100

253 23.8 17.2 25.8

125 13.2 8.1 12.1#





#89

Benzo(k)fluoranthene

Concen: 14.058 ng

RT: 15.051 min Scan# 21

Delta R.T. -0.023 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45

Instrument :

BNA_F

ClientSampleId :

WC04





Tgt Ion:252 Resp: 119390

Ion Ratio Lower Upper

252 100

253 23.8 17.4 26.2

125 13.2 8.3 12.5#



#90

Benzo(a)pyrene

Concen: 8.472 ng

RT: 15.415 min Scan# 2266

Delta R.T. 0.006 min

Lab File: BF133614.D

Acq: 07 Jun 2023 04:45




Tgt Ion:252 Resp: 67147


Ion Ratio Lower Upper

252 100

253 23.3 17.4 26.0

125 12.6 9.4 14.0





#92

Benzo(g,h,i)perylene

Concen: 4.828 ng

RT: 17.380 min Scan# 20

Delta R.T. 0.006 min

Lab File: BF133614.D


Acq: 07 Jun 2023 04:45

Instrument :

BNA_F

ClientSampleId :

WC04



Tgt Ion:276 Resp: 33627

Ion Ratio Lower Upper

276 100

277 26.5 18.8 28.2

138 30.4 17.8 26.6#

