

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060623\
 Data File : BF133602.D
 Acq On : 06 Jun 2023 20:54
 Operator : CG\JU
 Sample : 03002-04
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB04

Quant Time: Jun 07 04:02:03 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 06 14:05:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.840	152	122665	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	434773	20.000	ng	0.00
39) Acenaphthene-d10	9.875	164	175450	20.000	ng	0.00
64) Phenanthrene-d10	11.363	188	296135	20.000	ng	0.00
76) Chrysene-d12	14.010	240	150027	20.000	ng	0.00
86) Perylene-d12	15.480	264	129891	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.451	112	610492	83.038	ng	0.00
7) Phenol-d6	6.463	99	726279	80.249	ng	0.00
23) Nitrobenzene-d5	7.398	82	438323	67.314	ng	0.00
42) 2,4,6-Tribromophenol	10.669	330	146180	83.583	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	759201	66.787	ng	0.00
79) Terphenyl-d14	12.957	244	607985	72.053	ng	0.00
Target Compounds						
66) n-Nitrosodiphenylamine	10.533	169	27241	2.833	ng	98
92) Benzo(g,h,i)perylene	17.345	276	22029	3.257	ng	# 67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060623\
Data File : BF133602.D
Acq On : 06 Jun 2023 20:54
Operator : CG\JU
Sample : 03002-04
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB04

Quant Time: Jun 07 04:02:03 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 06 14:05:12 2023
Response via : Initial Calibration

