

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060724\
 Data File : BF138149.D
 Acq On : 07 Jun 2024 09:36
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 07 11:12:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 06 05:19:29 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	92	0.00
2	1,4-Dioxane	40.000	38.297	4.3	90	-0.02
3	Pyridine	40.000	37.303	6.7	87	-0.01
4	n-Nitrosodimethylamine	40.000	37.455	6.4	86	-0.01
5 S	2-Fluorophenol	80.000	79.268	0.9	92	0.00
6	Aniline	40.000	37.892	5.3	88	0.00
7 S	Phenol-d6	80.000	78.478	1.9	91	0.00
8	2-Chlorophenol	40.000	40.116	-0.3	92	0.00
9	Benzaldehyde	40.000	45.298	-13.2	96	0.00
10 C	Phenol	40.000	39.325	1.7	91	0.00
11	bis(2-Chloroethyl)ether	40.000	38.421	3.9	89	0.00
12	1,3-Dichlorobenzene	40.000	39.618	1.0	92	0.00
13 C	1,4-Dichlorobenzene	40.000	39.826	0.4	93	0.00
14	1,2-Dichlorobenzene	40.000	39.598	1.0	91	0.00
15	Benzyl Alcohol	40.000	39.148	2.1	89	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.345	6.6	87	0.00
17	2-Methylphenol	40.000	38.984	2.5	91	0.00
18	Hexachloroethane	40.000	40.587	-1.5	93	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.106	7.2	87	0.00
20	3+4-Methylphenols	40.000	39.892	0.3	90	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	91	0.00
22	Acetophenone	40.000	39.013	2.5	89	0.00
23 S	Nitrobenzene-d5	80.000	92.655	-15.8	98	0.00
24	Nitrobenzene	40.000	43.582	-9.0	94	0.00
25	Isophorone	40.000	38.226	4.4	88	0.00
26 C	2-Nitrophenol	40.000	47.148	-17.9	123	0.00
27	2,4-Dimethylphenol	40.000	40.531	-1.3	92	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.841	2.9	89	0.00
29 C	2,4-Dichlorophenol	40.000	42.371	-5.9	93	0.00
30	1,2,4-Trichlorobenzene	40.000	40.898	-2.2	93	0.00
31	Naphthalene	40.000	40.135	-0.3	92	0.00
32	Benzoic acid	40.000	50.062	-25.2#	145	0.00
33	4-Chloroaniline	40.000	39.570	1.1	91	0.00
34 C	Hexachlorobutadiene	40.000	41.339	-3.3	94	0.00
35	Caprolactam	40.000	39.178	2.1	89	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.196	-0.5	90	0.00
37	2-Methylnaphthalene	40.000	40.064	-0.2	92	0.00
38	1-Methylnaphthalene	40.000	40.046	-0.1	91	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	90	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.758	-4.4	93	0.00
41 P	Hexachlorocyclopentadiene	40.000	45.803	-14.5	96	0.00
42 S	2,4,6-Tribromophenol	80.000	93.791	-17.2	97	0.00
43 C	2,4,6-Trichlorophenol	40.000	44.187	-10.5	95	0.00
44	2,4,5-Trichlorophenol	40.000	43.852	-9.6	94	0.00
45 S	2-Fluorobiphenyl	80.000	81.804	-2.3	93	0.00
46	1,1'-Biphenyl	40.000	40.682	-1.7	92	0.00
47	2-Chloronaphthalene	40.000	40.253	-0.6	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060724\
 Data File : BF138149.D
 Acq On : 07 Jun 2024 09:36
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 07 11:12:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 06 05:19:29 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.828	-7.1	102	0.00
49	Acenaphthylene	40.000	40.270	-0.7	91	0.00
50	Dimethylphthalate	40.000	39.991	0.0	90	0.00
51	2,6-Dinitrotoluene	40.000	42.977	-7.4	99	0.00
52 C	Acenaphthene	40.000	41.249	-3.1	94	0.00
53	3-Nitroaniline	40.000	43.110	-7.8	100	0.00
54 P	2,4-Dinitrophenol	40.000	50.383	-26.0#	152	0.00
55	Dibenzofuran	40.000	40.188	-0.5	91	0.00
56 P	4-Nitrophenol	40.000	46.033	-15.1	98	0.00
57	2,4-Dinitrotoluene	40.000	45.005	-12.5	103	0.00
58	Fluorene	40.000	40.488	-1.2	91	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	45.310	-13.3	95	0.00
60	Diethylphthalate	40.000	40.063	-0.2	89	0.00
61	4-Chlorophenyl-phenylether	40.000	40.704	-1.8	92	0.00
62	4-Nitroaniline	40.000	47.332	-18.3	94	0.00
63	Azobenzene	40.000	39.453	1.4	88	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	88	0.00
65	4,6-Dinitro-2-methylphenol	40.000	48.396	-21.0	131	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.530	-1.3	90	0.00
67	4-Bromophenyl-phenylether	40.000	41.118	-2.8	91	0.00
68	Hexachlorobenzene	40.000	40.802	-2.0	91	0.00
69	Atrazine	40.000	38.318	4.2	85	0.00
70 C	Pentachlorophenol	40.000	48.273	-20.7#	97	0.00
71	Phenanthrene	40.000	39.699	0.8	88	0.00
72	Anthracene	40.000	39.995	0.0	90	0.00
73	Carbazole	40.000	38.861	2.8	87	0.00
74	Di-n-butylphthalate	40.000	39.126	2.2	86	0.00
75 C	Fluoranthene	40.000	38.496	3.8	86	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	84	0.00
77	Benzydine	40.000	49.634	-24.1	77	0.00
78	Pyrene	40.000	40.689	-1.7	85	0.00
79 S	Terphenyl-d14	80.000	81.345	-1.7	88	0.00
80	Butylbenzylphthalate	40.000	40.822	-2.1	79	0.00
81	Benzo(a)anthracene	40.000	39.840	0.4	85	0.00
82	3,3'-Dichlorobenzidine	40.000	44.295	-10.7	93	0.00
83	Chrysene	40.000	39.707	0.7	83	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.262	4.3	77	0.00
85 c	Di-n-octyl phthalate	40.000	39.794	0.5	81	0.00
86 I	Perylene-d12	20.000	20.000	0.0	110	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.011	-7.5	118	0.00
88	Benzo(b)fluoranthene	40.000	36.887	7.8	105	0.00
89	Benzo(k)fluoranthene	40.000	37.732	5.7	104	0.00
90 C	Benzo(a)pyrene	40.000	40.357	-0.9	108	0.00
91	Dibenzo(a,h)anthracene	40.000	42.996	-7.5	118	0.00
92	Benzo(g,h,i)perylene	40.000	46.177	-15.4	124	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060724\
Data File : BF138149.D
Acq On : 07 Jun 2024 09:36
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 07 11:12:53 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 06 05:19:29 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1