

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060823\
 Data File : BF133633.D
 Acq On : 08 Jun 2023 09:21
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 08 11:43:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 08 11:43:32 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.541	0.492	9.1	80	0.00
3	Pyridine	1.518	1.364	10.1	81	0.00
4	n-Nitrosodimethylamine	0.828	0.732	11.6	79	0.00
5 S	2-Fluorophenol	1.199	1.105	7.8	85	0.00
6	Aniline	2.004	1.837	8.3	84	0.00
7 S	Phenol-d6	1.476	1.365	7.5	86	0.00
8	2-Chlorophenol	1.181	1.130	4.3	86	0.00
9	Benzaldehyde	0.826	0.669	19.0	81	0.00
10 C	Phenol	1.480	1.417	4.3	86	0.00
11	bis(2-Chloroethyl)ether	1.224	1.102	10.0	83	0.00
12	1,3-Dichlorobenzene	1.311	1.208	7.9	85	0.00
13 C	1,4-Dichlorobenzene	1.315	1.222	7.1	85	0.00
14	1,2-Dichlorobenzene	1.213	1.127	7.1	86	0.00
15	Benzyl Alcohol	1.120	1.083	3.3	86	0.00
16	2,2'-oxybis(1-Chloropropane	2.321	1.944	16.2	76	0.00
17	2-Methylphenol	1.119	1.050	6.2	87	0.00
18	Hexachloroethane	0.455	0.447	1.8	87	0.00
19 P	n-Nitroso-di-n-propylamine	0.955	0.844	11.6	83	0.00
20	3+4-Methylphenols	1.265	1.217	3.8	85	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
22	Acetophenone	0.461	0.416	9.8	85	0.00
23 S	Nitrobenzene-d5	0.300	0.350	-16.7	96	0.00
24	Nitrobenzene	0.319	0.355	-11.3	92	0.00
25	Isophorone	0.698	0.643	7.9	85	0.00
26 C	2-Nitrophenol	0.121	0.173	-43.0#	115	0.00
27	2,4-Dimethylphenol	0.297	0.281	5.4	87	0.00
28	bis(2-Chloroethoxy)methane	0.405	0.368	9.1	84	0.00
29 C	2,4-Dichlorophenol	0.267	0.267	0.0	89	0.00
30	1,2,4-Trichlorobenzene	0.289	0.274	5.2	88	0.00
31	Naphthalene	0.970	0.899	7.3	87	0.00
32	Benzoic acid	0.172	0.207	-20.3	105	0.00
33	4-Chloroaniline	0.416	0.395	5.0	88	0.00
34 C	Hexachlorobutadiene	0.176	0.168	4.5	88	0.00
35	Caprolactam	0.088	0.094	-6.8	94	0.00
36 C	4-Chloro-3-methylphenol	0.299	0.299	0.0	90	0.00
37	2-Methylnaphthalene	0.662	0.621	6.2	87	0.00
38	1-Methylnaphthalene	0.617	0.579	6.2	87	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
40	1,2,4,5-Tetrachlorobenzene	0.593	0.558	5.9	90	0.00
41 P	Hexachlorocyclopentadiene	0.304	0.302	0.7	88	0.00
42 S	2,4,6-Tribromophenol	0.199	0.231	-16.1	103	0.00
43 C	2,4,6-Trichlorophenol	0.379	0.389	-2.6	92	0.00
44	2,4,5-Trichlorophenol	0.440	0.458	-4.1	92	0.00
45 S	2-Fluorobiphenyl	1.296	1.226	5.4	90	0.00
46	1,1'-Biphenyl	1.625	1.502	7.6	89	0.00
47	2-Chloronaphthalene	1.178	1.089	7.6	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060823\
 Data File : BF133633.D
 Acq On : 08 Jun 2023 09:21
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 08 11:43:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 08 11:43:32 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.320	0.407	-27.2#	104	0.00
49	Acenaphthylene	2.010	1.888	6.1	89	0.00
50	Dimethylphthalate	1.406	1.359	3.3	92	0.00
51	2,6-Dinitrotoluene	0.227	0.298	-31.3#	106	0.00
52 C	Acenaphthene	1.217	1.180	3.0	93	0.00
53	3-Nitroaniline	0.291	0.376	-29.2#	107	0.00
54 P	2,4-Dinitrophenol	0.097	0.155	-59.8#	154#	0.00
55	Dibenzofuran	1.811	1.694	6.5	89	0.00
56 P	4-Nitrophenol	0.240	0.275	-14.6	103	0.00
57	2,4-Dinitrotoluene	0.272	0.393	-44.5#	116	0.00
58	Fluorene	1.327	1.255	5.4	92	0.00
59	2,3,4,6-Tetrachlorophenol	0.332	0.363	-9.3	96	0.00
60	Diethylphthalate	1.459	1.392	4.6	90	0.00
61	4-Chlorophenyl-phenylether	0.637	0.612	3.9	94	0.00
62	4-Nitroaniline	0.297	0.365	-22.9	103	0.00
63	Azobenzene	1.461	1.339	8.4	87	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
65	4,6-Dinitro-2-methylphenol	0.070	0.107	-52.9#	145	0.00
66 c	n-Nitrosodiphenylamine	0.649	0.602	7.2	92	0.00
67	4-Bromophenyl-phenylether	0.212	0.202	4.7	91	0.00
68	Hexachlorobenzene	0.222	0.216	2.7	94	0.00
69	Atrazine	0.178	0.166	6.7	89	0.00
70 C	Pentachlorophenol	0.133	0.147	-10.5	97	0.00
71	Phenanthrene	1.013	0.951	6.1	93	0.00
72	Anthracene	1.043	0.986	5.5	94	0.00
73	Carbazole	0.986	0.929	5.8	93	0.00
74	Di-n-butylphthalate	1.166	1.094	6.2	90	0.00
75 C	Fluoranthene	1.119	1.074	4.0	94	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
77	Benzydine	0.585	0.401	31.5#	75	0.00
78	Pyrene	1.702	1.576	7.4	92	0.00
79 S	Terphenyl-d14	1.125	1.049	6.8	96	0.00
80	Butylbenzylphthalate	0.673	0.678	-0.7	92	0.00
81	Benzo(a)anthracene	1.332	1.266	5.0	94	0.00
82	3,3'-Dichlorobenzidine	0.416	0.398	4.3	93	0.00
83	Chrysene	1.345	1.256	6.6	91	0.00
84	Bis(2-ethylhexyl)phthalate	0.842	0.872	-3.6	97	0.00
85 c	Di-n-octyl phthalate	1.402	1.418	-1.1	86	0.00
86 I	Perylene-d12	1.000	1.000	0.0	90	0.00
87	Indeno(1,2,3-cd)pyrene	1.274	1.246	2.2	91	0.00
88	Benzo(b)fluoranthene	1.302	1.168	10.3	83	0.00
89	Benzo(k)fluoranthene	1.270	1.207	5.0	86	0.00
90 C	Benzo(a)pyrene	1.185	1.119	5.6	86	0.00
91	Dibenzo(a,h)anthracene	1.033	1.016	1.6	90	0.00
92	Benzo(g,h,i)perylene	1.042	1.039	0.3	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060823\
Data File : BF133633.D
Acq On : 08 Jun 2023 09:21
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 08 11:43:47 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 08 11:43:32 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1