

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060923\
 Data File : BF133694.D
 Acq On : 10 Jun 2023 01:59
 Operator : CG\JU
 Sample : 03092-02
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-1

Quant Time: Jun 10 02:37:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 09 15:37:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.839	152	116365	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	409722	20.000	ng	0.00
39) Acenaphthene-d10	9.880	164	163487	20.000	ng	0.00
64) Phenanthrene-d10	11.369	188	282908	20.000	ng	0.00
76) Chrysene-d12	14.015	240	145204	20.000	ng	0.00
86) Perylene-d12	15.480	264	116322	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	477659	71.461	ng	0.02
7) Phenol-d6	6.469	99	521889	59.983	ng	0.00
23) Nitrobenzene-d5	7.398	82	393403	52.297	ng	-0.01
42) 2,4,6-Tribromophenol	10.669	330	127349	61.537	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	658317	57.241	ng	0.00
79) Terphenyl-d14	12.957	244	580671	61.878	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060923\
Data File : BF133694.D
Acq On : 10 Jun 2023 01:59
Operator : CG\JU
Sample : 03092-02
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-1

Quant Time: Jun 10 02:37:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 09 15:37:41 2023
Response via : Initial Calibration

