

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061119\
 Data File : BF114946.D
 Acq On : 11 Jun 2019 15:11
 Operator : HP/JU
 Sample : K3250-01RE 2X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SCARSDALE-COMPOSITE-1RE

Manual Integrations
 APPROVED

mohammad
 6/12/2019 9:03:51 AM

Quant Time: Jun 11 16:38:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 16:44:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	253496	20.00	ng	0.00
21) Naphthalene-d8	7.94	136	1015501	20.00	ng	0.00
39) Acenaphthene-d10	9.69	164	495887	20.00	ng	0.00
64) Phenanthrene-d10	11.16	188	871030	20.00	ng	0.00
76) Chrysene-d12	13.79	240	561282	20.00	ng	0.00
87) Perylene-d12	15.20	264	373309	20.00	ng	0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.26	112	293547	19.60	ng	-0.01
7) Phenol-d6	6.30	99	375260	20.77	ng	-0.01
23) Nitrobenzene-d5	7.22	82	219203	13.26	ng	-0.01
42) 2,4,6-Tribromophenol	10.47	330	86694	16.07	ng	-0.01
45) 2-Fluorobiphenyl	9.02	172	424122	13.57	ng	0.00
79) Terphenyl-d14	12.74	244	359604	12.74	ng	0.00
Target Compounds						
71) Phenanthrene	11.19	178	86918	2.057	ng	97
78) Pyrene	12.60	202	298452	6.105	ng	96
83) Chrysene	13.82	228	238438m	5.721	ng	
92) Benzo(g,h,i)perylene	16.92	276	92420	5.200	ng	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061119\
Data File : BF114946.D
Acq On : 11 Jun 2019 15:11
Operator : HP/JU
Sample : K3250-01RE 2X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SCARSDALE-COMPOSITE-1RE

Manual Integrations
APPROVED

mohammad
6/12/2019 9:03:51 AM

Quant Time: Jun 11 16:38:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 16:44:00 2019
Response via : Initial Calibration

