

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061125\
 Data File : BF142731.D
 Acq On : 11 Jun 2025 13:18
 Operator : RC/JU
 Sample : PB168378BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB168378BL

Quant Time: Jun 11 13:43:41 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 11 05:56:09 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	74692	20.000	ng	0.00
21) Naphthalene-d8	8.175	136	287648	20.000	ng	0.00
39) Acenaphthene-d10	9.934	164	160376	20.000	ng	0.00
64) Phenanthrene-d10	11.422	188	295895	20.000	ng	0.00
76) Chrysene-d12	14.063	240	168179	20.000	ng	0.00
86) Perylene-d12	15.563	264	142470	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	375634	85.653	ng	0.01
7) Phenol-d6	6.516	99	443888	86.006	ng	0.00
23) Nitrobenzene-d5	7.451	82	416859	79.311	ng	0.00
42) 2,4,6-Tribromophenol	10.722	330	145994	83.060	ng	0.00
45) 2-Fluorobiphenyl	9.251	172	944326	78.228	ng	0.00
79) Terphenyl-d14	13.010	244	984688	80.415	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061125\
Data File : BF142731.D
Acq On : 11 Jun 2025 13:18
Operator : RC/JU
Sample : PB168378BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB168378BL

Quant Time: Jun 11 13:43:41 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061125.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 11 05:56:09 2025
Response via : Initial Calibration

