

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061218\
 Data File : BF106347.D
 Acq On : 12 Jun 2018 16:27
 Operator : JU/SJ
 Sample : J3247-03 2X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OR-03-061118

Manual Integrations
 APPROVED

Sohil
 6/13/2018 4:56:12 PM

Quant Time: Jun 13 02:12:27 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 11 15:00:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	162794	20.00	ng	0.00
21) Naphthalene-d8	8.25	136	561353	20.00	ng	0.00
38) Acenaphthene-d10	10.01	164	225793	20.00	ng	0.00
63) Phenanthrene-d10	11.51	188	444439	20.00	ng	0.00
75) Chrysene-d12	14.17	240	442188	20.00	ng	0.00
86) Perylene-d12	15.71	264	322018	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.59	112	507651	52.78	ng	0.00
7) Phenol-d6	6.59	99	601848	49.53	ng	0.00
23) Nitrobenzene-d5	7.53	82	362828	38.02	ng	0.00
41) 2,4,6-Tribromophenol	10.80	330	120123	55.29	ng	0.00
44) 2-Fluorobiphenyl	9.32	172	604256	38.16	ng	0.00
78) Terphenyl-d14	13.09	244	714650	40.14	ng	0.00
Target Compounds						
31) Naphthalene	8.28	128	118781	4.681	ng	98
49) Dimethylphthalate	9.72	163	91331	5.662	ng	# 99
51) Acenaphthene	10.04	154	32751	2.367	ng	97
54) Dibenzofuran	10.22	168	43069	2.312	ng	96
57) Fluorene	10.56	166	65862	4.462	ng	100
70) Phenanthrene	11.53	178	693194	31.848	ng	98
71) Anthracene	11.58	178	299884	13.753	ng	100
74) Fluoranthene	12.72	202	1275684	54.958	ng	97
77) Pyrene	12.95	202	1117937	40.383	ng	97
80) Benzo(a)anthracene	14.15	228	595165	22.618	ng	98
82) Chrysene	14.19	228	506680m	21.090	ng	
85) Indeno(1,2,3-cd)pyrene	17.28	276	160659m	8.379	ng	
87) Benzo(b)fluoranthene	15.25	252	488802m	25.111	ng	
88) Benzo(k)fluoranthene	15.28	252	130850m	6.787	ng	
89) Benzo(a)pyrene	15.65	252	325907	18.614	ng	97
90) Dibenzo(a,h)anthracene	17.29	278	37879	2.595	ng	# 87
91) Benzo(a,h,i)perylene	17.76	276	150170	10.586	ng	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061218\
Data File : BF106347.D
Acq On : 12 Jun 2018 16:27
Operator : JU/SJ
Sample : J3247-03 2X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OR-03-061118

Manual Integrations
APPROVED

Sohil
6/13/2018 4:56:12 PM

Quant Time: Jun 13 02:12:27 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 11 15:00:22 2018
Response via : Initial Calibration

