

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061221\
 Data File : BF124278.D
 Acq On : 12 Jun 2021 10:34
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Jun 14 04:23:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 10 01:53:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.839	152	43223	20.000	ng	-0.01
21) Naphthalene-d8	8.128	136	164294	20.000	ng	# 0.00
39) Acenaphthene-d10	9.880	164	98479	20.000	ng	-0.01
64) Phenanthrene-d10	11.369	188	177999	20.000	ng	0.00
76) Chrysene-d12	14.015	240	106866	20.000	ng	0.00
86) Perylene-d12	15.486	264	78231	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.469	112	210038	73.151	ng	-0.01
7) Phenol-d6	6.487	99	276475	71.635	ng	-0.01
23) Nitrobenzene-d5	7.410	82	287343	69.366	ng	0.00
42) 2,4,6-Tribromophenol	10.675	330	92786	66.503	ng	0.00
45) 2-Fluorobiphenyl	9.204	172	559281	71.465	ng	0.00
79) Terphenyl-d14	12.957	244	569500	73.166	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.575	88	43527	34.640	ng	91
3) Pyridine	3.334	79	115463	35.879	ng	# 78
4) n-Nitrosodimethylamine	3.293	42	66289	34.666	ng	# 94
6) Aniline	6.510	93	157927	35.825	ng	# 88
8) 2-Chlorophenol	6.634	128	115983	36.302	ng	96
9) Benzaldehyde	6.392	77	61851	36.053	ng	97
10) Phenol	6.498	94	150218	36.015	ng	# 65
11) bis(2-Chloroethyl)ether	6.581	93	108704	35.750	ng	94
12) 1,3-Dichlorobenzene	6.857	146	139102	37.243	ng	97
13) 1,4-Dichlorobenzene	6.857	146	139102	37.129	ng	96
14) 1,2-Dichlorobenzene	7.016	146	125496	35.016	ng	97
15) Benzyl Alcohol	6.987	79	125825	36.164	ng	91
16) 2,2'-oxybis(1-Chloropr...	7.122	45	152133	32.072	ng	81
17) 2-Methylphenol	7.104	107	93376	35.952	ng	# 90
18) Hexachloroethane	7.351	117	54217	36.748	ng	94
19) n-Nitroso-di-n-propyla...	7.263	70	98970	36.326	ng	# 84
20) 3+4-Methylphenols	7.257	107	122833	35.693	ng	# 78
22) Acetophenone	7.257	105	164566	35.137	ng	# 96
24) Nitrobenzene	7.428	77	144716	34.824	ng	93
25) Isophorone	7.669	82	254773	34.130	ng	# 97
26) 2-Nitrophenol	7.745	139	66711	35.937	ng	# 88
27) 2,4-Dimethylphenol	7.786	122	90484	34.567	ng	96
28) bis(2-Chloroethoxy)met...	7.875	93	128829	34.812	ng	99
29) 2,4-Dichlorophenol	7.992	162	107674	35.549	ng	93
30) 1,2,4-Trichlorobenzene	8.069	180	120568	35.520	ng	96
31) Naphthalene	8.151	128	343846	35.005	ng	99
32) Benzoic acid	7.922	122	55886	34.793	ng	91
33) 4-Chloroaniline	8.198	127	128711	35.230	ng	96
34) Hexachlorobutadiene	8.263	225	83554	34.849	ng	96
35) Caprolactam	8.575	113	20521	24.662	ng	# 66
36) 4-Chloro-3-methylphenol	8.686	107	116211	35.068	ng	91
37) 2-Methylnaphthalene	8.839	142	236530	34.338	ng	98
38) 1-Methylnaphthalene	8.939	142	226255	35.149	ng	98
40) 1,2,4,5-Tetrachloroben...	9.004	216	123372	35.359	ng	97
41) Hexachlorocyclopentadiene	8.992	237	56286	34.582	ng	96
43) 2,4,6-Trichlorophenol	9.122	196	78970	34.702	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061221\
 Data File : BF124278.D
 Acq On : 12 Jun 2021 10:34
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Quant Time: Jun 14 04:23:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 10 01:53:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.169	196	83285	34.093	ng	# 90
46) 1,1'-Biphenyl	9.304	154	291072	35.013	ng	99
47) 2-Chloronaphthalene	9.328	162	233703	34.538	ng	97
48) 2-Nitroaniline	9.428	65	84227	34.492	ng	96
49) Acenaphthylene	9.745	152	348372	35.419	ng	98
50) Dimethylphthalate	9.604	163	285359	35.024	ng	99
51) 2,6-Dinitrotoluene	9.669	165	66952	35.839	ng	92
52) Acenaphthene	9.916	154	221100	34.417	ng	96
53) 3-Nitroaniline	9.839	138	60825	34.788	ng	99
54) 2,4-Dinitrophenol	9.951	184	35330	38.877	ng	90
55) Dibenzofuran	10.092	168	352523	35.061	ng	96
56) 4-Nitrophenol	10.016	139	52086	41.684	ng	92
57) 2,4-Dinitrotoluene	10.075	165	90253	36.632	ng	# 90
58) Fluorene	10.433	166	279338	36.194	ng	96
59) 2,3,4,6-Tetrachlorophenol	10.210	232	70756	35.742	ng	# 87
60) Diethylphthalate	10.304	149	286919	34.894	ng	97
61) 4-Chlorophenyl-phenyle...	10.422	204	135994	34.462	ng	97
62) 4-Nitroaniline	10.457	138	60534	34.433	ng	# 79
63) Azobenzene	10.580	77	290274	33.866	ng	95
65) 4,6-Dinitro-2-methylph...	10.492	198	50707	35.448	ng	# 59
66) n-Nitrosodiphenylamine	10.545	169	231072	34.166	ng	95
67) 4-Bromophenyl-phenylether	10.910	248	86371	35.352	ng	# 86
68) Hexachlorobenzene	10.980	284	94659	34.207	ng	97
69) Atrazine	11.069	200	67552	35.284	ng	96
70) Pentachlorophenol	11.180	266	43704	32.887	ng	96
71) Phenanthrene	11.398	178	395704	35.587	ng	96
72) Anthracene	11.445	178	387578	34.611	ng	98
73) Carbazole	11.604	167	334040	34.230	ng	97
74) Di-n-butylphthalate	11.927	149	435674	34.739	ng	98
75) Fluoranthene	12.586	202	387030	33.160	ng	96
77) Benzidine	12.704	184	83651	34.806	ng	# 93
78) Pyrene	12.816	202	387993	37.809	ng	98
80) Butylbenzylphthalate	13.427	149	152903	36.730	ng	98
81) Benzo(a)anthracene	14.004	228	274298	35.234	ng	99
82) 3,3'-Dichlorobenzidine	13.963	252	82006	35.906	ng	97
83) Chrysene	14.039	228	263660	35.102	ng	98
84) Bis(2-ethylhexyl)phtha...	13.986	149	170090	34.349	ng	96
85) Di-n-octyl phthalate	14.598	149	235611	35.652	ng	# 92
87) Indeno(1,2,3-cd)pyrene	16.980	276	232642	36.838	ng	96
88) Benzo(b)fluoranthene	15.086	252	196118	32.272	ng	96
89) Benzo(k)fluoranthene	15.086	252	196118	34.090	ng	97
90) Benzo(a)pyrene	15.427	252	191580	36.295	ng	97
91) Dibenzo(a,h)anthracene	16.992	278	199521	36.960	ng	98
92) Benzo(g,h,i)perylene	17.427	276	201840	37.695	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 14 04:23:36 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 10 01:53:56 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.839 min Scan# 808
 Delta R.T. -0.012 min
 Lab File: BF124278.D
 Acq: 12 Jun 2021 10:34




Instrument :
 BNA_F
 ClientSampleID :
 SSTDCCC040


Tgt Ion:152 Resp: 43223
 Ion Ratio Lower Upper
 152 100
 150 170.3 127.0 190.6
 115 71.4 54.3 81.5



#2
 1,4-Dioxane
 Concen: 34.640 ng
 RT: 2.575 min Scan# 83
 Delta R.T. -0.023 min
 Lab File: BF124278.D
 Acq: 12 Jun 2021 10:34

Tgt Ion: 88 Resp: 43527
 Ion Ratio Lower Upper
 88 100
 58 95.8 83.8 125.8
 43 34.2 32.2 48.4









#3
Pyridine
Concen: 35.879 ng
RT: 3.334 min Scan# 212
Delta R.T. -0.023 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 115463
Ion Ratio Lower Upper
79 100
52 82.3 87.4 131.0#
51 41.2 41.4 62.0#



#4
n-Nitrosodimethylamine
Concen: 34.666 ng
RT: 3.293 min Scan# 205
Delta R.T. -0.023 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion: 42 Resp: 66289
Ion Ratio Lower Upper
42 100
74 97.8 83.5 125.3
44 5.5 5.7 8.5#









#5
2-Fluorophenol
Concen: 73.151 ng
RT: 5.469 min Scan# 575
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 210038
Ion Ratio Lower Upper
112 100
64 67.2 61.4 92.0
63 42.3 33.0 49.4



#6
Aniline
Concen: 35.825 ng
RT: 6.510 min Scan# 752
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion: 93 Resp: 157927
Ion Ratio Lower Upper
93 100
66 55.3 36.4 54.6#
65 27.0 19.4 29.0









#7
Phenol-d6
Concen: 71.635 ng
RT: 6.487 min Scan# 748
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 99	Resp: 276475
Ion Ratio	Lower Upper
99 100	
42 23.0	19.8 29.8
71 46.1	31.1 46.7



#8
2-Chlorophenol
Concen: 36.302 ng
RT: 6.634 min Scan# 773
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion: 128	Resp: 115983
Ion Ratio	Lower Upper
128 100	
130 32.6	12.6 52.6
64 56.5	41.5 81.5









#9
Benzaldehyde
Concen: 36.053 ng
RT: 6.392 min Scan# 732
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 77	Resp: 61851
Ion Ratio	Lower Upper
77 100	
105 94.6	75.2 115.2
106 86.3	71.6 111.6



#10
Phenol
Concen: 36.015 ng
RT: 6.498 min Scan# 750
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion: 94	Resp: 150218
Ion Ratio	Lower Upper
94 100	
65 51.4	18.1 58.1
66 80.0	29.0 69.0









#11

bis(2-Chloroethyl)ether
Concen: 35.750 ng
RT: 6.581 min Scan# 764
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion: 93 Resp: 108704
Ion Ratio Lower Upper
93 100
63 89.5 76.1 116.1
95 30.5 11.5 51.5




#12

1,3-Dichlorobenzene
Concen: 37.243 ng
RT: 6.857 min Scan# 811
Delta R.T. 0.065 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:146 Resp: 139102
Ion Ratio Lower Upper
146 100
148 60.7 50.9 76.3
75 41.7 34.2 51.4





#13

1,4-Dichlorobenzene

Concen: 37.129 ng

RT: 6.857 min Scan# 811

Delta R.T. -0.012 min

Lab File: BF124278.D

Acq: 12 Jun 2021 10:34





Tgt Ion:146 Resp: 139102

Ion Ratio Lower Upper

146 100

148 60.7 49.9 74.9

111 45.6 39.5 59.3



#14

1,2-Dichlorobenzene

Concen: 35.016 ng

RT: 7.016 min Scan# 838

Delta R.T. -0.006 min

Lab File: BF124278.D

Acq: 12 Jun 2021 10:34




Tgt Ion:146 Resp: 125496


Ion Ratio Lower Upper

146 100

148 66.4 51.6 77.4

111 50.4 38.4 57.6









#15
Benzyl Alcohol
Concen: 36.164 ng
RT: 6.987 min Scan# 833
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 79	Resp: 125825
Ion Ratio	Lower Upper
79	100
108	58.4 55.8 83.8
77	64.6 49.9 74.9



#16
2,2'-oxybis(1-chloropropane)
Concen: 32.072 ng
RT: 7.122 min Scan# 856
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion: 45	Resp: 152133
Ion Ratio	Lower Upper
45	100
77	17.9 0.0 30.0
79	13.2 0.0 27.3









#17
2-Methylphenol
Concen: 35.952 ng
RT: 7.104 min Scan# 853
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:	107	Resp:	93376
Ion	Ratio	Lower	Upper
107	100		
108	106.2	89.8	134.8
77	59.4	40.9	61.3
79	61.8	40.2	60.4



#18
Hexachloroethane
Concen: 36.748 ng
RT: 7.351 min Scan# 895
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:	117	Resp:	54217
Ion	Ratio	Lower	Upper
117	100		
119	89.2	77.8	116.8
201	87.2	67.8	101.8









#19
n-Nitroso-di-n-propylamine
Concen: 36.326 ng
RT: 7.263 min Scan# 880
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:	70	Resp:	98970
Ion Ratio	Lower	Upper	
70	100		
42	58.0	60.1	90.1#
101	8.3	7.4	11.2
130	19.0	14.2	21.2



#20
3+4-Methylphenols
Concen: 35.693 ng
RT: 7.257 min Scan# 879
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34


Tgt Ion:	107	Resp:	122833
Ion Ratio	Lower	Upper	
107	100		
108	85.3	71.5	111.5
77	143.0	82.0	122.0#
79	37.7	14.1	54.1








#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.128 min Scan# 1027
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040







Tgt Ion:136 Resp: 164294
Ion Ratio Lower Upper
136 100
137 9.7 8.7 13.1
54 8.1 9.8 14.6#
68 4.4 4.1 6.1



#22
Acetophenone
Concen: 35.137 ng
RT: 7.257 min Scan# 879
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:105 Resp: 164566
Ion Ratio Lower Upper
105 100
71 9.0 4.2 6.4#
51 44.9 33.6 50.4
120 20.0 16.3 24.5





#23

Nitrobenzene-d5

Concen: 69.366 ng

RT: 7.410 min Scan# 905

Delta R.T. -0.006 min

Lab File: BF124278.D

Acq: 12 Jun 2021 10:34

Instrument :





BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion: 82 Resp: 287343

Ion	Ratio	Lower	Upper
82	100		
128	35.9	29.4	44.2
54	60.9	54.1	81.1



#24

Nitrobenzene

Concen: 34.824 ng

RT: 7.428 min Scan# 908




Delta R.T. -0.006 min


Lab File: BF124278.D

Acq: 12 Jun 2021 10:34

Tgt Ion: 77 Resp: 144716

Ion	Ratio	Lower	Upper
77	100		
123	38.1	34.5	51.7
65	17.5	12.2	18.2









#25
Isophorone
Concen: 34.130 ng
RT: 7.669 min Scan# 949
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 254773
Ion Ratio Lower Upper
82 100
95 9.1 5.8 8.6#
138 15.3 11.8 17.8



#26
2-Nitrophenol
Concen: 35.937 ng
RT: 7.745 min Scan# 962
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion: 139 Resp: 66711
Ion Ratio Lower Upper
139 100
109 44.6 28.6 43.0#
65 62.5 56.8 85.2









#27
2,4-Dimethylphenol
Concen: 34.567 ng
RT: 7.786 min Scan# 969
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:	122	Resp:	90484
Ion	Ratio	Lower	Upper
122	100		
107	129.2	106.3	159.5
121	61.2	45.4	68.0



#28
bis(2-Chloroethoxy)methane
Concen: 34.812 ng
RT: 7.875 min Scan# 984
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:	93	Resp:	128829
Ion	Ratio	Lower	Upper
93	100		
95	33.6	26.5	39.7
123	11.5	8.4	12.6









#29
2,4-Dichlorophenol
Concen: 35.549 ng
RT: 7.992 min Scan# 1004
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 107674
Ion Ratio Lower Upper
162 100
164 63.4 44.1 84.1
98 37.4 27.3 67.3



#30
1,2,4-Trichlorobenzene
Concen: 35.520 ng
RT: 8.069 min Scan# 1017
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:180 Resp: 120568
Ion Ratio Lower Upper
180 100
182 92.4 77.2 115.8
145 31.0 26.0 39.0









#31
Naphthalene
Concen: 35.005 ng
RT: 8.151 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 343846
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 13.9 10.8 16.2



#32
Benzoic acid
Concen: 34.793 ng
RT: 7.922 min Scan# 992
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:122 Resp: 55886
Ion Ratio Lower Upper
122 100
105 115.0 107.8 147.8
77 96.0 81.6 121.6









#33
4-Chloroaniline
Concen: 35.230 ng
RT: 8.198 min Scan# 1039
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:127	Resp:	128711
Ion Ratio	Lower	Upper
127	100	
129	34.6	25.7 38.5
65	42.2	34.5 51.7
92	25.4	17.4 26.0



#34
Hexachlorobutadiene
Concen: 34.849 ng
RT: 8.263 min Scan# 1050
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:225	Resp:	83554
Ion Ratio	Lower	Upper
225	100	
223	58.0	50.4 75.6
227	63.1	51.4 77.0









#35
 Caprolactam
 Concen: 24.662 ng
 RT: 8.575 min Scan# 1103
 Delta R.T. -0.006 min
 Lab File: BF124278.D
 Acq: 12 Jun 2021 10:34




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:113 Resp: 20521
 Ion Ratio Lower Upper
 113 100
 55 203.8 254.9 294.9#
 56 155.3 179.4 219.4#



#36
 4-Chloro-3-methylphenol
 Concen: 35.068 ng
 RT: 8.686 min Scan# 1122
 Delta R.T. -0.012 min
 Lab File: BF124278.D
 Acq: 12 Jun 2021 10:34

Tgt Ion:107 Resp: 116211
 Ion Ratio Lower Upper
 107 100
 144 23.3 16.3 24.5
 142 70.4 50.1 75.1









#37
2-Methylnaphthalene
Concen: 34.338 ng
RT: 8.839 min Scan# 1148
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:142 Resp: 236530
Ion Ratio Lower Upper
142 100
141 91.5 71.8 107.6



#38
1-Methylnaphthalene
Concen: 35.149 ng
RT: 8.939 min Scan# 1165
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:142 Resp: 226255
Ion Ratio Lower Upper
142 100
141 92.7 72.8 109.2









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.880 min Scan# 1325
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:164	Resp:	98479
Ion Ratio	Lower	Upper
164	100	
162	104.6	80.7 121.1
160	47.6	37.3 55.9



#40
1,2,4,5-Tetrachlorobenzene
Concen: 35.359 ng
RT: 9.004 min Scan# 1176
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:216	Resp:	123372
Ion Ratio	Lower	Upper
216	100	
214	78.6	63.8 95.8
179	21.7	17.9 26.9
108	18.7	18.6 28.0









#41
Hexachlorocyclopentadiene
Concen: 34.582 ng
RT: 8.992 min Scan# 1174
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:237	Resp:	56286
Ion Ratio	Lower	Upper
237	100	
235	56.2	38.8 78.8
272	11.0	0.0 32.4



#42
2,4,6-Tribromophenol
Concen: 66.503 ng
RT: 10.675 min Scan# 1460
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:330	Resp:	92786
Ion Ratio	Lower	Upper
330	100	
332	100.2	77.1 115.7
141	39.8	27.9 41.9









#43
2,4,6-Trichlorophenol
Concen: 34.702 ng
RT: 9.122 min Scan# 1196
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:196 Resp: 78970
Ion Ratio Lower Upper
196 100
198 100.6 78.2 117.2
200 29.3 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 34.093 ng
RT: 9.169 min Scan# 1204
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:196 Resp: 83285
Ion Ratio Lower Upper
196 100
198 97.8 76.5 114.7
97 48.8 52.9 79.3#
132 27.3 25.3 37.9









#45
2-Fluorobiphenyl
Concen: 71.465 ng
RT: 9.204 min Scan# 1210
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:172 Resp: 559281
Ion Ratio Lower Upper
172 100
171 35.3 28.1 42.1
170 23.0 19.2 28.8



#46
1,1'-Biphenyl
Concen: 35.013 ng
RT: 9.304 min Scan# 1227
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:154 Resp: 291072
Ion Ratio Lower Upper
154 100
153 40.2 20.7 60.7
76 12.9 0.0 31.7









#47
2-Chloronaphthalene
Concen: 34.538 ng
RT: 9.328 min Scan# 1231
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 233703
Ion Ratio Lower Upper
162 100
127 41.8 36.2 54.2
164 31.3 25.4 38.2



#48
2-Nitroaniline
Concen: 34.492 ng
RT: 9.428 min Scan# 1248
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion: 65 Resp: 84227
Ion Ratio Lower Upper
65 100
92 56.0 47.1 70.7
138 84.5 70.2 105.2





#49

Acenaphthylene

Concen: 35.419 ng

RT: 9.745 min Scan# 1302

Delta R.T. -0.006 min

Lab File: BF124278.D

Acq: 12 Jun 2021 10:34

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:152 Resp: 348372

Ion Ratio Lower Upper

152 100

151 21.4 16.0 24.0

153 13.0 10.4 15.6



#50

Dimethylphthalate

Concen: 35.024 ng

RT: 9.604 min Scan# 1278

Delta R.T. -0.012 min

Lab File: BF124278.D

Acq: 12 Jun 2021 10:34




Tgt Ion:163 Resp: 285359


Ion Ratio Lower Upper

163 100

194 4.2 3.1 4.7

164 9.6 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 35.839 ng
RT: 9.669 min Scan# 1289
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 66952
Ion Ratio Lower Upper
165 100
63 79.4 67.7 101.5
89 64.7 58.8 88.2



#52
Acenaphthene
Concen: 34.417 ng
RT: 9.916 min Scan# 1331
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:154 Resp: 221100
Ion Ratio Lower Upper
154 100
153 113.0 88.1 132.1
152 50.4 43.4 65.0









#53
3-Nitroaniline
Concen: 34.788 ng
RT: 9.839 min Scan# 1318
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 60825
Ion Ratio Lower Upper
138 100
108 10.3 10.2 15.4
92 128.5 103.3 154.9



#54
2,4-Dinitrophenol
Concen: 38.877 ng
RT: 9.951 min Scan# 1337
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:184 Resp: 35330
Ion Ratio Lower Upper
184 100
63 81.8 73.5 110.3
154 72.9 52.5 78.7









#55
Dibenzofuran
Concen: 35.061 ng
RT: 10.092 min Scan# 1361
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 352523
Ion Ratio Lower Upper
168 100
139 38.6 33.2 49.8
169 12.8 10.8 16.2



#56
4-Nitrophenol
Concen: 41.684 ng
RT: 10.016 min Scan# 1348
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:139 Resp: 52086
Ion Ratio Lower Upper
139 100
109 102.6 63.2 103.2
65 130.0 110.0 150.0









#57
2,4-Dinitrotoluene
Concen: 36.632 ng
RT: 10.075 min Scan# 1358
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:	165	Resp:	90253
Ion Ratio	Lower	Upper	
165	100		
63	63.3	43.5	65.3
89	76.2	67.7	101.5
182	3.3	2.1	3.1#



#58
Fluorene
Concen: 36.194 ng
RT: 10.433 min Scan# 1419
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:	166	Resp:	279338
Ion Ratio	Lower	Upper	
166	100		
165	96.0	73.5	110.3
167	12.6	10.7	16.1









#59
2,3,4,6-Tetrachlorophenol
Concen: 35.742 ng
RT: 10.210 min Scan# 1381
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:232 Resp: 70756
Ion Ratio Lower Upper
232 100
131 41.3 43.2 64.8#
130 1.5 2.2 3.2#
166 29.8 27.3 40.9



#60
Diethylphthalate
Concen: 34.894 ng
RT: 10.304 min Scan# 1397
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:149 Resp: 286919
Ion Ratio Lower Upper
149 100
177 21.3 15.4 23.2
150 11.9 9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 34.462 ng
RT: 10.422 min Scan# 1417
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 135994
Ion Ratio Lower Upper
204 100
206 33.0 25.7 38.5
141 71.9 55.5 83.3



#62
4-Nitroaniline
Concen: 34.433 ng
RT: 10.457 min Scan# 1423
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:138 Resp: 60534
Ion Ratio Lower Upper
138 100
92 57.1 35.8 75.8
108 113.4 63.6 103.6#









#63
Azobenzene
Concen: 33.866 ng
RT: 10.580 min Scan# 1444
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:	77	Resp:	290274
Ion	Ratio	Lower	Upper
77	100		
182	20.0	1.7	41.7
105	18.0	1.7	41.7
51	37.8	15.9	55.9



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.369 min Scan# 1578
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:	188	Resp:	177999
Ion	Ratio	Lower	Upper
188	100		
94	7.3	6.9	10.3
80	10.9	8.2	12.2









#65
4,6-Dinitro-2-methylphenol
Concen: 35.448 ng
RT: 10.492 min Scan# 1429
Delta R.T. 0.000 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 50707
Ion Ratio Lower Upper
198 100
51 41.9 64.8 104.8#
105 36.6 42.6 82.6#



#66
n-Nitrosodiphenylamine
Concen: 34.166 ng
RT: 10.545 min Scan# 1438
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:169 Resp: 231072
Ion Ratio Lower Upper
169 100
168 71.9 53.3 79.9
167 35.4 29.9 44.9









#67
4-Bromophenyl-phenylether
Concen: 35.352 ng
RT: 10.910 min Scan# 1500
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 86371
Ion Ratio Lower Upper
248 100
250 105.0 77.4 116.2
141 97.7 63.7 95.5#



#68
Hexachlorobenzene
Concen: 34.207 ng
RT: 10.980 min Scan# 1512
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:284 Resp: 94659
Ion Ratio Lower Upper
284 100
142 36.1 28.3 42.5
249 31.9 23.5 35.3









#69
Atrazine
Concen: 35.284 ng
RT: 11.069 min Scan# 1527
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:200 Resp: 67552
Ion Ratio Lower Upper
200 100
173 25.4 6.9 46.9
215 43.3 26.5 66.5



#70
Pentachlorophenol
Concen: 32.887 ng
RT: 11.180 min Scan# 1546
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:266 Resp: 43704
Ion Ratio Lower Upper
266 100
268 63.2 51.6 77.4
264 60.3 52.6 78.8









#71
Phenanthrene
Concen: 35.587 ng
RT: 11.398 min Scan# 1583
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 395704
Ion Ratio Lower Upper
178 100
176 17.6 15.6 23.4
179 14.4 12.5 18.7



#72
Anthracene
Concen: 34.611 ng
RT: 11.445 min Scan# 1591
Delta R.T. -0.012 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34


Tgt Ion:178 Resp: 387578
Ion Ratio Lower Upper
178 100
176 17.9 15.4 23.0
179 16.3 12.6 18.8








#73
 Carbazole
 Concen: 34.230 ng
 RT: 11.604 min Scan# 1618
 Delta R.T. -0.006 min
 Lab File: BF124278.D
 Acq: 12 Jun 2021 10:34


Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040





Tgt Ion:167	Resp:	334040
Ion Ratio	Lower	Upper
167	100	
166	22.8	17.0 25.6
139	14.4	10.4 15.6




#74
 Di-n-butylphthalate
 Concen: 34.739 ng
 RT: 11.927 min Scan# 1673
 Delta R.T. -0.006 min
 Lab File: BF124278.D
 Acq: 12 Jun 2021 10:34



Tgt Ion:149	Resp:	435674
Ion Ratio	Lower	Upper
149	100	
150	8.8	7.8 11.8
104	6.1	4.6 6.8









#75
Fluoranthene
Concen: 33.160 ng
RT: 12.586 min Scan# 1785
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:202	Resp:	387030
Ion Ratio	Lower	Upper
202	100	
101	8.3	0.0 30.9
203	16.4	0.0 37.5



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.015 min Scan# 2028
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34


Tgt Ion:240	Resp:	106866
Ion Ratio	Lower	Upper
240	100	
120	7.4	7.2 10.8
236	23.3	19.0 28.4








#77
Benzidine
Concen: 34.806 ng
RT: 12.704 min Scan# 1805
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34


Instrument :
BNA_F
ClientSampleId :
SSTDCCC040





Tgt Ion:184 Resp: 83651
Ion Ratio Lower Upper
184 100
185 13.6 12.1 18.1
183 14.1 7.8 11.8#




#78
Pyrene
Concen: 37.809 ng
RT: 12.816 min Scan# 1824
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34



Tgt Ion:202 Resp: 387993
Ion Ratio Lower Upper
202 100
200 19.7 16.1 24.1
203 17.0 14.8 22.2









#79
 Terphenyl-d14
 Concen: 73.166 ng
 RT: 12.957 min Scan# 1848
 Delta R.T. -0.006 min
 Lab File: BF124278.D
 Acq: 12 Jun 2021 10:34




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:244 Resp: 569500
 Ion Ratio Lower Upper
 244 100
 212 7.1 5.7 8.5
 122 9.4 9.7 14.5#



#80
 Butylbenzylphthalate
 Concen: 36.730 ng
 RT: 13.427 min Scan# 1928
 Delta R.T. -0.006 min
 Lab File: BF124278.D
 Acq: 12 Jun 2021 10:34

Tgt Ion:149 Resp: 152903
 Ion Ratio Lower Upper
 149 100
 91 78.2 63.6 95.4
 206 19.7 14.5 21.7









#81
Benzo(a)anthracene
Concen: 35.234 ng
RT: 14.004 min Scan# 2026
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:228	Resp:	274298
Ion Ratio	Lower	Upper
228	100	
226	27.1	22.4 33.6
229	20.5	16.0 24.0



#82
3,3'-Dichlorobenzidine
Concen: 35.906 ng
RT: 13.963 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:252	Resp:	82006
Ion Ratio	Lower	Upper
252	100	
254	66.2	51.1 76.7
126	11.5	7.7 11.5









#83
Chrysene
Concen: 35.102 ng
RT: 14.039 min Scan# 2032
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 263660
Ion Ratio Lower Upper
228 100
226 29.1 24.3 36.5
229 19.1 16.2 24.2



#84
Bis(2-ethylhexyl)phthalate
Concen: 34.349 ng
RT: 13.986 min Scan# 2023
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:149 Resp: 170090
Ion Ratio Lower Upper
149 100
167 24.7 21.8 32.6
279 4.6 3.3 4.9









#85
Di-n-octyl phthalate
Concen: 35.652 ng
RT: 14.598 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 235611
Ion Ratio Lower Upper
149 100
167 1.6 1.2 1.8
43 11.1 11.8 17.6#



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.486 min Scan# 2278
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:264 Resp: 78231
Ion Ratio Lower Upper
264 100
260 23.2 18.7 28.1
265 20.9 17.8 26.8









#87
 Indeno(1,2,3-cd)pyrene
 Concen: 36.838 ng
 RT: 16.980 min Scan# 2532
 Delta R.T. 0.000 min
 Lab File: BF124278.D
 Acq: 12 Jun 2021 10:34




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:276 Resp: 232642
 Ion Ratio Lower Upper
 276 100
 138 22.5 16.2 24.2
 277 26.2 20.1 30.1



#88
 Benzo(b)fluoranthene
 Concen: 32.272 ng
 RT: 15.086 min Scan# 2210
 Delta R.T. 0.024 min
 Lab File: BF124278.D
 Acq: 12 Jun 2021 10:34

Tgt Ion:252 Resp: 196118
 Ion Ratio Lower Upper
 252 100
 253 20.8 18.1 27.1
 125 9.3 6.2 9.4









#89
Benzo(k)fluoranthene
Concen: 34.090 ng
RT: 15.086 min Scan# 2210
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 196118
Ion Ratio Lower Upper
252 100
253 20.8 17.9 26.9
125 9.3 7.0 10.6



#90
Benzo(a)pyrene
Concen: 36.295 ng
RT: 15.427 min Scan# 2268
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:252 Resp: 191580
Ion Ratio Lower Upper
252 100
253 21.3 18.4 27.6
125 10.4 8.0 12.0









#91
Dibenzo(a,h)anthracene
Concen: 36.960 ng
RT: 16.992 min Scan# 2534
Delta R.T. 0.000 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 199521
Ion Ratio Lower Upper
278 100
139 12.2 10.2 15.2
279 22.9 19.4 29.0



#92
Benzo(g,h,i)perylene
Concen: 37.695 ng
RT: 17.427 min Scan# 2608
Delta R.T. -0.006 min
Lab File: BF124278.D
Acq: 12 Jun 2021 10:34

Tgt Ion:276 Resp: 201840
Ion Ratio Lower Upper
276 100
277 24.1 18.2 27.2
138 18.2 13.8 20.6

