

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061221\
 Data File : BF124278.D
 Acq On : 12 Jun 2021 10:34
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 14 04:23:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 10 01:53:56 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	-0.01
2	1,4-Dioxane	40.000	34.640	13.4	91	-0.02
3	Pyridine	40.000	35.879	10.3	94	-0.02
4	n-Nitrosodimethylamine	40.000	34.666	13.3	90	-0.02
5 S	2-Fluorophenol	80.000	73.151	8.6	96	-0.01
6	Aniline	40.000	35.825	10.4	96	0.00
7 S	Phenol-d6	80.000	71.635	10.5	95	-0.01
8	2-Chlorophenol	40.000	36.302	9.2	94	-0.01
9	Benzaldehyde	40.000	36.053	9.9	99	0.00
10 C	Phenol	40.000	36.015	10.0	96	-0.01
11	bis(2-Chloroethyl)ether	40.000	35.750	10.6	93	0.00
12	1,3-Dichlorobenzene	40.000	37.090	7.3	97	0.00
13 C	1,4-Dichlorobenzene	40.000	37.129	7.2	97	-0.01
14	1,2-Dichlorobenzene	40.000	35.016	12.5	92	0.00
15	Benzyl Alcohol	40.000	36.164	9.6	97	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	32.072	19.8	87	0.00
17	2-Methylphenol	40.000	35.952	10.1	95	-0.01
18	Hexachloroethane	40.000	36.748	8.1	97	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	36.326	9.2	100	0.00
20	3+4-Methylphenols	40.000	35.693	10.8	95	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22	Acetophenone	40.000	35.137	12.2	96	0.00
23 S	Nitrobenzene-d5	80.000	69.366	13.3	96	0.00
24	Nitrobenzene	40.000	34.824	12.9	94	0.00
25	Isophorone	40.000	34.130	14.7	95	0.00
26 C	2-Nitrophenol	40.000	35.937	10.2	98	0.00
27	2,4-Dimethylphenol	40.000	34.567	13.6	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	34.812	13.0	100	0.00
29 C	2,4-Dichlorophenol	40.000	35.549	11.1	98	0.00
30	1,2,4-Trichlorobenzene	40.000	35.520	11.2	99	0.00
31	Naphthalene	40.000	35.005	12.5	97	0.00
32	Benzoic acid	40.000	34.793	13.0	101	0.00
33	4-Chloroaniline	40.000	35.230	11.9	97	-0.01
34 C	Hexachlorobutadiene	40.000	34.849	12.9	96	-0.01
35	Caprolactam	40.000	35.119	12.2	99	0.00
36 C	4-Chloro-3-methylphenol	40.000	35.068	12.3	102	-0.01
37	2-Methylnaphthalene	40.000	34.338	14.2	94	0.00
38	1-Methylnaphthalene	40.000	35.149	12.1	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	-0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	35.359	11.6	101	-0.01
41 P	Hexachlorocyclopentadiene	40.000	34.582	13.5	98	-0.01
42 S	2,4,6-Tribromophenol	80.000	66.503	16.9	93	0.00
43 C	2,4,6-Trichlorophenol	40.000	34.702	13.2	95	0.00
44	2,4,5-Trichlorophenol	40.000	34.093	14.8	95	0.00
45 S	2-Fluorobiphenyl	80.000	71.465	10.7	100	0.00
46	1,1'-Biphenyl	40.000	35.013	12.5	100	0.00
47	2-Chloronaphthalene	40.000	34.538	13.7	99	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061221\
 Data File : BF124278.D
 Acq On : 12 Jun 2021 10:34
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 14 04:23:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 10 01:53:56 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	34.492	13.8	97	-0.01
49	Acenaphthylene	40.000	35.419	11.5	99	0.00
50	Dimethylphthalate	40.000	35.024	12.4	102	-0.01
51	2,6-Dinitrotoluene	40.000	35.839	10.4	105	-0.01
52 C	Acenaphthene	40.000	34.417	14.0	97	-0.01
53	3-Nitroaniline	40.000	34.788	13.0	99	-0.01
54 P	2,4-Dinitrophenol	40.000	38.877	2.8	113	0.00
55	Dibenzofuran	40.000	35.061	12.3	100	0.00
56 P	4-Nitrophenol	40.000	41.684	-4.2	112	0.00
57	2,4-Dinitrotoluene	40.000	36.632	8.4	99	-0.01
58	Fluorene	40.000	36.194	9.5	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	35.742	10.6	103	-0.01
60	Diethylphthalate	40.000	34.894	12.8	98	0.00
61	4-Chlorophenyl-phenylether	40.000	34.462	13.8	99	0.00
62	4-Nitroaniline	40.000	34.433	13.9	95	0.00
63	Azobenzene	40.000	33.866	15.3	98	-0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	35.448	11.4	97	0.00
66 c	n-Nitrosodiphenylamine	40.000	34.166	14.6	94	0.00
67	4-Bromophenyl-phenylether	40.000	35.352	11.6	102	-0.01
68	Hexachlorobenzene	40.000	34.207	14.5	95	-0.01
69	Atrazine	40.000	35.284	11.8	89	-0.01
70 C	Pentachlorophenol	40.000	32.887	17.8	90	0.00
71	Phenanthrene	40.000	35.587	11.0	98	0.00
72	Anthracene	40.000	34.611	13.5	98	-0.01
73	Carbazole	40.000	34.230	14.4	95	0.00
74	Di-n-butylphthalate	40.000	34.739	13.2	97	0.00
75 C	Fluoranthene	40.000	33.160	17.1	92	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	93	0.00
77	Benzydine	40.000	34.806	13.0	74	0.00
78	Pyrene	40.000	37.809	5.5	92	0.00
79 S	Terphenyl-d14	80.000	73.166	8.5	89	0.00
80	Butylbenzylphthalate	40.000	36.730	8.2	90	0.00
81	Benzo(a)anthracene	40.000	35.234	11.9	88	0.00
82	3,3'-Dichlorobenzidine	40.000	35.906	10.2	89	0.00
83	Chrysene	40.000	35.102	12.2	87	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	34.349	14.1	84	0.00
85 c	Di-n-octyl phthalate	40.000	35.652	10.9	89	0.00
86 I	Perylene-d12	20.000	20.000	0.0	94	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.838	7.9	96	0.00
88	Benzo(b)fluoranthene	40.000	36.203	9.5	95	0.00
89	Benzo(k)fluoranthene	40.000	34.090	14.8	87	0.00
90 C	Benzo(a)pyrene	40.000	36.295	9.3	94	0.00
91	Dibenzo(a,h)anthracene	40.000	36.960	7.6	97	0.00
92	Benzo(g,h,i)perylene	40.000	37.695	5.8	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061221\
Data File : BF124278.D
Acq On : 12 Jun 2021 10:34
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 14 04:23:36 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 10 01:53:56 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0