

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061319\
 Data File : BF115008.D
 Acq On : 13 Jun 2019 18:17
 Operator : HP/JU
 Sample : K3345-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 C-MW-6_061119

Quant Time: Jun 14 04:09:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 12 15:42:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.65	152	298454	20.00	ng	0.00
21) Naphthalene-d8	7.93	136	1134266	20.00	ng	0.00
39) Acenaphthene-d10	9.67	164	553863	20.00	ng	0.00
64) Phenanthrene-d10	11.15	188	1056706	20.00	ng	-0.01
76) Chrysene-d12	13.77	240	733645	20.00	ng	-0.01
87) Perylene-d12	15.14	264	692001	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.25	112	1013707	56.71	ng	0.00
7) Phenol-d6	6.29	99	808134	37.48	ng	-0.01
23) Nitrobenzene-d5	7.21	82	1608089	85.34	ng	0.00
42) 2,4,6-Tribromophenol	10.46	330	681749	114.94	ng	0.00
45) 2-Fluorobiphenyl	9.00	172	2859237	87.49	ng	0.00
79) Terphenyl-d14	12.73	244	2957865	75.83	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061319\
Data File : BF115008.D
Acq On : 13 Jun 2019 18:17
Operator : HP/JU
Sample : K3345-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
C-MW-6_061119

Quant Time: Jun 14 04:09:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 12 15:42:37 2019
Response via : Initial Calibration

