

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061319\
 Data File : BF115010.D
 Acq On : 13 Jun 2019 19:10
 Operator : HP/JU
 Sample : K3345-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 C-MW-5_061119

Quant Time: Jun 14 04:12:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 12 15:42:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.65	152	249239	20.00	ng	0.00
21) Naphthalene-d8	7.93	136	958544	20.00	ng	0.00
39) Acenaphthene-d10	9.67	164	451877	20.00	ng	0.00
64) Phenanthrene-d10	11.15	188	810080	20.00	ng	-0.01
76) Chrysene-d12	13.77	240	543311	20.00	ng	-0.01
87) Perylene-d12	15.14	264	542906	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.25	112	741867	49.70	ng	0.00
7) Phenol-d6	6.28	99	590227	32.78	ng	-0.02
23) Nitrobenzene-d5	7.21	82	1152264	72.36	ng	0.00
42) 2,4,6-Tribromophenol	10.46	330	450543	93.11	ng	-0.01
45) 2-Fluorobiphenyl	9.00	172	2180233	81.77	ng	0.00
79) Terphenyl-d14	12.73	244	2108436	72.99	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061319\
Data File : BF115010.D
Acq On : 13 Jun 2019 19:10
Operator : HP/JU
Sample : K3345-03
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
C-MW-5_061119

Quant Time: Jun 14 04:12:56 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 12 15:42:37 2019
Response via : Initial Calibration

