

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061419\
 Data File : BF115065.D
 Acq On : 15 Jun 2019 2:35
 Operator : HP/JU
 Sample : K3366-03 10X
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Jun 15 04:57:33 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 12 15:42:37 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.64	152	290795	20.00	ng	0.00
21) Naphthalene-d8	7.93	136	1131571	20.00	ng	0.00
39) Acenaphthene-d10	9.67	164	509859	20.00	ng	0.00
64) Phenanthrene-d10	11.15	188	841914	20.00	ng	0.00
76) Chrysene-d12	13.77	240	523268	20.00	ng	-0.01
87) Perylene-d12	15.15	264	513042	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.25	112	73794	4.24	ng	-0.01
7) Phenol-d6	6.28	99	159040	7.57	ng	-0.02
23) Nitrobenzene-d5	7.20	82	132851	7.07	ng	-0.01
42) 2,4,6-Tribromophenol	10.46	330	6239	1.14	ng	0.00
45) 2-Fluorobiphenyl	9.00	172	313997	10.44	ng	0.00
79) Terphenyl-d14	12.73	244	231346	8.32	ng	-0.01
Target Compounds						
9) Benzaldehyde	6.41	77	1956	Below Cal		Qvalue 85
51) 2,6-Dinitrotoluene	9.67	165	66553	7.652 ng	#	37
57) 2,4-Dinitrotoluene	9.67	165	66553	6.012 ng	#	29
58) Fluorene	10.20	166	118	Below Cal	#	9


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061419\
Data File : BF115065.D
Acq On : 15 Jun 2019 2:35
Operator : HP/JU
Sample : K3366-03 10X
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Jun 15 04:57:33 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 12 15:42:37 2019
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.64 min Scan# 775
 Delta R.T. -0.01 min
 Lab File: BF115065.D
 Acq: 15 Jun 2019 2:35

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
152	100		
150	155.0	127.4	191.0
115	59.2	46.0	69.0




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E






#5
 2-Fluorophenol
 Concen: 4.237 ng
 RT: 5.25 min Scan# 537
 Delta R.T. -0.01 min
 Lab File: BF115065.D
 Acq: 15 Jun 2019 2:35

Tgt Ion	Ratio	Lower	Upper
112	100		
64	49.1	41.6	62.4
63	26.8	20.9	31.3



Abundance Ion 112.00 (111.70 to 112.70): E
 Ion 64.00 (63.70 to 64.70): BF1
 Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 7.571 ng

RT: 6.28 min Scan# 713

Delta R.T. -0.02 min

Lab File: BF115065.D

Acq: 15 Jun 2019 2:35

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 99 Resp: 159040

Ion Ratio Lower Upper

99 100


42 14.7 13.4 20.2

71 30.0 26.4 39.6




Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1

6.28



Time-->



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 7.93 min Scan# 993

Delta R.T. -0.01 min

Lab File: BF115065.D

Acq: 15 Jun 2019 2:35

Tgt Ion: 136 Resp: 1131571


Ion Ratio Lower Upper

136 100

137 11.3 9.0 13.4


54 6.7 5.4 8.0

68 5.7 4.6 7.0




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1

7.93




Time-->






#23
Nitrobenzene-d5
Concen: 7.067 ng
RT: 7.20 min Scan# 870
Delta R.T. -0.01 min
Lab File: BF115065.D
Acq: 15 Jun 2019 2:35

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
82	100		
128	49.5	39.9	59.9
54	43.0	37.3	55.9




Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 128.00 (127.70 to 128.70): BF1
Ion 54.00 (53.70 to 54.70): BF1






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.67 min Scan# 1290
Delta R.T. -0.01 min
Lab File: BF115065.D
Acq: 15 Jun 2019 2:35

Tgt Ion	Ratio	Lower	Upper
164	100		
162	103.0	82.6	124.0
160	46.8	36.9	55.3



Abundance Ion 164.00 (163.70 to 164.70): BF1
Ion 162.00 (161.70 to 162.70): BF1
Ion 160.00 (159.70 to 160.70): BF1





#42

2,4,6-Tribromophenol

Concen: 1.143 ng

RT: 10.46 min Scan# 1424

Delta R.T. -0.01 min

Lab File: BF115065.D

Acq: 15 Jun 2019 2:35

Instrument :

BNA_F

ClientSampleId :


Tot Ion:330 Resp: 6239

Ion Ratio Lower Upper

330 100

332 100.0 77.6 116.4




141 33.0 27.2 40.8



Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#45

2-Fluorobiphenyl

Concen: 10.438 ng

RT: 9.00 min Scan# 1176

Delta R.T. -0.01 min

Lab File: BF115065.D

Acq: 15 Jun 2019 2:35


Tgt Ion:172 Resp: 313997

Ion Ratio Lower Upper

172 100

171 35.0 29.8 44.6



170 22.9 20.6 31.0




Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E









#51
 2,6-Dinitrotoluene
 Concen: 7.652 ng
 RT: 9.67 min Scan# 1290
 Delta R.T. 0.20 min
 Lab File: BF115065.D
 Acq: 15 Jun 2019 2:35




Instrument :
 BNA_F
 ClientSampled :


Tot Ion	Ratio	Lower	Upper
165	100		
63	1.4	30.6	46.0#
89	1.2	34.4	51.6#



#57
 2,4-Dinitrotoluene
 Concen: 6.012 ng
 RT: 9.67 min Scan# 1290
 Delta R.T. -0.20 min
 Lab File: BF115065.D
 Acq: 15 Jun 2019 2:35

Tgt Ion	Ratio	Lower	Upper
165	100		
63	1.4	29.4	44.2#
89	1.2	50.6	75.8#
182	0.0	2.2	3.2#





#58

Fluorene

Concen: Below Cal

RT: 10.20 min Scan# 1380

Delta R.T. -0.02 min

Lab File: BF115065.D

Acq: 15 Jun 2019 2:35

Instrument :

BNA_F

ClientSampled :





Tot Ion:166 Resp: 118

Ion Ratio Lower Upper

166 100

165 0.0 77.6 116.4#

167 0.0 11.0 16.4#



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.15 min Scan# 1541

Delta R.T. -0.01 min

Lab File: BF115065.D

Acq: 15 Jun 2019 2:35




Tgt Ion:188 Resp: 841914


Ion Ratio Lower Upper

188 100

94 8.7 7.2 10.8

80 9.3 7.4 11.2









#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.77 min Scan# 1986
Delta R.T. -0.01 min
Lab File: BF115065.D
Acq: 15 Jun 2019 2:35

Instrument :
BNA_F
ClientSampleId :

Tot Ion:240 Resp: 523268
Ion Ratio Lower Upper
240 100
120 13.6 10.0 15.0
236 26.4 21.6 32.4




Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E






#79
Terphenyl-d14
Concen: 8.316 ng
RT: 12.73 min Scan# 1809
Delta R.T. -0.01 min
Lab File: BF115065.D
Acq: 15 Jun 2019 2:35

Tgt Ion:244 Resp: 231346
Ion Ratio Lower Upper
244 100
212 5.9 5.8 8.8
122 10.7 9.4 14.0



Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E





#87
Pervlene-d12
Concen: 20.000 ng
RT: 15.15 min Scan# 2221
Delta R.T. -0.00 min
Lab File: BF115065.D
Acq: 15 Jun 2019 2:35

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
264	100		
260	23.9	19.9	29.9
265	21.3	17.6	26.4

