

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061419\
 Data File : BF115028.D
 Acq On : 14 Jun 2019 10:04
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 17 03:48:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 12 15:42:37 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	0.00
2	1,4-Dioxane	0.561	0.577	-2.9	119	-0.01
3	Pyridine	1.578	1.586	-0.5	119	-0.01
4	n-Nitrosodimethylamine	0.560	0.582	-3.9	119	0.00
5 S	2-Fluorophenol	1.198	1.232	-2.8	120	0.00
6	Aniline	1.935	1.974	-2.0	118	0.00
7 S	Phenol-d6	1.445	1.494	-3.4	119	0.00
8	2-Chlorophenol	1.358	1.422	-4.7	121	0.00
9	Benzaldehyde	0.831	0.868	-4.5	121	0.00
10 C	Phenol	1.617	1.635	-1.1	119	0.00
11	bis(2-Chloroethyl)ether	1.255	1.319	-5.1	121	0.00
12	1,3-Dichlorobenzene	1.586	1.627	-2.6	120	0.00
13 C	1,4-Dichlorobenzene	1.595	1.625	-1.9	118	0.00
14	1,2-Dichlorobenzene	1.501	1.465	2.4	116	0.00
15	Benzyl Alcohol	1.083	0.897	17.2	97	0.00
16	2,2'-oxybis(1-Chloropropane	1.679	1.686	-0.4	116	0.00
17	2-Methylphenol	1.133	1.185	-4.6	119	0.00
18	Hexachloroethane	0.576	0.592	-2.8	119	0.00
19 P	n-Nitroso-di-n-propylamine	0.878	0.904	-3.0	122	0.00
20	3+4-Methylphenols	1.348	1.388	-3.0	124	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
22	Acetophenone	0.453	0.479	-5.7	124	0.00
23 S	Nitrobenzene-d5	0.332	0.344	-3.6	120	0.00
24	Nitrobenzene	0.338	0.358	-5.9	120	0.00
25	Isophorone	0.622	0.652	-4.8	122	0.00
26 C	2-Nitrophenol	0.188	0.205	-9.0	124	0.00
27	2,4-Dimethylphenol	0.272	0.296	-8.8	125	0.00
28	bis(2-Chloroethoxy)methane	0.399	0.417	-4.5	121	0.00
29 C	2,4-Dichlorophenol	0.287	0.311	-8.4	123	0.00
30	1,2,4-Trichlorobenzene	0.332	0.346	-4.2	120	0.00
31	Naphthalene	0.946	0.975	-3.1	119	0.00
32	Benzoic acid	0.197	0.192	2.5	114	0.02
33	4-Chloroaniline	0.437	0.464	-6.2	121	0.00
34 C	Hexachlorobutadiene	0.184	0.193	-4.9	121	0.00
35	Caprolactam	0.096	0.105	-9.4	126	0.01
36 C	4-Chloro-3-methylphenol	0.289	0.318	-10.0	126	0.00
37	2-Methylnaphthalene	0.631	0.646	-2.4	120	0.00
38	1-Methylnaphthalene	0.597	0.621	-4.0	121	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	122	0.00
40	1,2,4,5-Tetrachlorobenzene	0.591	0.600	-1.5	119	0.00
41 P	Hexachlorocyclopentadiene	0.291	0.290	0.3	112	0.00
42 S	2,4,6-Tribromophenol	0.214	0.224	-4.7	125	0.00
43 C	2,4,6-Trichlorophenol	0.441	0.438	0.7	116	0.00
44	2,4,5-Trichlorophenol	0.446	0.475	-6.5	126	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061419\
 Data File : BF115028.D
 Acq On : 14 Jun 2019 10:04
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 17 03:48:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 12 15:42:37 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.180	1.168	1.0	117	0.00
46	1,1'-Biphenyl	1.603	1.547	3.5	119	0.00
47	2-Chloronaphthalene	1.296	1.310	-1.1	120	0.00
48	2-Nitroaniline	0.383	0.412	-7.6	126	0.00
49	Acenaphthylene	1.993	1.926	3.4	118	0.00
50	Dimethylphthalate	1.504	1.552	-3.2	124	0.00
51	2,6-Dinitrotoluene	0.341	0.358	-5.0	123	0.00
52 C	Acenaphthene	1.140	1.148	-0.7	120	0.00
53	3-Nitroaniline	0.418	0.455	-8.9	128	0.00
54 P	2,4-Dinitrophenol	0.165	0.184	-11.5	126	0.00
55	Dibenzofuran	1.718	1.656	3.6	118	0.00
56 P	4-Nitrophenol	0.288	0.299	-3.8	120	0.00
57	2,4-Dinitrotoluene	0.434	0.461	-6.2	124	0.00
58	Fluorene	1.257	1.192	5.2	117	0.00
59	2,3,4,6-Tetrachlorophenol	0.360	0.385	-6.9	126	0.00
60	Diethylphthalate	1.438	1.469	-2.2	123	0.00
61	4-Chlorophenyl-phenylether	0.620	0.585	5.6	118	0.00
62	4-Nitroaniline	0.420	0.463	-10.2	130	0.00
63	Azobenzene	1.258	1.286	-2.2	120	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	121	0.00
65	4,6-Dinitro-2-methylphenol	0.112	0.123	-9.8	125	0.00
66 c	n-Nitrosodiphenylamine	0.598	0.608	-1.7	120	0.00
67	4-Bromophenyl-phenylether	0.215	0.221	-2.8	123	0.00
68	Hexachlorobenzene	0.236	0.246	-4.2	122	0.00
69	Atrazine	0.193	0.186	3.6	115	0.00
70 C	Pentachlorophenol	0.128	0.144	-12.5	128	0.00
71	Phenanthrene	1.033	1.018	1.5	120	0.00
72	Anthracene	1.042	1.013	2.8	120	0.00
73	Carbazole	0.909	0.955	-5.1	124	0.00
74	Di-n-butylphthalate	1.146	1.106	3.5	117	0.00
75 C	Fluoranthene	1.056	1.046	0.9	121	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	121	0.00
77	Benzidine	0.809	0.774	4.3	111	0.00
78	Pyrene	1.822	1.825	-0.2	123	0.00
79 S	Terphenyl-d14	1.063	1.023	3.8	120	0.00
80	Butylbenzylphthalate	0.851	0.873	-2.6	123	0.00
81	Benzo(a)anthracene	1.497	1.523	-1.7	121	0.00
82	3,3'-Dichlorobenzidine	0.574	0.625	-8.9	124	0.00
83	Chrysene	1.488	1.553	-4.4	125	0.00
84	Bis(2-ethylhexyl)phthalate	1.006	0.958	4.8	115	0.00
85 c	Di-n-octyl phthalate	1.732	1.741	-0.5	118	0.00
86	Indeno(1,2,3-cd)pyrene	1.307	1.302	0.4	124	0.00
87 I	Perylene-d12	1.000	1.000	0.0	121	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061419\
Data File : BF115028.D
Acq On : 14 Jun 2019 10:04
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 17 03:48:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 12 15:42:37 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.330	1.424	-7.1	130	0.00
89	Benzo(k)fluoranthene	1.167	1.224	-4.9	121	0.00
90 C	Benzo(a)pyrene	1.147	1.215	-5.9	126	0.00
91	Dibenzo(a,h)anthracene	0.988	1.005	-1.7	123	0.00
92	Benzo(q,h,i)perylene	1.008	1.030	-2.2	125	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0