

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061623\
 Data File : BF133823.D
 Acq On : 17 Jun 2023 00:13
 Operator : CG\JU
 Sample : 03260-01DL 5X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-1DL

Quant Time: Jun 17 01:03:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 16 14:54:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.863	152	86346	20.000	ng	0.00
21) Naphthalene-d8	8.145	136	297715	20.000	ng	0.00
39) Acenaphthene-d10	9.898	164	118668	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	206907	20.000	ng	# 0.01
76) Chrysene-d12	14.057	240	128945	20.000	ng	# 0.00
86) Perylene-d12	15.562	264	112537	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	89473	18.039	ng	0.00
7) Phenol-d6	6.492	99	104107	16.125	ng	0.00
23) Nitrobenzene-d5	7.422	82	61503	11.252	ng	0.00
42) 2,4,6-Tribromophenol	10.692	330	24049	16.010	ng	0.00
45) 2-Fluorobiphenyl	9.216	172	121610	14.568	ng	0.00
79) Terphenyl-d14	12.998	244	100720	12.086	ng	0.00
Target Compounds						
84) Bis(2-ethylhexyl)phtha...	14.039	149	468074	74.966	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061623\
Data File : BF133823.D
Acq On : 17 Jun 2023 00:13
Operator : CG\JU
Sample : 03260-01DL 5X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-1DL

Quant Time: Jun 17 01:03:29 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 16 14:54:03 2023
Response via : Initial Calibration

