



2,4-Dinitrophenol

Response Ratio



Resp Ratio = $1.92 \times 10^{-1} \times \text{Amt} - 7.07 \times 10^{-2}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear


Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF062119.M
Calibration Table Last Updated: Mon Jun 24 04:21:29 2019



Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF062119.M
Calibration Table Last Updated: Mon Jun 24 04:21:29 2019

4-Chlorophenyl-phenylether


Response Ratio



Resp Ratio = $4.88 \times 10^{-1} \times \text{Amt} + 1.95 \times 10^{-1}$
Coef of Det (r^2) = 0.990 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF062119.M
Calibration Table Last Updated: Mon Jun 24 04:21:29 2019

4,6-Dinitro-2-methylphenol



Resp Ratio = $1.19 \times 10^{-1} \times \text{Amt} - 2.94 \times 10^{-2}$
Coef of Det (r^2) = 0.999 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF062119.M
Calibration Table Last Updated: Mon Jun 24 04:21:29 2019