

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062119\
 Data File : BF115085.D
 Acq On : 21 Jun 2019 18:38
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 22 04:53:50 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 21 17:55:32 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.612	0.615	-0.5	100	0.00
3	Pyridine	1.724	1.736	-0.7	100	0.00
4	n-Nitrosodimethylamine	0.676	0.669	1.0	99	0.00
5 S	2-Fluorophenol	1.296	1.321	-1.9	100	0.00
6	Aniline	2.162	2.207	-2.1	100	0.00
7 S	Phenol-d6	1.573	1.608	-2.2	100	0.00
8	2-Chlorophenol	1.457	1.509	-3.6	101	0.00
9	Benzaldehyde	1.026	1.078	-5.1	100	0.00
10 C	Phenol	1.843	1.873	-1.6	100	0.00
11	bis(2-Chloroethyl)ether	1.358	1.394	-2.7	102	0.00
12	1,3-Dichlorobenzene	1.617	1.673	-3.5	102	0.00
13 C	1,4-Dichlorobenzene	1.622	1.666	-2.7	100	0.00
14	1,2-Dichlorobenzene	1.502	1.566	-4.3	101	0.00
15	Benzyl Alcohol	1.145	1.198	-4.6	100	0.00
16	2,2'-oxybis(1-Chloropropane	1.970	2.014	-2.2	101	0.00
17	2-Methylphenol	1.203	1.217	-1.2	101	0.00
18	Hexachloroethane	0.585	0.600	-2.6	101	0.00
19 P	n-Nitroso-di-n-propylamine	1.007	1.026	-1.9	102	0.00
20	3+4-Methylphenols	1.389	1.430	-3.0	101	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
22	Acetophenone	0.458	0.469	-2.4	100	0.00
23 S	Nitrobenzene-d5	0.325	0.340	-4.6	102	0.00
24	Nitrobenzene	0.358	0.373	-4.2	101	0.00
25	Isophorone	0.666	0.674	-1.2	101	0.00
26 C	2-Nitrophenol	0.177	0.190	-7.3	104	0.00
27	2,4-Dimethylphenol	0.290	0.301	-3.8	101	0.00
28	bis(2-Chloroethoxy)methane	0.421	0.437	-3.8	101	0.00
29 C	2,4-Dichlorophenol	0.299	0.312	-4.3	101	0.00
30	1,2,4-Trichlorobenzene	0.330	0.346	-4.8	101	0.00
31	Naphthalene	0.982	1.034	-5.3	101	0.00
32	Benzoic acid	0.207	0.213	-2.9	118	0.00
33	4-Chloroaniline	0.449	0.458	-2.0	99	0.00
34 C	Hexachlorobutadiene	0.181	0.193	-6.6	103	0.00
35	Caprolactam	0.100	0.101	-1.0	101	0.00
36 C	4-Chloro-3-methylphenol	0.303	0.310	-2.3	100	0.00
37	2-Methylnaphthalene	0.662	0.689	-4.1	101	0.00
38	1-Methylnaphthalene	0.632	0.662	-4.7	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	0.565	0.595	-5.3	101	0.00
41 P	Hexachlorocyclopentadiene	0.335	0.353	-5.4	103	0.00
42 S	2,4,6-Tribromophenol	0.211	0.226	-7.1	104	0.00
43 C	2,4,6-Trichlorophenol	0.436	0.452	-3.7	100	0.00
44	2,4,5-Trichlorophenol	0.449	0.471	-4.9	104	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062119\
 Data File : BF115085.D
 Acq On : 21 Jun 2019 18:38
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 22 04:53:50 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 21 17:55:32 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.177	1.256	-6.7	101	0.00
46	1,1'-Biphenyl	1.600	1.620	-1.3	100	0.00
47	2-Chloronaphthalene	1.298	1.364	-5.1	101	0.00
48	2-Nitroaniline	0.378	0.399	-5.6	102	0.00
49	Acenaphthylene	2.022	2.139	-5.8	101	0.00
50	Dimethylphthalate	1.471	1.547	-5.2	101	0.00
51	2,6-Dinitrotoluene	0.340	0.356	-4.7	102	0.00
52 C	Acenaphthene	1.266	1.332	-5.2	101	0.00
53	3-Nitroaniline	0.415	0.437	-5.3	102	0.00
54 P	2,4-Dinitrophenol	0.150	0.162	-8.0	108	0.00
55	Dibenzofuran	1.816	1.845	-1.6	100	0.00
56 P	4-Nitrophenol	0.332	0.349	-5.1	100	0.00
57	2,4-Dinitrotoluene	0.439	0.474	-8.0	104	0.00
58	Fluorene	1.352	1.294	4.3	100	0.00
59	2,3,4,6-Tetrachlorophenol	0.377	0.394	-4.5	101	0.00
60	Diethylphthalate	1.479	1.558	-5.3	103	0.00
61	4-Chlorophenyl-phenylether	0.645	0.630	2.3	102	0.00
62	4-Nitroaniline	0.416	0.437	-5.0	102	0.00
63	Azobenzene	1.382	1.451	-5.0	102	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	0.100	0.107	-7.0	106	0.00
66 c	n-Nitrosodiphenylamine	0.623	0.646	-3.7	102	0.00
67	4-Bromophenyl-phenylether	0.217	0.226	-4.1	103	0.00
68	Hexachlorobenzene	0.235	0.244	-3.8	103	0.00
69	Atrazine	0.200	0.207	-3.5	101	0.00
70 C	Pentachlorophenol	0.150	0.161	-7.3	104	0.00
71	Phenanthrene	1.077	1.076	0.1	102	0.00
72	Anthracene	1.087	1.099	-1.1	101	0.00
73	Carbazole	0.971	1.012	-4.2	102	0.00
74	Di-n-butylphthalate	1.122	1.199	-6.9	104	0.00
75 C	Fluoranthene	1.074	1.126	-4.8	102	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
77	Benzidine	0.888	0.928	-4.5	107	0.00
78	Pyrene	1.537	1.522	1.0	103	0.00
79 S	Terphenyl-d14	0.941	0.944	-0.3	104	0.00
80	Butylbenzylphthalate	0.678	0.707	-4.3	109	0.00
81	Benzo(a)anthracene	1.435	1.464	-2.0	106	0.00
82	3,3'-Dichlorobenzidine	0.518	0.543	-4.8	106	0.00
83	Chrysene	1.315	1.325	-0.8	105	0.00
84	Bis(2-ethylhexyl)phthalate	0.889	0.952	-7.1	112	0.00
85 c	Di-n-octyl phthalate	1.493	1.612	-8.0	112	0.00
86	Indeno(1,2,3-cd)pyrene	1.556	1.651	-6.1	108	0.00
87 I	Perylene-d12	1.000	1.000	0.0	108	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062119\
Data File : BF115085.D
Acq On : 21 Jun 2019 18:38
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 22 04:53:50 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 21 17:55:32 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.248	1.337	-7.1	109	0.00
89	Benzo(k)fluoranthene	1.119	1.055	5.7	106	0.00
90 C	Benzo(a)pyrene	1.125	1.157	-2.8	106	0.00
91	Dibenzo(a,h)anthracene	1.050	1.137	-8.3	108	0.00
92	Benzo(g,h,i)perylene	1.063	1.133	-6.6	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0