

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062123\
 Data File : BF133869.D
 Acq On : 21 Jun 2023 14:11
 Operator : CG\JU
 Sample : 03304-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 JC-CONCRETE-0616

Quant Time: Jun 21 15:13:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 13:43:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.857	152	112711	20.000	ng	0.00
21) Naphthalene-d8	8.139	136	431505	20.000	ng	0.00
39) Acenaphthene-d10	9.898	164	211142	20.000	ng	0.00
64) Phenanthrene-d10	11.392	188	371767	20.000	ng	0.00
76) Chrysene-d12	14.051	240	189805	20.000	ng	0.00
86) Perylene-d12	15.557	264	227072	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	571843	88.325	ng	0.02
7) Phenol-d6	6.498	99	724077	85.920	ng	0.00
23) Nitrobenzene-d5	7.422	82	515468	65.065	ng	0.00
42) 2,4,6-Tribromophenol	10.692	330	192607	72.065	ng	0.00
45) 2-Fluorobiphenyl	9.216	172	969423	65.267	ng	0.00
79) Terphenyl-d14	12.986	244	780756	63.649	ng	0.00
Target Compounds						
						Qvalue
55) Dibenzofuran	10.098	168	73309	3.861	ng	99
71) Phenanthrene	11.416	178	653820	34.664	ng	99
75) Fluoranthene	12.610	202	425844	21.261	ng	99
78) Pyrene	12.839	202	272677	15.422	ng	99
81) Benzo(a)anthracene	14.039	228	45228	3.445	ng	98
83) Chrysene	14.074	228	43244	3.482	ng	95
88) Benzo(b)fluoranthene	15.104	252	31490m	2.283	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062123\
Data File : BF133869.D
Acq On : 21 Jun 2023 14:11
Operator : CG\JU
Sample : 03304-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
JC-CONCRETE-0616

Quant Time: Jun 21 15:13:25 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 20 13:43:46 2023
Response via : Initial Calibration

