

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062218\
 Data File : BF106724.D
 Acq On : 22 Jun 2018 22:21
 Operator : JU/SJ
 Sample : J3594-05
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HC-SOIL-3

Quant Time: Jun 23 00:14:30 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 22 11:34:03 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.96	152	50150	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	186789	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	82259	20.00	ng	0.00
63) Phenanthrene-d10	11.49	188	140334	20.00	ng	-0.01
75) Chrysene-d12	14.14	240	146924	20.00	ng	-0.01
86) Perylene-d12	15.67	264	122251	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.61	112	328879	111.05	ng	0.00
7) Phenol-d6	6.60	99	412918	111.64	ng	0.00
23) Nitrobenzene-d5	7.52	82	259020	77.06	ng	0.00
41) 2,4,6-Tribromophenol	10.80	330	108893	114.21	ng	0.00
44) 2-Fluorobiphenyl	9.32	172	429012	88.96	ng	0.00
78) Terphenyl-d14	13.08	244	430510	70.98	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.71	163	59622	9.664	ng	Qvalue 100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062218\
Data File : BF106724.D
Acq On : 22 Jun 2018 22:21
Operator : JU/SJ
Sample : J3594-05
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HC-SOIL-3

Quant Time: Jun 23 00:14:30 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 22 11:34:03 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.96 min Scan# 786
Delta R.T. -0.00 min
Lab File: BF106724.D
Acq: 22 Jun 2018 22:21

Instrument :
BNA_F
ClientSampleId :
HC-SOIL-3

Tot Ion:152 Resp: 50150
Ion Ratio Lower Upper
152 100
150 161.0 124.6 186.8
115 59.9 50.1 75.1




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 111.046 ng
RT: 5.61 min Scan# 556
Delta R.T. 0.01 min
Lab File: BF106724.D
Acq: 22 Jun 2018 22:21

Tgt Ion:112 Resp: 328879
Ion Ratio Lower Upper
112 100
64 57.1 47.9 71.9
63 28.4 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 111.644 ng

RT: 6.60 min Scan# 725

Delta R.T. -0.01 min

Lab File: BF106724.D

Acq: 22 Jun 2018 22:21

Instrument :

BNA_F

ClientSampled :

HC-SOIL-3


Tot Ion: 99 Resp: 412918

Ion Ratio Lower Upper

99 100

42 16.3 14.5 21.7




71 34.1 25.4 38.0



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.24 min Scan# 1004

Delta R.T. -0.01 min

Lab File: BF106724.D

Acq: 22 Jun 2018 22:21

Tgt Ion: 136 Resp: 186789


Ion Ratio Lower Upper

136 100

137 10.6 8.9 13.3

54 7.6 6.6 9.8

68 6.3 5.0 7.4






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1





#23

Nitrobenzene-d5

Concen: 77.064 ng

RT: 7.52 min Scan# 882

Delta R.T. -0.01 min

Lab File: BF106724.D

Acq: 22 Jun 2018 22:21

Instrument :

BNA_F

ClientSampleId :

HC-SOIL-3


Tot Ion: 82 Resp: 259020

Ion Ratio Lower Upper




82 100

128 46.3 36.1 54.1

54 48.8 41.4 62.2



Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 128.00 (127.70 to 128.70): BF1
Ion 54.00 (53.70 to 54.70): BF1



#38

Acenaphthene-d10

Concen: 20.000 ng

RT: 10.00 min Scan# 1303

Delta R.T. -0.01 min

Lab File: BF106724.D

Acq: 22 Jun 2018 22:21


Tgt Ion:164 Resp: 82259

Ion Ratio Lower Upper



164 100


162 95.5 86.1 129.1

160 43.8 38.3 57.5



Abundance Ion 164.00 (163.70 to 164.70): BF1
Ion 162.00 (161.70 to 162.70): BF1
Ion 160.00 (159.70 to 160.70): BF1









#41
 2,4,6-Tribromophenol
 Concen: 114.215 ng
 RT: 10.80 min Scan# 1438
 Delta R.T. -0.01 min
 Lab File: BF106724.D
 Acq: 22 Jun 2018 22:21




Instrument :
 BNA_F
 ClientSampleId :
 HC-SOIL-3


Tot Ion	Ratio	Lower	Upper
330	100		
332	94.9	77.0	115.6
141	23.5	29.9	44.9#



#44
 2-Fluorobiphenyl
 Concen: 88.965 ng
 RT: 9.32 min Scan# 1187
 Delta R.T. -0.01 min
 Lab File: BF106724.D
 Acq: 22 Jun 2018 22:21

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.1	29.7	44.5
170	23.7	19.6	29.4





#49

Dimethylphthalate

Concen: 9.664 ng

RT: 9.71 min Scan# 1253

Delta R.T. -0.02 min

Lab File: BF106724.D

Acq: 22 Jun 2018 22:21

Instrument :

BNA_F

ClientSampleId :

HC-SOIL-3


Tot Ion:163 Resp: 59622

Ion Ratio Lower Upper

163 100

194 3.7 2.7 4.1




164 10.2 8.2 12.2



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.49 min Scan# 1556

Delta R.T. -0.01 min

Lab File: BF106724.D

Acq: 22 Jun 2018 22:21


Tgt Ion:188 Resp: 140334

Ion Ratio Lower Upper

188 100

94 9.4 8.5 12.7



80 10.3 9.2 13.8




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1





#75

Chrysene-d12

Concen: 20.000 ng

RT: 14.14 min Scan# 2007

Delta R.T. -0.01 min

Lab File: BF106724.D

Acq: 22 Jun 2018 22:21

Instrument :

BNA_F

ClientSampleId :

HC-SOIL-3


Tot Ion:240 Resp: 146924

Ion Ratio Lower Upper

240 100

120 8.7 9.5 14.3#




236 25.8 20.7 31.1



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#78

Terphenyl-d14

Concen: 70.977 ng

RT: 13.08 min Scan# 1826

Delta R.T. -0.01 min

Lab File: BF106724.D

Acq: 22 Jun 2018 22:21


Tgt Ion:244 Resp: 430510

Ion Ratio Lower Upper

244 100

212 6.3 5.4 8.0



122 8.6 11.0 16.4#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E






#86
Pervlene-d12
Concen: 20.000 ng
RT: 15.67 min Scan# 2267
Delta R.T. -0.02 min
Lab File: BF106724.D
Acq: 22 Jun 2018 22:21

Instrument :
BNA_F
ClientSampleId :
HC-SOIL-3

Tot Ion:	264	Resp:	122251
Ion	Ratio	Lower	Upper
264	100		
260	23.7	19.2	28.8
265	21.6	17.5	26.3



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E

