


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF062219\
 Data File : BF115136.D
 Acq On : 22 Jun 2019 23:29
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF062119.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 24 04:21:29 2019



AutoFind: Scans 1607, 1608, 1609; Background Corrected with Scan 1602

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.7	60770	PASS
68	69	0.00	2	1.7	1019	PASS
69	198	0.00	100	32.4	60056	PASS
70	69	0.00	2	0.7	446	PASS
127	198	10	80	47.7	88472	PASS
197	198	0.00	2	0.3	544	PASS
198	198	100	100	100.0	185602	PASS
199	198	5	9	6.7	12427	PASS
275	198	10	60	25.1	46634	PASS
365	198	1	100	2.9	5301	PASS
441	198	0.01	100	11.4	21140	PASS
442	198	50	100	69.2	128525	PASS
443	442	15	24	19.2	24632	PASS