

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062223\
 Data File : BF133914.D
 Acq On : 22 Jun 2023 22:15
 Operator : CG\JU
 Sample : 03320-17
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-16

Quant Time: Jun 23 05:05:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 23 02:21:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.857	152	101162	20.000	ng	0.00
21) Naphthalene-d8	8.139	136	394586	20.000	ng	0.00
39) Acenaphthene-d10	9.892	164	201297	20.000	ng	0.00
64) Phenanthrene-d10	11.386	188	424883	20.000	ng	0.00
76) Chrysene-d12	14.051	240	332500	20.000	ng	0.00
86) Perylene-d12	15.557	264	271126	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	830152	142.861	ng	0.02
7) Phenol-d6	6.498	99	933621	123.432	ng	0.00
23) Nitrobenzene-d5	7.422	82	748165	103.273	ng	0.00
42) 2,4,6-Tribromophenol	10.692	330	409364	160.657	ng	0.00
45) 2-Fluorobiphenyl	9.216	172	1363991	96.322	ng	0.00
79) Terphenyl-d14	12.992	244	1905419	88.672	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062223\
Data File : BF133914.D
Acq On : 22 Jun 2023 22:15
Operator : CG\JU
Sample : 03320-17
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-16

Quant Time: Jun 23 05:05:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 23 02:21:26 2023
Response via : Initial Calibration

