

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062223\
 Data File : BF133918.D
 Acq On : 23 Jun 2023 00:24
 Operator : CG\JU
 Sample : 03321-08
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-23

Quant Time: Jun 23 05:06:02 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 23 02:21:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.857	152	107005	20.000	ng	0.00
21) Naphthalene-d8	8.133	136	413214	20.000	ng	0.00
39) Acenaphthene-d10	9.892	164	218165	20.000	ng	0.00
64) Phenanthrene-d10	11.386	188	484366	20.000	ng	0.00
76) Chrysene-d12	14.045	240	380611	20.000	ng	-0.01
86) Perylene-d12	15.551	264	282235	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	787698	128.153	ng	0.02
7) Phenol-d6	6.498	99	888182	111.013	ng	0.00
23) Nitrobenzene-d5	7.422	82	720828	95.014	ng	0.00
42) 2,4,6-Tribromophenol	10.692	330	407805	147.671	ng	0.00
45) 2-Fluorobiphenyl	9.216	172	1329458	86.625	ng	0.00
79) Terphenyl-d14	12.992	244	1812441	73.683	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062223\
Data File : BF133918.D
Acq On : 23 Jun 2023 00:24
Operator : CG\JU
Sample : 03321-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-23

Quant Time: Jun 23 05:06:02 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 23 02:21:26 2023
Response via : Initial Calibration

